

Chemical Analysis Report

CEN-DIST-2020-09-23-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Ingram/Farrar/Tennessee**

Request ID: **RQ-2020-09-21-58**

Customer: **CEN-DIST**

Project ID: **SMP**

Send Reports to:

FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact

Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 09-OCT-2020 11:30



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Ingram East

Collection Date/Time: 09/22/2020 11:48

Field ID: G4CE0155

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2197598	EPA 180.1 Rev. 2.0	Turbidity	1.5		NTU	P387751	
	EPA 300.0 Rev. 2.1	Chloride	7.5		mg Cl/L	P388030	
		Sulfate	1.7		mg SO4/L	P388032	
	SM 2120 B-2011	Color (true)	26		PCU	P387763	
	SM 2320 B-2011	Total Alkalinity	8.2		mg CaCO3/L	P387828	
	SM 2540 C-2011	TDS	37		mg/L	P388196	
	SM 2540 D-2011	TSS	2	I	mg/L	P388199	
	SM 4500 F-C-2011	Fluoride	0.040	I	mg F/L	P387943	
2197600	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P387887	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.79		mg N/L	P387868	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P387787	
	EPA 365.1 Rev. 2.0	Total-P	0.011		mg P/L	P388156	
	SM 5310 B-2011	Organic Carbon	9.9		mg C/L	P387981	
2197602	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P387759	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Field Blank

Collection Date/Time: 09/22/2020 12:04

Field ID: FB092220

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2197599	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P387751	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P388030	
		Sulfate	0.20	U	mg SO4/L	P388032	
	SM 2120 B-2011	Color (true)	2.5	U	PCU	P387763	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P387828	
	SM 2540 C-2011	TDS	15	U	mg/L	P388196	
	SM 2540 D-2011	TSS	2	U	mg/L	P388199	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P387943	
2197601	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P387887	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P387868	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P387787	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P388156	
	SM 5310 B-2011	Organic Carbon	0.50	U	mg C/L	P387981	
2197603	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P387759	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387751

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387887

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387868

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387787

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P388156

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387759

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P387763

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P387828

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P388196

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P388199

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P387981

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387887

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	94.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387868

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387787

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P388156

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387759

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.5		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P387763

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.4		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P387828

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	93 - 104

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	100		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P387981

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197327	Chloride	103		P	90 - 110
2197587	Chloride	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197327	Sulfate	100		P	90 - 110
2197587	Sulfate	98.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387887

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197593	Ammonia-N	102	99.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387868

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197600	Kjeldahl Nitrogen	96.5	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387787

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197561	NO2NO3-N	94.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P388156

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197552	Total-P	104	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387759

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197562	O-Phosphate-P	98.7	99.3	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197717	Fluoride	100	100	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P387981

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197588	Organic Carbon	94.9	98.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387751

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197566	Turbidity	0.367	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197327	Chloride	0.509	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197327	Sulfate	0.774	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387887

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197593	Ammonia-N	1.86	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387868

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197600	Kjeldahl Nitrogen	2.62	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387787

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197561	NO2NO3-N	0.400	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P388156

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197552	Total-P	3.37	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387759

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197562	O-Phosphate-P	0.485	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P387763

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197592	Color (true)	0.446	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P387828

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197717	Total Alkalinity	0.488	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P388196

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197560	TDS	4.18	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197717	Fluoride	0.0	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P387981

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197588	Organic Carbon	2.99	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A101160

Included Lab Sample IDs: 2197598, 2197599

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A101162

Included Lab Sample IDs: 2197602, 2197603

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	100	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A101163

Included Lab Sample IDs: 2197598, 2197599

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.1	99.4	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A101175

Included Lab Sample IDs: 2197600, 2197601

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	98.5	97.5	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A101193

Included Lab Sample IDs: 2197598, 2197599

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.7	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A101210

Included Lab Sample IDs: 2197600, 2197601

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	95.9	98.6	P/P	90 - 110
Ammonia-N	98.6	99.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A101228

Included Lab Sample IDs: 2197600, 2197601

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.9	100	P/P	90 - 110
Kjeldahl Nitrogen	99.7	94.9	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A101239

Included Lab Sample IDs: 2197598, 2197599

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-2011

Run ID: A101239

Included Lab Sample IDs: 2197598, 2197599

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	104	104	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A101262

Included Lab Sample IDs: 2197600, 2197601

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.2	98.4	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A101281

Included Lab Sample IDs: 2197598, 2197599

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101384

Included Lab Sample IDs: 2197600, 2197601

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	94.5	94.5	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					0.367	
EPA 300.0 Rev. 2.1	Chloride	101	103	104		0.509	
	Sulfate	98.8	100	98.4		0.774	
EPA 350.1 Rev. 2.0	Ammonia-N	94.6	102	99.6			1.86
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.0	96.5	101			2.62
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	98.5	94.8				0.400
EPA 365.1 Rev. 2.0	Total-P	106	104	107			3.37
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.5	98.7	99.3			0.485
SM 2120 B-2011	Color (true)	98.4				0.446	
SM 2320 B-2011	Total Alkalinity	101				0.488	
SM 2540 C-2011	TDS					4.18	
SM 4500 F-C-2011	Fluoride	100	100	100			0.0
SM 5310 B-2011	Organic Carbon	97.5	94.9	98.6			2.99

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2197598, 2197599
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2197598, 2197599
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2197598, 2197599
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2197600, 2197601
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2197600, 2197601
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2197600, 2197601
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2197600, 2197601
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2197602, 2197603
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2197598, 2197599
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2197598, 2197599
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2197598, 2197599
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2197598, 2197599
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2197598, 2197599
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2197600, 2197601

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/23/2020			09/23/2020 15:04	Yen Ta	2197598, 2197599
EPA 300.0 Rev. 2.1	09/23/2020			09/30/2020 02:00	Elliott Rodriguez	2197598
	09/23/2020			09/30/2020 02:13	Elliott Rodriguez	2197599
EPA 350.1 Rev. 2.0	09/23/2020			09/26/2020 13:57	Ping Hua	2197600
	09/23/2020			09/26/2020 14:14	Ping Hua	2197601
EPA 351.2 Rev. 2.0	09/23/2020	09/25/2020 11:45	David Boose	09/28/2020 11:34	Elliott Rodriguez	2197600
	09/23/2020	09/25/2020 11:45	David Boose	09/28/2020 11:49	Elliott Rodriguez	2197601
EPA 353.2 Rev. 2.0	09/23/2020			09/24/2020 10:21	Beverly Y. Sanders	2197600
	09/23/2020			09/24/2020 10:22	Beverly Y. Sanders	2197601
EPA 365.1 Rev. 2.0	09/23/2020	10/02/2020 14:22	Yen Ta	10/06/2020 13:32	Lipi Saha	2197600
	09/23/2020	10/02/2020 14:22	Yen Ta	10/06/2020 13:33	Lipi Saha	2197601
EPA 365.1 Rev. 2.0 dissolved	09/23/2020			09/23/2020 15:32	Virginia P. Brown	2197602
	09/23/2020			09/23/2020 15:33	Virginia P. Brown	2197603
SM 2120 B-2011	09/23/2020			09/23/2020 16:43	Benjamin T. Nordmann	2197598
	09/23/2020			09/23/2020 16:44	Benjamin T. Nordmann	2197599
SM 2320 B-2011	09/23/2020			09/24/2020 17:20	Dale L. Simmons	2197598
	09/23/2020			09/24/2020 17:29	Dale L. Simmons	2197599
SM 2540 C-2011	09/23/2020			09/28/2020 15:25	Yijie Li	2197598, 2197599
SM 2540 D-2011	09/23/2020			09/28/2020 15:25	Yijie Li	2197598, 2197599
SM 4500 F-C-2011	09/23/2020			09/25/2020 20:13	Dale L. Simmons	2197598
	09/23/2020			09/25/2020 20:18	Dale L. Simmons	2197599
SM 5310 B-2011	09/23/2020			09/28/2020 09:27	David Boose	2197600
	09/23/2020			09/28/2020 09:40	David Boose	2197601