

Chemical Analysis Report

CEN-DIST-2020-10-20-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SJR Above Puzzle**
Request ID: **RQ-2020-09-28-42**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 10-NOV-2020 11:09

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Haines @ Center

Collection Date/Time: 10/19/2020 13:49

Field ID: G3CE0039

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2203506	EPA 180.1 Rev. 2.0	Turbidity	2.3		NTU	P388977	
	EPA 300.0 Rev. 2.1	Chloride	22		mg Cl/L	P389373	
		Sulfate	4.6		mg SO4/L	P389376	
		Color (true)	120		PCU	P388985	
	SM 2320 B-2011	Total Alkalinity	39		mg CaCO3/L	P389229	
	SM 2540 C-2011	TDS	112		mg/L	P389382	
	SM 2540 D-2011	TSS	3	I	mg/L	P389310	
	SM 4500 F-C-2011	Fluoride	0.17		mg F/L	P389118	
2203507	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P389208	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P389188	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	I	mg N/L	P389018	
	EPA 365.1 Rev. 2.0	Total-P	0.073		mg P/L	P389251	
	SM 5310 B-2011	Organic Carbon	15		mg C/L	P389069	
2203508	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.028		mg P/L	P388988	
2203511	EPA 200.7 Rev. 4.4	Barium	10.9		ug/L	P389029	
		Calcium	11.7		mg/L	P389029	
		Iron	140		ug/L	P389029	
		Magnesium	3.70		mg/L	P389029	
		Potassium	6.9		mg/L	P389029	
		Sodium	14.2		mg/L	P389029	
		Zinc	5.0	U	ug/L	P389029	
	EPA 200.8 Rev. 5.4	Aluminum	27		ug/L	P389029	
		Antimony	0.070	I	ug/L	P389029	
		Arsenic	1.21		ug/L	P389029	
		Beryllium	0.010	U	ug/L	P389029	
		Boron**	46.3		ug/L	P389029	
		Cadmium	0.020	U	ug/L	P389029	
		Chromium	0.40	U	ug/L	P389029	
		Copper	0.93		ug/L	P389029	
		Lead	0.20	U	ug/L	P389029	
		Nickel	0.50	U	ug/L	P389029	
		Selenium	0.20	U	ug/L	P389029	
		Silver	0.010	U	ug/L	P389029	
		Thallium	0.10	U	ug/L	P389029	
		Hardness	44.3		mg/L	P389029	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P388977

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P389029

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P389029

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389373

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389376

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389208

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389188

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389018

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389251

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P388988

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P388985

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P389229

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2011
Batch ID: P389382

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P389310

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P389118

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-2011
Batch ID: P389069

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P389029

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	103		P	85 - 115
Calcium	106		P	85 - 115
Iron	105		P	85 - 115
Magnesium	105		P	85 - 115
Potassium	103		P	85 - 115
Sodium	105		P	85 - 115
Zinc	100		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P389029

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	109		P	85 - 115
Antimony	103		P	85 - 115
Arsenic	101		P	85 - 115
Beryllium	101		P	85 - 115
Boron	104		P	85 - 115
Cadmium	102		P	85 - 115
Chromium	101		P	85 - 115
Copper	97.6		P	85 - 115
Lead	97.1		P	85 - 115
Nickel	98.5		P	85 - 115
Selenium	101		P	85 - 115
Silver	97.4		P	85 - 115
Thallium	98.5		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389373

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389376

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389208

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389188

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	107		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389018

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389251

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P388988

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.9		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P388985

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P389229

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P389118

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.3		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P389069

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.3		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P389029

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203379	Barium	105		P	80 - 120
2203379	Calcium	104		P	80 - 120
2203379	Iron	107		P	80 - 120
2203379	Magnesium	109		P	80 - 120
2203379	Potassium	103		P	80 - 120
2203379	Sodium	101		P	80 - 120
2203379	Zinc	102		P	80 - 120
2203607	Barium	108	104	P/P	80 - 120
2203607	Calcium	108	106	P/P	80 - 120
2203607	Iron	109	107	P/P	80 - 120
2203607	Magnesium	110	108	P/P	80 - 120
2203607	Potassium	103	102	P/P	80 - 120
2203607	Sodium	101		P	80 - 120
2203607	Zinc	106	103	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P389029

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203379	Aluminum	101		P	80 - 120
2203379	Antimony	99.2		P	80 - 120
2203379	Arsenic	101		P	80 - 120
2203379	Beryllium	99.8		P	80 - 120
2203379	Boron	102		P	80 - 120
2203379	Cadmium	102		P	80 - 120
2203379	Chromium	99.4		P	80 - 120
2203379	Copper	96.7		P	80 - 120
2203379	Lead	95.1		P	80 - 120
2203379	Nickel	97.4		P	80 - 120
2203379	Selenium	97.0		P	80 - 120
2203379	Silver	96.6		P	80 - 120
2203379	Thallium	98.0		P	80 - 120
2203607	Aluminum	103	104	P/P	80 - 120
2203607	Antimony	102	102	P/P	80 - 120
2203607	Arsenic	100	102	P/P	80 - 120
2203607	Beryllium	96.7	99.2	P/P	80 - 120
2203607	Boron	101	104	P/P	80 - 120
2203607	Cadmium	103	103	P/P	80 - 120
2203607	Chromium	99.7	98.7	P/P	80 - 120
2203607	Copper	97.3	97.6	P/P	80 - 120
2203607	Lead	95.9	96.2	P/P	80 - 120
2203607	Nickel	97.8	98.0	P/P	80 - 120
2203607	Selenium	94.3	95.7	P/P	80 - 120
2203607	Silver	95.7	96.9	P/P	80 - 120
2203607	Thallium	98.0	98.7	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389373

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203363	Chloride	106		P	90 - 110
2203407	Chloride	103		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389376

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203363	Sulfate	104		P	90 - 110
2203407	Sulfate	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389208

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203335	Ammonia-N	106	105	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389188

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203604	Kjeldahl Nitrogen	102	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389018

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203507	NO2NO3-N	98.9	99.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389251

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203408	Total-P	105	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P388988

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203388	O-Phosphate-P	98.9	98.7	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P389118

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203350	Fluoride	103	105	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P389069

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2203460	Organic Carbon	97.6	97.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P388977

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203584	Turbidity	1.30	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P389029

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203607	Barium	3.86	Spike	P	0 - 20
2203607	Calcium	1.37	Spike	P	0 - 20
2203607	Iron	1.27	Spike	P	0 - 20
2203607	Magnesium	1.50	Spike	P	0 - 20
2203607	Potassium	0.574	Spike	P	0 - 20
2203607	Sodium	0.0891	Spike	P	0 - 20
2203607	Zinc	3.26	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P389029

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203607	Aluminum	1.35	Spike	P	0 - 20
2203607	Antimony	0.260	Spike	P	0 - 20
2203607	Arsenic	1.49	Spike	P	0 - 20
2203607	Beryllium	2.12	Spike	P	0 - 20
2203607	Boron	2.74	Spike	P	0 - 20
2203607	Cadmium	0.0805	Spike	P	0 - 20
2203607	Chromium	0.975	Spike	P	0 - 20
2203607	Copper	0.263	Spike	P	0 - 20
2203607	Lead	0.347	Spike	P	0 - 20
2203607	Nickel	0.263	Spike	P	0 - 20
2203607	Selenium	1.45	Spike	P	0 - 20
2203607	Silver	1.29	Spike	P	0 - 20
2203607	Thallium	0.747	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389373

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203363	Chloride	0.487	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P389376

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203363	Sulfate	0.720	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P389208

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203335	Ammonia-N	0.800	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P389188

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203604	Kjeldahl Nitrogen	1.76	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P389018

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203507	NO2NO3-N	0.494	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P389251

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203408	Total-P	0.937	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P388988

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203388	O-Phosphate-P	0.178	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P388985

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203456	Color (true)	0.194	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P389229

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2204122	Total Alkalinity	1.10	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P389382

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203456	TDS	1.08	Sample	P	0 - 10

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-2011
Batch ID: P389118

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203350	Fluoride	2.01	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P389069

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2203460	Organic Carbon	0.116	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A101640

Included Lab Sample IDs: 2203506

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	100	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A101644

Included Lab Sample IDs: 2203506

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A101645

Included Lab Sample IDs: 2203508

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.9	99.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A101657

Included Lab Sample IDs: 2203507

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A101673

Included Lab Sample IDs: 2203507

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.6	98.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A101694

Included Lab Sample IDs: 2203506

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A101702

Included Lab Sample IDs: 2203511

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	100	100	P/P	90 - 110
Calcium	100	100	P/P	90 - 110
Magnesium	100	100	P/P	90 - 110
Potassium	98.4	97.5	P/P	90 - 110
Sodium	100	99.3	P/P	90 - 110
Zinc	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A101719

Included Lab Sample IDs: 2203511

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Iron	105	106	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A101740

Included Lab Sample IDs: 2203507

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	102	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A101749

Included Lab Sample IDs: 2203506

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.5	99.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A101768

Included Lab Sample IDs: 2203507

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	107	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A101777

Included Lab Sample IDs: 2203511

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Antimony	99.3	99.7	P/P	90 - 110
Arsenic	98.4	97.9	P/P	90 - 110
Beryllium	98.7	99.4	P/P	90 - 110
Boron	101	102	P/P	90 - 110
Cadmium	100	99.2	P/P	90 - 110
Chromium	96.2	95.9	P/P	90 - 110
Copper	94.6	95.7	P/P	90 - 110
Lead	95.4	95.3	P/P	90 - 110
Nickel	96.4	96.7	P/P	90 - 110
Selenium	99.4	98.7	P/P	90 - 110
Silver	99.1	99.2	P/P	90 - 110
Thallium	92.7	91.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101793

Included Lab Sample IDs: 2203507

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A101798

Included Lab Sample IDs: 2203506

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A101828

Included Lab Sample IDs: 2203511

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	99.7	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					1.30	
EPA 200.7 Rev. 4.4	Barium	103	105	108	104		3.86
	Calcium	106	104	108	106		1.37
	Iron	105	107	109	107		1.27
	Magnesium	105	109	110	108		1.50
	Potassium	103	103	103	102		0.574
	Sodium	105	101	101			0.0891
	Zinc	100	102	106	103		3.26
EPA 200.8 Rev. 5.4	Aluminum	109	103	104	101		1.35
	Antimony	103	102	102	99.2		0.260
	Arsenic	101	100	102	101		1.49
	Beryllium	101	96.7	99.2	99.8		2.12
	Boron	104	101	104	102		2.74
	Cadmium	102	103	103	102		0.0805
	Chromium	101	99.7	98.7	99.4		0.975
	Copper	97.6	97.3	97.6	96.7		0.263
	Lead	97.1	95.9	96.2	95.1		0.347
	Nickel	98.5	97.8	98.0	97.4		0.263
	Selenium	101	94.3	95.7	97.0		1.45
	Silver	97.4	95.7	96.9	96.6		1.29
	Thallium	98.5	98.0	98.7	98.0		0.747
EPA 300.0 Rev. 2.1	Chloride	102	106	103		0.487	
	Sulfate	103	104	103		0.720	
EPA 350.1 Rev. 2.0	Ammonia-N	102	106	105			0.800
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	107	102	100			1.76
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	98.9	99.4			0.494
EPA 365.1 Rev. 2.0	Total-P	103	105	106			0.937
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.9	98.9	98.7			0.178
SM 2120 B-2011	Color (true)	101				0.194	
SM 2320 B-2011	Total Alkalinity	101				1.10	
SM 2540 C-2011	TDS					1.08	
SM 4500 F-C-2011	Fluoride	99.3	103	105			2.01
SM 5310 B-2011	Organic Carbon	99.3	97.6	97.7			0.116

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2203506
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2203511
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2203511
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2203506
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2203506
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2203507
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2203507
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2203507
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2203507
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2203508
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2203506
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2203506
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2203511

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2203506
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2203506
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2203506
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2203507

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	10/20/2020			10/20/2020 16:02	Yen Ta	2203506
EPA 200.7 Rev. 4.4	10/20/2020	10/21/2020 15:25	Elliott D. Healy	10/22/2020 17:04	Betina Topolski	2203511
	10/20/2020	10/21/2020 15:25	Elliott D. Healy	10/23/2020 11:25	Betina Topolski	2203511
EPA 200.8 Rev. 5.4	10/20/2020	10/21/2020 15:25	Elliott D. Healy	10/26/2020 17:35	Alexander Thompson	2203511
	10/20/2020	10/21/2020 15:25	Elliott D. Healy	10/28/2020 16:26	Alexander Thompson	2203511
EPA 300.0 Rev. 2.1	10/20/2020			10/28/2020 02:30	Lipi Saha	2203506
EPA 350.1 Rev. 2.0	10/20/2020			10/25/2020 10:43	Ping Hua	2203507
EPA 351.2 Rev. 2.0	10/20/2020	10/23/2020 14:00	David Boose	10/26/2020 12:45	Elliott Rodriguez	2203507
EPA 353.2 Rev. 2.0	10/20/2020			10/21/2020 10:42	Beverly Y. Sanders	2203507
EPA 365.1 Rev. 2.0	10/20/2020	10/26/2020 13:20	Yen Ta	10/27/2020 12:29	Benjamin T. Nordmann	2203507
EPA 365.1 Rev. 2.0 dissolved	10/20/2020			10/20/2020 16:38	Samantha Davila	2203508
SM 2120 B-2011	10/20/2020			10/20/2020 17:22	Benjamin T. Nordmann	2203506
SM 2320 B-2011	10/20/2020			10/24/2020 01:48	Samantha Davila	2203506
SM 2340B	10/20/2020			10/22/2020 17:04	Betina Topolski	2203511
SM 2540 C-2011	10/20/2020			10/21/2020 15:19	Yijie Li	2203506
SM 2540 D-2011	10/20/2020			10/21/2020 15:19	Yijie Li	2203506
SM 4500 F-C-2011	10/20/2020			10/22/2020 11:20	Samantha Davila	2203506
SM 5310 B-2011	10/20/2020			10/22/2020 00:51	Lauren Lees	2203507