

Chemical Analysis Report

SO-DIST-2020-02-18-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **2020 SMP : Myakka**
Request ID: **RQ-2020-02-17-51**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Nicole Reistad

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 12-MAR-2020 10:54



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Horse Creek @ SR 72

Collection Date/Time: 02/17/2020 09:25

Field ID: G3SD0157

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158520	EPA 180.1 Rev. 2.0	Turbidity	2.5		NTU	P379138	
	EPA 300.0 Rev. 2.1	Chloride	26		mg Cl/L	P379718	
		Sulfate	240		mg SO4/L	P379817	
	SM 2120 B-2011	Color (true)	94		PCU	P379207	
	SM 2320 B-2011	Total Alkalinity	58		mg CaCO3/L	P379416	
	SM 2540 C-2011	TDS	435		mg/L	P379821	
	SM 2540 D-2011	TSS	3	I	mg/L	P379601	
	SM 4500 F-C-2011	Fluoride	0.47		mg F/L	P379421	
2158522	EPA 350.1 Rev. 2.0	Ammonia-N	0.026		mg N/L	P379461	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.91		mg N/L	P379313	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.27		mg N/L	P379608	
	EPA 365.1 Rev. 2.0	Total-P	0.38		mg P/L	P379776	
	SM 5310 B-2011	Organic Carbon	17		mg C/L	P379361	
2158524	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.30		mg P/L	P379144	
2158561	EPA 8321B	Aldicarb	0.020	U	ug/L	P379101	
		Aldicarb Sulfone	0.0020	U	ug/L	P379101	
		Aldicarb Sulfoxide	0.0020	U	ug/L	P379101	
		Carbaryl	0.0020	U	ug/L	P379101	
		Carbofuran	0.0020	U	ug/L	P379101	
		3-Hydroxycarbofuran	0.0020	U	ug/L	P379101	
		Methiocarb	0.0020	U	ug/L	P379101	
		Methomyl	0.0020	U	ug/L	P379101	
		Oxamyl	0.0020	U	ug/L	P379101	
		Propoxur	0.0020	U	ug/L	P379101	
		2,4-D	0.0042	I	ug/L	P379190	
		Acesulfame K**	0.10	U	ug/L	P379179	
		AMPA**	0.25	I	ug/L	P379179	
2158563		Sucralose**	0.010	U	ug/L	P379179	
		Acetaminophen**	0.0080	U	ug/L	P379190	
		Endothall**	0.25	U	ug/L	P379179	
		Acetamiprid**	0.0020	U	ug/L	P379190	RPD
		Glufosinate**	0.10	U	ug/L	P379179	MS
		Glyphosate**	0.10	U	ug/L	P379179	MS
		Afidopyropen**	0.0020	UJ	ug/L	P379190	
		Bentazon	0.27		ug/L	P379190	
		Benzovindiflupyr**	0.0020	UJ	ug/L	P379190	
		Carbamazepine**	8.0E-04	U	ug/L	P379190	
		Clothianidin**	0.019		ug/L	P379190	
		Dinotefuran**	0.014		ug/L	P379190	
		Diuron	0.0020	U	ug/L	P379190	
		Fenuron	0.016	U	ug/L	P379190	
		Fluridone**	8.0E-04	U	ug/L	P379190	
		Imazapyr**	0.028		ug/L	P379190	

Field ID: G3SD0157

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158563	EPA 8321B	Hydrocodone**	0.0020	U	ug/L	P379190	
		Ibuprofen**	0.020	U	ug/L	P379190	
		Imidacloprid	0.16		ug/L	P379190	
		Linuron	0.0040	U	ug/L	P379190	
		Mandestrobin**	0.0020	UJ	ug/L	P379190	
		MCP	0.0023	I	ug/L	P379190	
		Naproxen**	0.0080	U	ug/L	P379190	
		Primidone**	0.0040	U	ug/L	P379190	
		Pyraclostrobin**	0.0020	U	ug/L	P379190	
		Thiamethoxam**	0.047		ug/L	P379190	
		Tolfenpyrad**	0.0020	U	ug/L	P379190	
		Triclopyr**	0.0040	U	ug/L	P379190	
2158573	EPA 8270E	Aldrin	4.3	U	ng/L	P379288	
		Alpha-BHC	2.2	U	ng/L	P379288	
		Alpha-Chlordane	1.1	U	ng/L	P379288	
		Beta-BHC	11	U	ng/L	P379288	
		Bifenthrin**	4.3	U	ng/L	P379288	
		Delta-BHC	8.7	U	ng/L	P379288	
		Gamma-BHC	8.7	U	ng/L	P379288	
		Carbophenothion	11	U	ng/L	P379288	
		Chlorothalonil	2.2	UJ	ng/L	P379288	RPD
		Cis-Nonachlor**	1.1	U	ng/L	P379288	
		Cypermethrin	22	U	ng/L	P379288	
		DDD-p,p'	6.5	U	ng/L	P379288	
		DDE-p,p'	6.5	U	ng/L	P379288	
		DDT-p,p'	7.6	U	ng/L	P379288	
		Dicofol	33	U	ng/L	P379288	
		Dieldrin	4.3	U	ng/L	P379288	
		Endosulfan I	4.3	U	ng/L	P379288	
		Endosulfan II	4.3	U	ng/L	P379288	
		Endosulfan Sulfate	2.2	U	ng/L	P379288	
		Endrin	2.2	U	ng/L	P379288	
		Endrin Aldehyde	2.2	U	ng/L	P379288	
		Endrin Ketone	2.2	U	ng/L	P379288	
		Ethalfuralin**	3.3	U	ng/L	P379288	
		Gamma-Chlordane	1.1	U	ng/L	P379288	
		Heptachlor	1.6	U	ng/L	P379288	
		Heptachlor Epoxide	2.2	U	ng/L	P379288	
		Hexachlorobenzene	2.2	U	ng/L	P379288	
		Methoxychlor	4.3	U	ng/L	P379288	
		Mirex	2.2	U	ng/L	P379288	
		Permethrin	11	U	ng/L	P379288	
		Trans-Nonachlor**	1.1	U	ng/L	P379288	
		Trifluralin	2.7	U	ng/L	P379288	
	EPA 8276	Chlordane	5.4	UJ	ng/L	P379288	LCS, RPD

Field ID: G3SD0157

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158573	EPA 8276	Toxaphene	33	U	ng/L	P379288	
2158575	EPA 8270E	Acetochlor	0.24	U	ng/L	P379289	
		Alachlor	0.24	U	ng/L	P379289	
		Ametryn	3.9		ng/L	P379289	
		Atrazine	43		ng/L	P379289	
		Atrazine Desethyl	8.0		ng/L	P379289	RPD
		Azinphos Methyl	1.9	U	ng/L	P379289	
		Bromacil	150		ng/L	P379289	
		Butylate	0.39	U	ng/L	P379289	RPD
		Carbophenothion	0.48	U	ng/L	P379289	
		Chlorpyrifos Ethyl	0.24	U	ng/L	P379289	
		Chlorpyrifos Methyl	0.048	U	ng/L	P379289	
		Cyanazine	3.1	U	ng/L	P379289	
		Demeton	1.5	UJ	ng/L	P379289	RPD
		Diazinon	0.12	U	ng/L	P379289	
		Dimethenamid**	0.097	U	ng/L	P379289	
		Disulfoton	0.48	U	ng/L	P379289	
		Dithiopyr**	0.18	I	ng/L	P379289	
		EPTC	1.2	U	ng/L	P379289	RPD
		Ethion	0.15	U	ng/L	P379289	
		Ethoprop	0.097	U	ng/L	P379289	
		Fenamiphos	0.48	U	ng/L	P379289	
		Fipronil	0.24	U	ng/L	P379289	
		Fipronil Desulfinyl**	0.12	U	ng/L	P379289	
		Fipronil Sulfide	0.15	U	ng/L	P379289	
		Fipronil Sulfone	0.15	U	ng/L	P379289	
		Flumioxazin**	1.7	U	ng/L	P379289	
		Fonofos	0.19	U	ng/L	P379289	
		Hexazinone	23		ng/L	P379289	
		Imazalil**	0.73	U	ng/L	P379289	MS, RPD
		Malathion	0.34	U	ng/L	P379289	
		Metalaxyl	43		ng/L	P379289	
		Metolachlor	140		ng/L	P379289	
		Metribuzin	4.7	I	ng/L	P379289	
		Mevinphos	0.48	U	ng/L	P379289	
		Molinate	0.29	U	ng/L	P379289	
		Norflurazon	84		ng/L	P379289	
		Oxadiazon**	1.9		ng/L	P379289	
		Parathion Ethyl	0.24	U	ng/L	P379289	
		Parathion Methyl	0.53	U	ng/L	P379289	
		Pendimethalin	0.97	U	ng/L	P379289	
		Phorate	0.24	U	ng/L	P379289	
		Prodiamine**	0.17	U	ng/L	P379289	RPD
		Prometon	0.87	U	ng/L	P379289	
		Prometryn	0.19	U	ng/L	P379289	

Field ID: G3SD0157

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158575	EPA 8270E	Pronamide**	0.15	U	ng/L	P379289	
		Propiconazole**	6.0		ng/L	P379289	
		Simazine	82		ng/L	P379289	
		Sulfentrazone**	2.1	J	ng/L	P379289	RPD
		Tebuconazole**	0.48	UJ	ng/L	P379289	RPD
		Terbufos	0.097	U	ng/L	P379289	
		Terbuthylazine	0.41	U	ng/L	P379289	
2158579	EPA 200.7 Rev. 4.4	Barium	14.4		ug/L	P379279	
		Calcium	78.4		mg/L	P379279	
		Iron	170		ug/L	P379279	
		Magnesium	33.3		mg/L	P379279	
		Potassium	8.3		mg/L	P379279	
		Sodium	17.9		mg/L	P379279	
		Zinc	5.0	U	ug/L	P379279	
	EPA 200.8 Rev. 5.4	Aluminum	67		ug/L	P379279	
		Antimony	0.24		ug/L	P379279	
		Arsenic	0.73		ug/L	P379279	
		Boron**	34.3		ug/L	P379279	
		Cadmium	0.045	I	ug/L	P379279	
		Chromium	0.40	U	ug/L	P379279	
		Copper	0.40	U	ug/L	P379279	
		Lead	0.20	U	ug/L	P379279	
		Nickel	0.87	I	ug/L	P379279	
		Selenium	0.20	U	ug/L	P379279	
		Silver	0.010	U	ug/L	P379279	
		Thallium	0.10	U	ug/L	P379279	
		Hardness	333		mg/L	P379279	
	SM 2340B						

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

EPA 8321B: J - second source check standards for some analytes were not available for analysis. MS accuracy for bentazon and imazapyr could not be assessed due to high concentration of parameter in the spiked sample.

EPA 8270E: Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8276: Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8270E: MS accuracy for some parameters not assessed due to high content of these parameters in the sample spiked. Refer to the Lab Analysis Report for an explanation of QC Codes.

Sample Location: Brandy Branch @ SR 70

Collection Date/Time: 02/17/2020 10:05

Field ID: G3SD0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158526	EPA 180.1 Rev. 2.0	Turbidity	0.90		NTU	P379138	
	EPA 300.0 Rev. 2.1	Chloride	31		mg Cl/L	P379718	
		Sulfate	710		mg SO4/L	P379721	
	SM 2120 B-2011	Color (true)	45	A	PCU	P379207	
	SM 2320 B-2011	Total Alkalinity	56		mg CaCO3/L	P379416	
	SM 2540 C-2011	TDS	1.12E+03		mg/L	P379821	
	SM 2540 D-2011	TSS	2	I	mg/L	P379601	
	SM 4500 F-C-2011	Fluoride	0.62		mg F/L	P379421	
2158527	EPA 350.1 Rev. 2.0	Ammonia-N	0.028		mg N/L	P379461	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.77		mg N/L	P379313	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.44		mg N/L	P379608	
	EPA 365.1 Rev. 2.0	Total-P	0.25		mg P/L	P379776	
	SM 5310 B-2011	Organic Carbon	11		mg C/L	P379361	
2158528	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.21		mg P/L	P379144	
2158565	EPA 8321B	Aldicarb	0.020	U	ug/L	P379101	
		Aldicarb Sulfone	0.0020	U	ug/L	P379101	
		Aldicarb Sulfoxide	0.0020	U	ug/L	P379101	
		Carbaryl	0.0020	U	ug/L	P379101	
		Carbofuran	0.0020	U	ug/L	P379101	
		3-Hydroxycarbofuran	0.0020	U	ug/L	P379101	
		Methiocarb	0.0020	U	ug/L	P379101	
		Methomyl	0.0020	U	ug/L	P379101	
		Oxamyl	0.0048	I	ug/L	P379101	
		Propoxur	0.0020	U	ug/L	P379101	
		2,4-D	0.0064	I	ug/L	P379190	
		Acesulfame K**	0.10	U	ug/L	P379179	
		AMPA**	1.0	U	ug/L	P379179	
		Sucralose**	0.010	U	ug/L	P379179	
2158566		Acetaminophen**	0.0080	U	ug/L	P379190	
		Endothall**	0.25	U	ug/L	P379179	
		Acetamiprid**	0.0020	U	ug/L	P379190	RPD
		Glufosinate**	1.0	U	ug/L	P379179	MS
		Glyphosate**	1.0	U	ug/L	P379179	MS
		Afidopyropen**	0.0020	UJ	ug/L	P379190	
		Bentazon	0.93		ug/L	P379190	
		Benzovindiflupyr**	0.0020	UJ	ug/L	P379190	
		Carbamazepine**	8.0E-04	U	ug/L	P379190	
		Clothianidin**	0.035		ug/L	P379190	
		Dinotefuran**	0.0020	U	ug/L	P379190	
		Diuron	0.0051	I	ug/L	P379190	
		Fenuron	0.016	U	ug/L	P379190	
		Fluridone**	8.0E-04	U	ug/L	P379190	
		Imazapyr**	0.0055	I	ug/L	P379190	

Field ID: G3SD0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158566	EPA 8321B	Hydrocodone**	0.0020	U	ug/L	P379190	
		Ibuprofen**	0.020	U	ug/L	P379190	
		Imidacloprid	0.33		ug/L	P379190	
		Linuron	0.0040	U	ug/L	P379190	
		Mandestrobin**	0.0020	UJ	ug/L	P379190	
		MCP	0.0020	U	ug/L	P379190	
		Naproxen**	0.0080	U	ug/L	P379190	
		Primidone**	0.0040	U	ug/L	P379190	
		Pyraclostrobin**	0.0020	U	ug/L	P379190	
		Thiamethoxam**	0.083		ug/L	P379190	
		Tolfenpyrad**	0.0020	U	ug/L	P379190	
		Triclopyr**	0.0040	U	ug/L	P379190	
2158577	EPA 8270E	Aldrin	4.4	U	ng/L	P379288	
		Alpha-BHC	2.2	U	ng/L	P379288	
		Alpha-Chlordane	1.1	U	ng/L	P379288	
		Beta-BHC	11	U	ng/L	P379288	
		Bifenthrin**	4.4	U	ng/L	P379288	
		Delta-BHC	8.8	U	ng/L	P379288	
		Gamma-BHC	8.8	U	ng/L	P379288	
		Carbophenothion	11	U	ng/L	P379288	
		Chlorothalonil	2.2	UJ	ng/L	P379288	RPD
		Cis-Nonachlor**	1.1	U	ng/L	P379288	
		Cypermethrin	22	U	ng/L	P379288	
		DDD-p,p'	6.6	U	ng/L	P379288	
		DDE-p,p'	6.6	U	ng/L	P379288	
		DDT-p,p'	7.7	U	ng/L	P379288	
		Dicofol	33	U	ng/L	P379288	
		Dieldrin	4.4	U	ng/L	P379288	
		Endosulfan I	4.4	U	ng/L	P379288	
		Endosulfan II	4.4	U	ng/L	P379288	
		Endosulfan Sulfate	2.2	U	ng/L	P379288	
		Endrin	2.2	U	ng/L	P379288	
		Endrin Aldehyde	2.2	U	ng/L	P379288	
		Endrin Ketone	2.2	U	ng/L	P379288	
		Ethalfuralin**	3.3	U	ng/L	P379288	
		Gamma-Chlordane	1.1	U	ng/L	P379288	
		Heptachlor	1.7	U	ng/L	P379288	
		Heptachlor Epoxide	2.2	U	ng/L	P379288	
		Hexachlorobenzene	2.2	U	ng/L	P379288	
		Methoxychlor	4.4	U	ng/L	P379288	
		Mirex	2.2	U	ng/L	P379288	
		Permethrin	11	U	ng/L	P379288	
		Trans-Nonachlor**	1.1	U	ng/L	P379288	
		Trifluralin	2.8	U	ng/L	P379288	
	EPA 8276	Chlordane	5.5	UJ	ng/L	P379288	LCS, RPD

Field ID: G3SD0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158577	EPA 8276	Toxaphene	33	UJ	ng/L	P379288	CCV
2158578	EPA 8270E	Acetochlor	0.26	U	ng/L	P379289	
		Alachlor	0.26	U	ng/L	P379289	
		Ametryn	30		ng/L	P379289	
		Atrazine	41		ng/L	P379289	
		Atrazine Desethyl	17		ng/L	P379289	RPD
		Azinphos Methyl	2.0	U	ng/L	P379289	
		Bromacil	300		ng/L	P379289	
		Butylate	0.41	U	ng/L	P379289	RPD
		Carbophenothion	0.51	U	ng/L	P379289	
		Chlorpyrifos Ethyl	0.26	U	ng/L	P379289	
		Chlorpyrifos Methyl	0.051	U	ng/L	P379289	
		Cyanazine	3.3	U	ng/L	P379289	
		Demeton	1.5	UJ	ng/L	P379289	RPD
		Diazinon	0.13	U	ng/L	P379289	
		Dimethenamid**	0.10	U	ng/L	P379289	
		Disulfoton	0.51	U	ng/L	P379289	
		Dithiopyr**	0.10	U	ng/L	P379289	
		EPTC	1.3	U	ng/L	P379289	RPD
		Ethion	0.15	U	ng/L	P379289	
		Ethoprop	0.10	U	ng/L	P379289	
		Fenamiphos	0.51	U	ng/L	P379289	
		Fipronil	0.26	U	ng/L	P379289	
		Fipronil Desulfinyl**	0.13	U	ng/L	P379289	
		Fipronil Sulfide	0.15	U	ng/L	P379289	
		Fipronil Sulfone	0.15	U	ng/L	P379289	
		Flumioxazin**	1.8	U	ng/L	P379289	
		Fonofos	0.20	U	ng/L	P379289	
		Hexazinone	0.64	I	ng/L	P379289	
		Imazalil**	0.77	U	ng/L	P379289	MS, RPD
		Malathion	0.36	U	ng/L	P379289	
		Metalaxyl	17		ng/L	P379289	
		Metolachlor	380		ng/L	P379289	
		Metribuzin	3.3	U	ng/L	P379289	
		Mevinphos	0.51	U	ng/L	P379289	
		Molinate	0.31	U	ng/L	P379289	
		Norflurazon	130		ng/L	P379289	
		Oxadiazon**	6.1		ng/L	P379289	
		Parathion Ethyl	0.26	U	ng/L	P379289	
		Parathion Methyl	0.56	U	ng/L	P379289	
		Pendimethalin	1.0	U	ng/L	P379289	
		Phorate	0.26	U	ng/L	P379289	
		Prodiamine**	0.18	U	ng/L	P379289	RPD
		Prometon	0.92	U	ng/L	P379289	
		Prometryn	0.38	I	ng/L	P379289	

Field ID: G3SD0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158578	EPA 8270E	Pronamide**	0.15	U	ng/L	P379289	
		Propiconazole**	16		ng/L	P379289	
		Simazine	410		ng/L	P379289	
		Sulfentrazone**	14	J	ng/L	P379289	RPD
		Tebuconazole**	0.51	UJ	ng/L	P379289	RPD
		Terbufos	0.10	U	ng/L	P379289	
		Terbuthylazine	0.43	U	ng/L	P379289	
2158581	EPA 200.7 Rev. 4.4	Barium	23.0		ug/L	P379279	
		Calcium	204		mg/L	P379279	
		Iron	110	I	ug/L	P379279	
		Magnesium	84.4		mg/L	P379279	
		Potassium	16.0		mg/L	P379279	
		Sodium	16.4		mg/L	P379279	
		Zinc	5.0	U	ug/L	P379279	
	EPA 200.8 Rev. 5.4	Aluminum	53		ug/L	P379279	
		Antimony	0.25		ug/L	P379279	
		Arsenic	0.67		ug/L	P379279	
		Boron**	56.3		ug/L	P379279	
		Cadmium	0.054	I	ug/L	P379279	
		Chromium	0.40	U	ug/L	P379279	
		Copper	0.40	U	ug/L	P379279	
		Lead	0.20	U	ug/L	P379279	
		Nickel	1.1	I	ug/L	P379279	
		Selenium	0.20	U	ug/L	P379279	
		Silver	0.010	U	ug/L	P379279	
		Thallium	0.10	U	ug/L	P379279	
		Hardness	857		mg/L	P379279	
	SM 2340B						

Ref. Method and Comment:

EPA 8321B: J - second source check standards for some analytes were not available for analysis. MS accuracy for bentazon and imazapyr could not be assessed due to high concentration of parameter in the spiked sample.

EPA 8321B: MDL for some parameters elevated due to matrix interference.

EPA 8270E: Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8276: Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8270E: MS accuracy for some parameters not assessed due to high content of these parameters in the sample spiked. Refer to the Lab Analysis Report for an explanation of QC Codes.

Sample Location: Lake Myakka (Upper)

Collection Date/Time: 02/17/2020 11:30

Field ID: G3SD0133

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158521	EPA 180.1 Rev. 2.0	Turbidity	10	A	NTU	P379138	
	EPA 300.0 Rev. 2.1	Chloride	27		mg Cl/L	P379718	
		Sulfate	240		mg SO4/L	P379817	
	SM 2120 B-2011	Color (true)	68		PCU	P379207	
	SM 2320 B-2011	Total Alkalinity	37		mg CaCO3/L	P379416	
	SM 2540 C-2011	TDS	482	A	mg/L	P379821	
	SM 2540 D-2011	TSS	10	A	mg/L	P379601	
	SM 4500 F-C-2011	Fluoride	0.25		mg F/L	P379421	
2158523	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P379461	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.88		mg N/L	P379313	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P379608	
	EPA 365.1 Rev. 2.0	Total-P	0.24		mg P/L	P379776	
	SM 5310 B-2011	Organic Carbon	14		mg C/L	P379361	
2158525	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.13		mg P/L	P379144	
2158562	EPA 8321B	Aldicarb	0.020	U	ug/L	P379101	
		Aldicarb Sulfone	0.0020	U	ug/L	P379101	
		Aldicarb Sulfoxide	0.0020	U	ug/L	P379101	
		Carbaryl	0.0020	U	ug/L	P379101	
		Carbofuran	0.0020	U	ug/L	P379101	
		3-Hydroxycarbofuran	0.0020	U	ug/L	P379101	
		Methiocarb	0.0020	U	ug/L	P379101	
		Methomyl	0.0020	U	ug/L	P379101	
		Oxamyl	0.0020	U	ug/L	P379101	
		Propoxur	0.0020	U	ug/L	P379101	
		2,4-D	0.0020	U	ug/L	P379190	
		Acesulfame K**	0.10	U	ug/L	P379179	
		AMPA**	0.45		ug/L	P379179	
		Sucralose**	0.18		ug/L	P379179	
2158564		Acetaminophen**	0.0080	U	ug/L	P379190	
		Endothall**	0.25	U	ug/L	P379179	
		Acetamiprid**	0.0020	U	ug/L	P379190	RPD
		Glufosinate**	0.10	U	ug/L	P379179	MS
		Glyphosate**	0.28	I	ug/L	P379179	MS
		Afidopyropen**	0.0020	UJ	ug/L	P379190	
		Bentazon	0.0030	I	ug/L	P379190	
		Benzovindiflupyr**	0.0020	UJ	ug/L	P379190	
		Carbamazepine**	8.2E-04	I	ug/L	P379190	
		Clothianidin**	0.0034	I	ug/L	P379190	
		Dinotefuran**	0.065		ug/L	P379190	
		Diuron	0.0020	U	ug/L	P379190	
		Fenuron	0.016	U	ug/L	P379190	
		Fluridone**	0.0036		ug/L	P379190	
		Imazapyr**	0.011	I	ug/L	P379190	

Field ID: G3SD0133

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158564	EPA 8321B	Hydrocodone**	0.0020	U	ug/L	P379190	
		Ibuprofen**	0.020	U	ug/L	P379190	
		Imidacloprid	0.030		ug/L	P379190	
		Linuron	0.0040	U	ug/L	P379190	
		Mandestrobin**	0.0020	UJ	ug/L	P379190	
		MCP	0.0020	U	ug/L	P379190	
		Naproxen**	0.0080	U	ug/L	P379190	
		Primidone**	0.0040	U	ug/L	P379190	
		Pyraclostrobin**	0.0020	U	ug/L	P379190	
		Thiamethoxam**	0.014		ug/L	P379190	
		Tolfenpyrad**	0.0020	U	ug/L	P379190	
		Triclopyr**	0.0040	U	ug/L	P379190	
2158574	EPA 8270E	Aldrin	4.4	U	ng/L	P379288	
		Alpha-BHC	2.2	U	ng/L	P379288	
		Alpha-Chlordane	1.1	U	ng/L	P379288	
		Beta-BHC	11	U	ng/L	P379288	
		Bifenthrin**	4.4	U	ng/L	P379288	
		Delta-BHC	8.7	U	ng/L	P379288	
		Gamma-BHC	8.7	U	ng/L	P379288	
		Carbophenothion	11	U	ng/L	P379288	
		Chlorothalonil	2.2	UJ	ng/L	P379288	RPD
		Cis-Nonachlor**	1.1	U	ng/L	P379288	
		Cypermethrin	22	U	ng/L	P379288	
		DDD-p,p'	6.5	U	ng/L	P379288	
		DDE-p,p'	6.5	U	ng/L	P379288	
		DDT-p,p'	7.6	U	ng/L	P379288	
		Dicofol	33	U	ng/L	P379288	
		Dieldrin	4.4	U	ng/L	P379288	
		Endosulfan I	4.4	U	ng/L	P379288	
		Endosulfan II	4.4	U	ng/L	P379288	
		Endosulfan Sulfate	2.2	U	ng/L	P379288	
		Endrin	2.2	U	ng/L	P379288	
		Endrin Aldehyde	2.2	U	ng/L	P379288	
		Endrin Ketone	2.2	U	ng/L	P379288	
		Ethalfuralin**	3.3	U	ng/L	P379288	
		Gamma-Chlordane	1.1	U	ng/L	P379288	
		Heptachlor	1.6	U	ng/L	P379288	
		Heptachlor Epoxide	2.2	U	ng/L	P379288	
		Hexachlorobenzene	2.2	U	ng/L	P379288	
		Methoxychlor	4.4	U	ng/L	P379288	
		Mirex	2.2	U	ng/L	P379288	
		Permethrin	11	U	ng/L	P379288	
		Trans-Nonachlor**	1.1	U	ng/L	P379288	
		Trifluralin	2.7	U	ng/L	P379288	
	EPA 8276	Chlordane	5.4	UJ	ng/L	P379288	LCS, RPD

Field ID: G3SD0133

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158574	EPA 8276	Toxaphene	33	U	ng/L	P379288	
2158576	EPA 8270E	Acetochlor	0.24	U	ng/L	P379289	
		Alachlor	0.24	U	ng/L	P379289	
		Ametryn	0.69	I	ng/L	P379289	
		Atrazine	4.7		ng/L	P379289	
		Atrazine Desethyl	1.4	UJ	ng/L	P379289	RPD
		Azinphos Methyl	4.8	U	ng/L	P379289	
		Bromacil	4.7		ng/L	P379289	
		Butylate	0.39	UJ	ng/L	P379289	RPD
		Carbophenothion	0.48	U	ng/L	P379289	
		Chlorpyrifos Ethyl	0.24	U	ng/L	P379289	
		Chlorpyrifos Methyl	0.048	U	ng/L	P379289	
		Cyanazine	3.1	U	ng/L	P379289	
		Demeton	6.8	UJ	ng/L	P379289	RPD
		Diazinon	0.12	U	ng/L	P379289	
		Dimethenamid**	0.097	U	ng/L	P379289	
		Disulfoton	0.48	U	ng/L	P379289	
		Dithiopyr**	0.19	I	ng/L	P379289	
		EPTC	1.2	UJ	ng/L	P379289	RPD
		Ethion	0.14	U	ng/L	P379289	
		Ethoprop	0.097	U	ng/L	P379289	
		Fenamiphos	0.48	U	ng/L	P379289	
		Fipronil	0.24	U	ng/L	P379289	
		Fipronil Desulfinyl**	0.12	U	ng/L	P379289	
		Fipronil Sulfide	0.14	U	ng/L	P379289	
		Fipronil Sulfone	0.26	I	ng/L	P379289	
		Flumioxazin**	1.7	U	ng/L	P379289	
		Fonofos	0.19	U	ng/L	P379289	
		Hexazinone	40		ng/L	P379289	
		Imazalil**	0.72	UJ	ng/L	P379289	MS, RPD
		Malathion	1.1	I	ng/L	P379289	
		Metalaxyl	1.3	I	ng/L	P379289	
		Metolachlor	31		ng/L	P379289	
		Metribuzin	64		ng/L	P379289	
		Mevinphos	0.48	U	ng/L	P379289	
		Molinate	0.29	U	ng/L	P379289	
		Norflurazon	5.8		ng/L	P379289	
		Oxadiazon**	0.72		ng/L	P379289	
		Parathion Ethyl	0.24	U	ng/L	P379289	
		Parathion Methyl	0.53	U	ng/L	P379289	
		Pendimethalin	0.97	U	ng/L	P379289	
		Phorate	0.24	U	ng/L	P379289	
		Prodiamine**	0.17	UJ	ng/L	P379289	RPD
		Prometon	0.87	U	ng/L	P379289	
		Prometryn	0.19	U	ng/L	P379289	

Field ID: G3SD0133

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158576	EPA 8270E	Pronamide**	0.14	U	ng/L	P379289	
		Propiconazole**	0.48	U	ng/L	P379289	
		Simazine	0.48	U	ng/L	P379289	
		Sulfentrazone**	0.48	UJ	ng/L	P379289	RPD
		Tebuconazole**	0.90	IJ	ng/L	P379289	RPD
		Terbufos	0.097	U	ng/L	P379289	
		Terbuthylazine	0.41	U	ng/L	P379289	
2158580	EPA 200.7 Rev. 4.4	Barium	19.3		ug/L	P379279	
		Calcium	63.5		mg/L	P379279	
		Iron	230		ug/L	P379279	
		Magnesium	31.2		mg/L	P379279	
		Potassium	9.1		mg/L	P379279	
		Sodium	18.1		mg/L	P379279	
		Zinc	5.0	U	ug/L	P379279	
	EPA 200.8 Rev. 5.4	Aluminum	158		ug/L	P379279	
		Antimony	0.058	I	ug/L	P379279	
		Arsenic	0.82		ug/L	P379279	
		Boron**	33.1		ug/L	P379279	
		Cadmium	0.020	U	ug/L	P379279	
		Chromium	0.40	U	ug/L	P379279	
		Copper	0.90		ug/L	P379279	
		Lead	0.25	I	ug/L	P379279	
		Nickel	0.50	U	ug/L	P379279	
		Selenium	0.20	U	ug/L	P379279	
		Silver	0.010	U	ug/L	P379279	
		Thallium	0.10	U	ug/L	P379279	
		Hardness	287		mg/L	P379279	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

EPA 8321B: J - second source check standards for some analytes were not available for analysis. MS accuracy for bentazon and imazapyr could not be assessed due to high concentration of parameter in the spiked sample.

EPA 8270E: Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8276: Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8270E: MS accuracy for some parameters not assessed due to high content of these parameters in the sample spiked. Refer to the Lab Analysis Report for an explanation of QC Codes.

Sample Location: Lake Myakka (Lower)

Collection Date/Time: 02/17/2020 12:55

Field ID: G3SD0107

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158529	EPA 180.1 Rev. 2.0	Turbidity	5.0		NTU	P379139	
	EPA 300.0 Rev. 2.1	Chloride	28		mg Cl/L	P379718	
		Sulfate	230		mg SO4/L	P379817	
	SM 2120 B-2011	Color (true)	68	A	PCU	P379208	
	SM 2320 B-2011	Total Alkalinity	31		mg CaCO3/L	P379416	
	SM 2540 C-2011	TDS	440		mg/L	P379821	
	SM 2540 D-2011	TSS	4	I	mg/L	P379601	
	SM 4500 F-C-2011	Fluoride	0.26		mg F/L	P379421	
2158530	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P379461	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.91		mg N/L	P379313	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.042		mg N/L	P379608	
	EPA 365.1 Rev. 2.0	Total-P	0.22		mg P/L	P379776	
	SM 5310 B-2011	Organic Carbon	14		mg C/L	P379361	
2158531	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.15		mg P/L	P379144	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

Sample Location: Howard Creek NE Branch

Collection Date/Time: 02/17/2020 13:50

Field ID: G3SD0158

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2158532	EPA 180.1 Rev. 2.0	Turbidity	14		NTU	P379138	
	EPA 300.0 Rev. 2.1	Chloride	45		mg Cl/L	P379718	
		Sulfate	20		mg SO4/L	P379721	
		Color (true)	160		PCU	P379207	
	SM 2320 B-2011	Total Alkalinity	72		mg CaCO3/L	P379416	
	SM 2540 C-2011	TDS	220		mg/L	P379821	
	SM 2540 D-2011	TSS	4	I	mg/L	P379601	
	SM 4500 F-C-2011	Fluoride	0.53		mg F/L	P379421	
2158533	EPA 350.1 Rev. 2.0	Ammonia-N	0.023		mg N/L	P379461	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.81		mg N/L	P379313	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.076		mg N/L	P379608	
	EPA 365.1 Rev. 2.0	Total-P	0.80		mg P/L	P379776	
	SM 5310 B-2011	Organic Carbon	17		mg C/L	P379361	
2158534	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.58		mg P/L	P379146	
2158582	EPA 200.7 Rev. 4.4	Barium	4.3		ug/L	P379279	
		Calcium	27.7		mg/L	P379279	
		Iron	3.72E+03		ug/L	P379279	
		Magnesium	10.0		mg/L	P379279	
		Potassium	2.3		mg/L	P379279	
		Sodium	27.7		mg/L	P379279	
		Zinc	5.0	U	ug/L	P379279	
	EPA 200.8 Rev. 5.4	Aluminum	67		ug/L	P379279	
		Antimony	0.091	I	ug/L	P379279	
		Arsenic	2.37		ug/L	P379279	
		Boron**	36.7		ug/L	P379279	
		Cadmium	0.020	U	ug/L	P379279	
		Chromium	1.0	I	ug/L	P379279	
		Copper	0.40	U	ug/L	P379279	
		Lead	0.20	U	ug/L	P379279	
		Nickel	0.92	I	ug/L	P379279	
		Selenium	0.20	U	ug/L	P379279	
		Silver	0.010	U	ug/L	P379279	
		Thallium	0.10	U	ug/L	P379279	
	SM 2340B	Hardness	110		mg/L	P379279	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P379138

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P379139

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P379279

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P379279

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379718

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379721

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379817

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P379461

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P379313

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P379608

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P379776

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379144

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379146

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270E
Batch ID: P379288

Component	Result	Code	Units
Aldrin	4.0	U	ng/L
Alpha-BHC	2.0	U	ng/L
Alpha-Chlordane	1.0	U	ng/L
Beta-BHC	10	U	ng/L
Bifenthrin	4.0	U	ng/L
Carbophenothion	10	U	ng/L
Chlorothalonil	2.0	U	ng/L
Cis-Nonachlor	1.0	U	ng/L
Cypermethrin	20	U	ng/L
DDD-p,p'	6.0	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E
Batch ID: P379288

Component	Result	Code	Units
DDE-p,p'	6.0	U	ng/L
DDT-p,p'	7.0	U	ng/L
Delta-BHC	8.0	U	ng/L
Dicofol	30	U	ng/L
Dieldrin	4.0	U	ng/L
Endosulfan I	4.0	U	ng/L
Endosulfan II	4.0	U	ng/L
Endosulfan Sulfate	2.0	U	ng/L
Endrin	2.0	U	ng/L
Endrin Aldehyde	2.0	U	ng/L
Endrin Ketone	2.0	U	ng/L
Ethalfuralin	3.0	U	ng/L
Gamma-BHC	8.0	U	ng/L
Gamma-Chlordane	1.0	U	ng/L
Heptachlor	1.5	U	ng/L
Heptachlor Epoxide	2.0	U	ng/L
Hexachlorobenzene	2.0	U	ng/L
Methoxychlor	4.0	U	ng/L
Mirex	2.0	U	ng/L
Permethrin	10	U	ng/L
Trans-Nonachlor	1.0	U	ng/L
Trifluralin	2.5	U	ng/L

Reference Method: EPA 8270E
Batch ID: P379289

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.60	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	2.0	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Carbophenothion	0.50	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.050	U	ng/L
Cyanazine	3.2	U	ng/L
Demeton	1.5	U	ng/L
Diazinon	0.12	U	ng/L
Dimethenamid	0.10	U	ng/L
Disulfoton	0.50	U	ng/L
Dithiopyr	0.10	U	ng/L
EPTC	1.2	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Fenamiphos	0.50	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Desulfinyl	0.12	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E
 Batch ID: P379289

Component	Result	Code	Units
Flumioxazin	1.8	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Imazalil	0.75	U	ng/L
Malathion	0.35	U	ng/L
Metalaxyl	0.75	U	ng/L
Metolachlor	0.50	U	ng/L
Metribuzin	3.2	U	ng/L
Mevinphos	0.50	U	ng/L
Molinate	0.30	U	ng/L
Norflurazon	0.50	U	ng/L
Oxadiazon	0.10	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.55	U	ng/L
Pendimethalin	1.0	U	ng/L
Phorate	0.25	U	ng/L
Prodiamine	0.18	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Pronamide	0.15	U	ng/L
Propiconazole	0.50	U	ng/L
Simazine	0.50	U	ng/L
Sulfentrazone	0.50	U	ng/L
Tebuconazole	0.50	U	ng/L
Terbufos	0.10	U	ng/L
Terbutylazine	0.42	U	ng/L

Reference Method: EPA 8276
 Batch ID: P379288

Component	Result	Code	Units
Chlordane	5.0	U	ng/L
Toxaphene	30	U	ng/L

Reference Method: EPA 8321B
 Batch ID: P379101

Component	Result	Code	Units
3-Hydroxycarbofuran	0.0020	U	ug/L
Aldicarb	0.020	U	ug/L
Aldicarb Sulfone	0.0020	U	ug/L
Aldicarb Sulfoxide	0.0020	U	ug/L
Carbaryl	0.0020	U	ug/L
Carbofuran	0.0020	U	ug/L
Methiocarb	0.0020	U	ug/L
Methomyl	0.0020	U	ug/L
Oxamyl	0.0020	U	ug/L
Propoxur	0.0020	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B
 Batch ID: P379179

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: EPA 8321B
 Batch ID: P379190

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyropen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Clothianidin	0.0020	U	ug/L
Dinotefuran	0.0020	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: SM 2120 B-2011
 Batch ID: P379207

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2120 B-2011
 Batch ID: P379208

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P379416

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P379821

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P379601

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P379421

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P379361

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P379279

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	106		P	85 - 115
Calcium	107		P	85 - 115
Iron	109		P	85 - 115
Magnesium	112		P	85 - 115
Potassium	107		P	85 - 115
Sodium	112		P	85 - 115
Zinc	106		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P379279

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	91.9		P	85 - 115
Antimony	104		P	85 - 115
Arsenic	100		P	85 - 115
Boron	95.9		P	85 - 115
Cadmium	104		P	85 - 115
Chromium	97.1		P	85 - 115
Copper	95.3		P	85 - 115
Lead	95.4		P	85 - 115
Nickel	94.4		P	85 - 115
Selenium	105		P	85 - 115
Silver	100		P	85 - 115
Thallium	99.4		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379718

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	96.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379721

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379817

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P379461

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.2		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P379313

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P379608

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P379776

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379144

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.9		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379146

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.2		P	90 - 110

Reference Method: EPA 8270E
Batch ID: P379288

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	35.3	32.7	P/P	20 - 87.0
Alpha-BHC	82.6	74.6	P/P	65 - 142
Alpha-Chlordane	76.2	71.9	P/P	58 - 106
Beta-BHC	109	98.3	P/P	65 - 179
Bifenthrin	69.9	63.6	P/P	38 - 128
Carbophenothion	75.2	74.0	P/P	20 - 144
Chlorothalonil	91.4	38.0	P/P	20 - 214
Cis-Nonachlor	78.4	75.2	P/P	56 - 113
Cypermethrin	88.8	76.5	P/P	37 - 130
DDD-p,p'	89.8	83.1	P/P	66 - 119
DDE-p,p'	75.3	72.8	P/P	53 - 108
DDT-p,p'	69.0	67.7	P/P	57 - 133
Delta-BHC	113	93.9	P/P	68 - 187
Dicofol	84.5	79.5	P/P	66 - 123
Dieldrin	86.9	80.3	P/P	65 - 111
Endosulfan I	87.9	79.0	P/P	65 - 115
Endosulfan II	100	84.3	P/P	68 - 133
Endosulfan Sulfate	94.9	78.2	P/P	62 - 127
Endrin	87.2	81.0	P/P	64 - 122
Endrin Aldehyde	105	88.3	P/P	61 - 132
Endrin Ketone	97.4	82.7	P/P	66 - 120
Ethalfuralin	69.4	64.4	P/P	57 - 127

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P379288

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Gamma-BHC	106	92.3	P/P	71 - 158
Gamma-Chlordane	69.3	67.3	P/P	53 - 106
Heptachlor	59.6	58.2	P/P	45 - 96.0
Heptachlor Epoxide	79.9	73.4	P/P	64 - 109
Hexachlorobenzene	68.5	63.3	P/P	46 - 103
Methoxychlor	78.0	75.7	P/P	69 - 149
Mirex	58.5	56.5	P/P	31 - 90.0
Permethrin	71.5	65.9	P/P	41 - 108
Trans-Nonachlor	74.5	71.5	P/P	50 - 107
Trifluralin	69.3	66.3	P/P	58 - 119

Reference Method: EPA 8270E
Batch ID: P379289

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	99.5	94.5	P/P	71 - 122
Alachlor	99.0	94.9	P/P	70 - 121
Ametryn	103	116	P/P	49 - 137
Atrazine	100	101	P/P	76 - 125
Atrazine Desethyl	85.0	80.3	P/P	31 - 144
Azinphos Methyl	133	111	P/P	67 - 150
Bromacil	89.5	99.6	P/P	36 - 121
Butylate	48.8	60.6	P/P	33 - 88.0
Carbophenothion	104	108	P/P	50 - 150
Chlorpyrifos Ethyl	104	103	P/P	73 - 117
Chlorpyrifos Methyl	97.9	92.9	P/P	68 - 121
Cyanazine	127	99.0	P/P	20 - 250
Demeton	78.6	56.8	P/P	30 - 123
Diazinon	91.9	89.8	P/P	70 - 128
Dimethenamid	92.4	94.4	P/P	66 - 122
Disulfoton	87.6	65.7	P/P	46 - 124
Dithiopyr	97.4	98.5	P/P	69 - 122
EPTC	46.5	57.9	P/P	31 - 90.0
Ethion	98.9	107	P/P	45 - 135
Ethoprop	87.0	79.0	P/P	59 - 118
Fenamiphos	105	102	P/P	30 - 136
Fipronil	87.5	108	P/P	57 - 127
Fipronil Desulfinyl	94.6	107	P/P	51 - 122
Fipronil Sulfide	88.8	107	P/P	57 - 119
Fipronil Sulfone	95.4	104	P/P	51 - 119
Flumioxazin	157	161	P/P	70 - 200
Fonofos	93.4	77.1	P/P	60 - 120
Hexazinone	96.5	116	P/P	57 - 142
Imazalil	94.4	90.8	P/P	50 - 150
Malathion	109	111	P/P	67 - 128
Metalaxyl	106	101	P/P	21 - 148
Metolachlor	92.3	94.5	P/P	70 - 120
Metribuzin	104	104	P/P	20 - 253
Mevinphos	79.5	69.5	P/P	47 - 116
Molinate	66.0	63.7	P/P	41 - 99.0
Norflurazon	91.8	109	P/P	54 - 127
Oxadiazon	95.2	107	P/P	65 - 122

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P379289

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Parathion Ethyl	95.4	95.5	P/P	71 - 113
Parathion Methyl	95.9	96.4	P/P	73 - 129
Pendimethalin	93.1	94.3	P/P	63 - 130
Phorate	90.0	75.7	P/P	48 - 118
Prodiamine	104	93.8	P/P	20 - 144
Prometon	94.8	98.6	P/P	53 - 120
Prometryn	102	103	P/P	63 - 119
Pronamide	92.7	104	P/P	50 - 150
Propiconazole	99.8	117	P/P	50 - 150
Simazine	101	101	P/P	77 - 129
Sulfentrazone	70.2	99.9	P/P	50 - 150
Tebuconazole	59.8	103	P/P	20 - 118
Terbufos	92.6	79.5	P/P	61 - 119
Terbutylazine	96.0	98.0	P/P	75 - 119

Reference Method: EPA 8276
Batch ID: P379288

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	157	98.2	F/P	47 - 119
Toxaphene	76.9	82.6	P/P	41 - 112

Reference Method: EPA 8321B
Batch ID: P379101

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
3-Hydroxycarbofuran	88.6		P	30 - 160
Aldicarb	74.0		P	30 - 160
Aldicarb Sulfone	95.0		P	30 - 160
Aldicarb Sulfoxide	96.0		P	30 - 160
Carbaryl	61.9		P	30 - 160
Carbofuran	70.2		P	30 - 160
Methiocarb	73.0		P	30 - 160
Methomyl	85.6		P	30 - 160
Oxamyl	92.9		P	30 - 160
Propoxur	71.7		P	30 - 160

Reference Method: EPA 8321B
Batch ID: P379179

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	95.5		P	40 - 150
AMPA	98.1		P	40 - 150
Endothall	86.8		P	40 - 150
Glufosinate	95.0		P	40 - 150
Glyphosate	93.1		P	40 - 140
Sucralose	113		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P379190

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P379190

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	133		P	30 - 160
Acetaminophen	88.1		P	30 - 160
Acetamiprid	82.0		P	30 - 160
Afidopyropen	71.5		P	30 - 160
Bentazon	76.6		P	30 - 160
Benzovindiflupyr	78.1		P	30 - 160
Carbamazepine	80.5		P	30 - 160
Clothianidin	70.9		P	30 - 160
Dinotefuran	79.0		P	30 - 160
Diuron	78.7		P	30 - 160
Fenuron	91.3		P	30 - 160
Fluridone	92.4		P	30 - 160
Hydrocodone	82.6		P	30 - 160
Ibuprofen	86.7		P	30 - 160
Imazapyr	90.1		P	30 - 160
Imidacloprid	89.6		P	30 - 160
Linuron	73.5		P	30 - 160
Mandestrobin	87.7		P	30 - 160
MCPP	131		P	30 - 160
Naproxen	83.9		P	30 - 160
Primidone	79.7		P	30 - 160
Pyraclostrobin	82.3		P	30 - 160
Thiamethoxam	86.9		P	30 - 160
Tolfenpyrad	58.6		P	30 - 160
Triclopyr	105		P	30 - 160

Reference Method: SM 2120 B-2011
Batch ID: P379207

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	85 - 115

Reference Method: SM 2120 B-2011
Batch ID: P379208

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P379416

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P379421

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.0		P	92 - 103

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P379361

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P379279

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158739	Barium	106		P	80 - 120
2158739	Calcium	102		P	80 - 120
2158739	Iron	110		P	80 - 120
2158739	Magnesium	103		P	80 - 120
2158739	Potassium	108		P	80 - 120
2158739	Sodium	103		P	80 - 120
2158739	Zinc	105		P	80 - 120
2159002	Barium	105	105	P/P	80 - 120
2159002	Calcium	109	105	P/P	80 - 120
2159002	Iron	109	105	P/P	80 - 120
2159002	Magnesium	116	111	P/P	80 - 120
2159002	Potassium	108	107	P/P	80 - 120
2159002	Sodium	106	105	P/P	80 - 120
2159002	Zinc	105	105	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P379279

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158739	Aluminum	96.9		P	80 - 120
2158739	Antimony	107		P	80 - 120
2158739	Arsenic	105		P	80 - 120
2158739	Boron	98.1		P	80 - 120
2158739	Cadmium	105		P	80 - 120
2158739	Chromium	99.6		P	80 - 120
2158739	Copper	97.0		P	80 - 120
2158739	Lead	96.7		P	80 - 120
2158739	Nickel	95.7		P	80 - 120
2158739	Selenium	109		P	80 - 120
2158739	Silver	101		P	80 - 120
2158739	Thallium	107		P	80 - 120
2159002	Aluminum	97.2	96.0	P/P	80 - 120
2159002	Antimony	104	104	P/P	80 - 120
2159002	Arsenic	102	101	P/P	80 - 120
2159002	Boron	95.6	95.6	P/P	80 - 120
2159002	Cadmium	102	101	P/P	80 - 120
2159002	Chromium	99.1	98.6	P/P	80 - 120
2159002	Copper	99.0	98.4	P/P	80 - 120
2159002	Lead	95.7	95.3	P/P	80 - 120
2159002	Nickel	97.3	97.9	P/P	80 - 120
2159002	Selenium	107	106	P/P	80 - 120
2159002	Silver	100	99.9	P/P	80 - 120
2159002	Thallium	105	106	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379718

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158479	Chloride	99.6		P	90 - 110
2158532	Chloride	102		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379721

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158479	Sulfate	96.8		P	90 - 110
2158532	Sulfate	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379817

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2159619	Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P379461

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158522	Ammonia-N	99.1	92.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P379313

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158527	Kjeldahl Nitrogen	104	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P379608

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158404	NO2NO3-N	98.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P379776

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158485	Total-P	103	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379144

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158486	O-Phosphate-P	95.0	96.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379146

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158698	O-Phosphate-P	92.0	92.8	P/P	90 - 110

Reference Method: EPA 8270E
Batch ID: P379288

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158658	Aldrin	44.5	42.2	P/P	32 - 82.0
2158658	Alpha-BHC	88.4	76.7	P/P	68 - 157
2158658	Alpha-Chlordane	76.3	74.6	P/P	54 - 108

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E
Batch ID: P379288

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158658	Beta-BHC	145	133	P/P	54 - 200
2158658	Bifenthrin	66.5	72.3	P/P	49 - 124
2158658	Carbophenothion	92.9	94.0	P/P	42 - 173
2158658	Chlorothalonil	93.9	67.1	P/P	10 - 289
2158658	Cis-Nonachlor	78.6	78.1	P/P	53 - 112
2158658	Cypermethrin	85.4	83.9	P/P	43 - 141
2158658	DDD-p,p'	90.5	88.1	P/P	58 - 125
2158658	DDE-p,p'	74.6	72.8	P/P	50 - 108
2158658	DDT-p,p'	67.9	66.5	P/P	52 - 140
2158658	Delta-BHC	140	131	P/P	61 - 228
2158658	Dicofol	75.8	70.3	P/P	59 - 130
2158658	Dieldrin	91.4	93.7	P/P	57 - 121
2158658	Endosulfan I	89.9	84.1	P/P	62 - 123
2158658	Endosulfan II	102	94.8	P/P	48 - 163
2158658	Endosulfan Sulfate	100	89.7	P/P	63 - 132
2158658	Endrin	86.9	82.7	P/P	58 - 128
2158658	Endrin Aldehyde	91.8	81.7	P/P	44 - 126
2158658	Endrin Ketone	107	95.2	P/P	68 - 127
2158658	Ethalfuralin	67.1	61.5	P/P	55 - 131
2158658	Gamma-BHC	135	120	P/P	54 - 206
2158658	Gamma-Chlordane	68.0	67.1	P/P	52 - 105
2158658	Heptachlor	55.6	54.3	P/P	47 - 94.0
2158658	Heptachlor Epoxide	82.7	77.1	P/P	61 - 113
2158658	Hexachlorobenzene	66.4	61.0	P/P	45 - 100
2158658	Methoxychlor	78.3	72.7	P/P	63 - 153
2158658	Mirex	56.3	57.2	P/P	38 - 90.0
2158658	Permethrin	71.8	70.4	P/P	45 - 112
2158658	Trans-Nonachlor	74.7	74.3	P/P	50 - 103
2158658	Trifluralin	69.1	64.7	P/P	57 - 122

Reference Method: EPA 8270E
Batch ID: P379289

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158576	Acetochlor	101	96.3	P/P	68 - 126
2158576	Alachlor	103	97.8	P/P	66 - 125
2158576	Ametryn	101	98.3	P/P	31 - 182
2158576	Atrazine Desethyl	60.9	84.6	P/P	13 - 161
2158576	Azinphos Methyl	125	122	P/P	54 - 162
2158576	Bromacil	88.2	78.7	P/P	34 - 144
2158576	Butylate	32.5	50.9	P/P	16 - 92.0
2158576	Carbophenothion	110	101	P/P	50 - 150
2158576	Chlorpyrifos Ethyl	123	108	P/P	63 - 124
2158576	Chlorpyrifos Methyl	103	99.9	P/P	64 - 125
2158576	Cyanazine	106	102	P/P	13 - 245
2158576	Demeton	85.9	87.5	P/P	20 - 154
2158576	Diazinon	97.0	93.7	P/P	66 - 141
2158576	Dimethenamid	100	94.3	P/P	50 - 130
2158576	Disulfoton	90.2	89.5	P/P	55 - 130
2158576	Dithiopyr	99.8	92.7	P/P	66 - 122
2158576	EPTC	29.1	41.4	P/P	15 - 93.0
2158576	Ethion	121	107	P/P	15 - 148

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E
 Batch ID: P379289

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158576	Ethoprop	95.7	96.1	P/P	59 - 126
2158576	Fenamiphos	100	94.3	P/P	39 - 137
2158576	Fipronil	109	100	P/P	47 - 131
2158576	Fipronil Desulfinyl	110	97.4	P/P	43 - 121
2158576	Fipronil Sulfide	112	98.5	P/P	29 - 134
2158576	Fipronil Sulfone	108	90.8	P/P	19 - 131
2158576	Flumioxazin	177	174	P/P	70 - 200
2158576	Fonofos	93.0	85.8	P/P	62 - 128
2158576	Imazalil	66.8	210	P/F	50 - 150
2158576	Malathion	115	104	P/P	54 - 135
2158576	Metalaxyl	105	100	P/P	46 - 136
2158576	Mevinphos	83.8	86.7	P/P	44 - 135
2158576	Molinate	57.2	68.8	P/P	31 - 104
2158576	Norflurazon	106	85.4	P/P	32 - 134
2158576	Oxadiazon	96.3	88.1	P/P	55 - 123
2158576	Parathion Ethyl	108	97.6	P/P	67 - 120
2158576	Parathion Methyl	109	105	P/P	70 - 140
2158576	Pendimethalin	110	101	P/P	59 - 145
2158576	Phorate	82.6	86.5	P/P	52 - 127
2158576	Prodiamine	110	76.7	P/P	11 - 157
2158576	Prometon	89.1	93.4	P/P	55 - 130
2158576	Prometryn	106	99.8	P/P	55 - 127
2158576	Pronamide	99.4	108	P/P	50 - 150
2158576	Propiconazole	112	111	P/P	50 - 150
2158576	Simazine	119	107	P/P	58 - 159
2158576	Sulfentrazone	100	90.6	P/P	50 - 150
2158576	Tebuconazole	97.4	76.4	P/P	10 - 122
2158576	Terbufos	97.9	98.6	P/P	65 - 131
2158576	Terbutylazine	102	95.1	P/P	63 - 131

Reference Method: EPA 8276
 Batch ID: P379288

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158600	Chlordane	97.3	95.1	P/P	34 - 129
2158600	Toxaphene	76.5	82.0	P/P	26 - 123

Reference Method: EPA 8321B
 Batch ID: P379101

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2157327	3-Hydroxycarbofuran	83.4	82.9	P/P	30 - 160
2157327	Aldicarb	78.2	64.9	P/P	30 - 160
2157327	Aldicarb Sulfone	103	101	P/P	30 - 160
2157327	Aldicarb Sulfoxide	44.3	45.6	P/P	30 - 160
2157327	Carbaryl	50.0	61.4	P/P	30 - 160
2157327	Carbofuran	67.6	69.0	P/P	30 - 160
2157327	Methiocarb	58.9	63.7	P/P	30 - 160
2157327	Methomyl	77.8	90.0	P/P	30 - 160
2157327	Oxamyl	96.1	98.0	P/P	30 - 160
2157327	Propoxur	62.4	66.7	P/P	30 - 160

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P379179

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158377	Acesulfame K	79.3	75.2	P/P	40 - 150
2158377	AMPA	83.7	77.7	P/P	40 - 150
2158377	Endothall	106	97.5	P/P	40 - 150
2158377	Glufosinate	195	246	F/F	40 - 150
2158377	Glyphosate	183	213	F/F	40 - 140
2158377	Sucralose	108	101	P/P	40 - 150

Reference Method: EPA 8321B
Batch ID: P379190

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158377	2,4-D	131	142	P/P	30 - 160
2158377	Acetaminophen	92.5	114	P/P	30 - 160
2158377	Acetamiprid	78.4	57.3	P/P	30 - 160
2158377	Afidopyropen	87.0	80.3	P/P	30 - 160
2158377	Benzovindiflupyr	75.4	82.7	P/P	30 - 160
2158377	Carbamazepine	66.9	70.5	P/P	30 - 160
2158377	Clothianidin	92.3	104	P/P	30 - 160
2158377	Dinotefuran	119	120	P/P	30 - 160
2158377	Diuron	63.8	65.4	P/P	30 - 160
2158377	Fenuron	77.4	71.1	P/P	30 - 160
2158377	Fluridone	88.0	90.0	P/P	30 - 160
2158377	Hydrocodone	78.4	95.1	P/P	30 - 160
2158377	Ibuprofen	79.4	82.2	P/P	30 - 160
2158377	Imidacloprid	157	138	P/P	30 - 160
2158377	Linuron	75.0	71.5	P/P	30 - 160
2158377	Mandestrobin	84.9	88.5	P/P	30 - 160
2158377	MCP	133	134	P/P	30 - 160
2158377	Naproxen	97.3	117	P/P	30 - 160
2158377	Primidone	85.7	88.2	P/P	30 - 160
2158377	Pyraclostrobin	83.5	88.0	P/P	30 - 160
2158377	Thiamethoxam	108	121	P/P	30 - 160
2158377	Tolfenpyrad	57.1	57.5	P/P	30 - 160
2158377	Triclopyr	115	130	P/P	30 - 160

Reference Method: SM 4500 F-C-2011
Batch ID: P379421

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158297	Fluoride	98.8	99.3	P/P	89 - 106

Reference Method: SM 5310 B-2011
Batch ID: P379361

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2158477	Organic Carbon	99.8	101	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P379138

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158521	Turbidity	0.0	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P379139

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158696	Turbidity	0.881	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P379279

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2159002	Barium	0.640	Spike	P	0 - 20
2159002	Calcium	2.88	Spike	P	0 - 20
2159002	Iron	3.66	Spike	P	0 - 20
2159002	Magnesium	3.50	Spike	P	0 - 20
2159002	Potassium	0.422	Spike	P	0 - 20
2159002	Sodium	0.981	Spike	P	0 - 20
2159002	Zinc	0.495	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P379279

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2159002	Aluminum	1.24	Spike	P	0 - 20
2159002	Antimony	0.00801	Spike	P	0 - 20
2159002	Arsenic	0.201	Spike	P	0 - 20
2159002	Boron	0.0281	Spike	P	0 - 20
2159002	Cadmium	0.446	Spike	P	0 - 20
2159002	Chromium	0.477	Spike	P	0 - 20
2159002	Copper	0.656	Spike	P	0 - 20
2159002	Lead	0.451	Spike	P	0 - 20
2159002	Nickel	0.588	Spike	P	0 - 20
2159002	Selenium	1.73	Spike	P	0 - 20
2159002	Silver	0.355	Spike	P	0 - 20
2159002	Thallium	0.647	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379718

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158479	Chloride	0.696	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379721

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158479	Sulfate	1.03	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P379817

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2159548	Sulfate	0.278	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P379461

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158522	Ammonia-N	4.86	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P379313

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158527	Kjeldahl Nitrogen	2.20	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P379608

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158404	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P379776

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158485	Total-P	0.615	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379144

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158486	O-Phosphate-P	1.01	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P379146

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158698	O-Phosphate-P	0.427	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P379288

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158658	Aldrin	5.31	Spike	P	0 - 30
2158658	Alpha-BHC	14.2	Spike	P	0 - 30
2158658	Alpha-Chlordane	2.20	Spike	P	0 - 30
2158658	Beta-BHC	9.28	Spike	P	0 - 30
2158658	Bifenthrin	8.30	Spike	P	0 - 30
2158658	Carbophenothion	1.15	Spike	P	0 - 30
2158658	Chlorothalonil	33.2	Spike	F	0 - 30
2158658	Cis-Nonachlor	0.671	Spike	P	0 - 30
2158658	Cypermethrin	1.76	Spike	P	0 - 30
2158658	DDD-p,p'	2.73	Spike	P	0 - 30
2158658	DDE-p,p'	2.49	Spike	P	0 - 30
2158658	DDT-p,p'	2.05	Spike	P	0 - 30
2158658	Delta-BHC	6.48	Spike	P	0 - 30
2158658	Dicofol	7.42	Spike	P	0 - 30
2158658	Dieldrin	2.53	Spike	P	0 - 30
2158658	Endosulfan I	6.69	Spike	P	0 - 30
2158658	Endosulfan II	7.51	Spike	P	0 - 30
2158658	Endosulfan Sulfate	11.3	Spike	P	0 - 30
2158658	Endrin	4.98	Spike	P	0 - 30
2158658	Endrin Aldehyde	11.7	Spike	P	0 - 30
2158658	Endrin Ketone	11.4	Spike	P	0 - 30
2158658	Ethalfuralin	8.70	Spike	P	0 - 30
2158658	Gamma-BHC	12.2	Spike	P	0 - 30
2158658	Gamma-Chlordane	1.30	Spike	P	0 - 30
2158658	Heptachlor	2.43	Spike	P	0 - 30
2158658	Heptachlor Epoxide	7.11	Spike	P	0 - 30
2158658	Hexachlorobenzene	8.48	Spike	P	0 - 30
2158658	Methoxychlor	7.44	Spike	P	0 - 30
2158658	Mirex	1.61	Spike	P	0 - 30
2158658	Permethrin	1.99	Spike	P	0 - 30
2158658	Trans-Nonachlor	0.547	Spike	P	0 - 30
2158658	Trifluralin	6.66	Spike	P	0 - 30
LFB	Aldrin	7.73	LCS	P	0 - 30
LFB	Alpha-BHC	10.1	LCS	P	0 - 30
LFB	Alpha-Chlordane	5.85	LCS	P	0 - 30
LFB	Beta-BHC	9.93	LCS	P	0 - 30
LFB	Bifenthrin	9.39	LCS	P	0 - 30
LFB	Carbophenothion	1.61	LCS	P	0 - 30
LFB	Chlorothalonil	82.5	LCS	F	0 - 30
LFB	Cis-Nonachlor	4.19	LCS	P	0 - 30
LFB	Cypermethrin	14.9	LCS	P	0 - 30
LFB	DDD-p,p'	7.85	LCS	P	0 - 30
LFB	DDE-p,p'	3.32	LCS	P	0 - 30
LFB	DDT-p,p'	1.90	LCS	P	0 - 30
LFB	Delta-BHC	18.7	LCS	P	0 - 30
LFB	Dicofol	6.10	LCS	P	0 - 30
LFB	Dieldrin	7.88	LCS	P	0 - 30
LFB	Endosulfan I	10.7	LCS	P	0 - 30
LFB	Endosulfan II	17.1	LCS	P	0 - 30
LFB	Endosulfan Sulfate	19.3	LCS	P	0 - 30
LFB	Endrin	7.40	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P379288

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Endrin Aldehyde	17.0	LCS	P	0 - 30
LFB	Endrin Ketone	16.3	LCS	P	0 - 30
LFB	Ethalfuralin	7.52	LCS	P	0 - 30
LFB	Gamma-BHC	13.3	LCS	P	0 - 30
LFB	Gamma-Chlordane	2.90	LCS	P	0 - 30
LFB	Heptachlor	2.44	LCS	P	0 - 30
LFB	Heptachlor Epoxide	8.39	LCS	P	0 - 30
LFB	Hexachlorobenzene	7.85	LCS	P	0 - 30
LFB	Methoxychlor	3.00	LCS	P	0 - 30
LFB	Mirex	3.53	LCS	P	0 - 30
LFB	Permethrin	8.11	LCS	P	0 - 30
LFB	Trans-Nonachlor	4.11	LCS	P	0 - 30
LFB	Trifluralin	4.42	LCS	P	0 - 30

Reference Method: EPA 8270E
Batch ID: P379289

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158576	Acetochlor	4.82	Spike	P	0 - 30
2158576	Alachlor	4.86	Spike	P	0 - 30
2158576	Ametryn	2.56	Spike	P	0 - 30
2158576	Atrazine	0.914	Spike	P	0 - 30
2158576	Atrazine Desethyl	32.6	Spike	F	0 - 30
2158576	Azinphos Methyl	2.79	Spike	P	0 - 30
2158576	Bromacil	6.53	Spike	P	0 - 30
2158576	Butylate	44.1	Spike	F	0 - 30
2158576	Carbophenothion	8.54	Spike	P	0 - 30
2158576	Chlorpyrifos Ethyl	12.8	Spike	P	0 - 30
2158576	Chlorpyrifos Methyl	3.51	Spike	P	0 - 30
2158576	Cyanazine	3.58	Spike	P	0 - 30
2158576	Demeton	1.84	Spike	P	0 - 30
2158576	Diazinon	3.38	Spike	P	0 - 30
2158576	Dimethenamid	5.93	Spike	P	0 - 30
2158576	Disulfoton	0.861	Spike	P	0 - 30
2158576	Dithiopyr	6.97	Spike	P	0 - 30
2158576	EPTC	34.9	Spike	F	0 - 30
2158576	Ethion	12.1	Spike	P	0 - 30
2158576	Ethoprop	0.412	Spike	P	0 - 30
2158576	Fenamiphos	6.01	Spike	P	0 - 30
2158576	Fipronil	8.25	Spike	P	0 - 30
2158576	Fipronil Desulfinyl	12.2	Spike	P	0 - 30
2158576	Fipronil Sulfide	12.6	Spike	P	0 - 30
2158576	Fipronil Sulfone	15.0	Spike	P	0 - 30
2158576	Flumioxazin	1.43	Spike	P	0 - 30
2158576	Fonofos	8.11	Spike	P	0 - 30
2158576	Hexazinone	2.10	Spike	P	0 - 30
2158576	Imazalil	103	Spike	F	0 - 30
2158576	Malathion	7.70	Spike	P	0 - 30
2158576	Metalaxyl	4.64	Spike	P	0 - 30
2158576	Metolachlor	0.666	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P379289

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158576	Metribuzin	5.83	Spike	P	0 - 30
2158576	Mevinphos	3.40	Spike	P	0 - 30
2158576	Molinate	18.4	Spike	P	0 - 30
2158576	Norflurazon	14.3	Spike	P	0 - 30
2158576	Oxadiazon	7.41	Spike	P	0 - 30
2158576	Parathion Ethyl	10.3	Spike	P	0 - 30
2158576	Parathion Methyl	3.53	Spike	P	0 - 30
2158576	Pendimethalin	8.53	Spike	P	0 - 30
2158576	Phorate	4.55	Spike	P	0 - 30
2158576	Prodiamine	35.8	Spike	F	0 - 30
2158576	Prometon	4.70	Spike	P	0 - 30
2158576	Prometryn	6.20	Spike	P	0 - 30
2158576	Pronamide	7.96	Spike	P	0 - 30
2158576	Propiconazole	0.837	Spike	P	0 - 30
2158576	Simazine	10.9	Spike	P	0 - 30
2158576	Sulfentrazone	10.3	Spike	P	0 - 30
2158576	Tebuconazole	22.2	Spike	P	0 - 30
2158576	Terbufos	0.749	Spike	P	0 - 30
2158576	Terbutylazine	6.68	Spike	P	0 - 30
LFB	Acetochlor	5.23	LCS	P	0 - 30
LFB	Alachlor	4.30	LCS	P	0 - 30
LFB	Ametryn	12.1	LCS	P	0 - 30
LFB	Atrazine	0.993	LCS	P	0 - 30
LFB	Atrazine Desethyl	5.76	LCS	P	0 - 30
LFB	Azinphos Methyl	18.7	LCS	P	0 - 30
LFB	Bromacil	10.6	LCS	P	0 - 30
LFB	Butylate	21.6	LCS	P	0 - 30
LFB	Carbophenothion	4.56	LCS	P	0 - 30
LFB	Chlorpyrifos Ethyl	0.380	LCS	P	0 - 30
LFB	Chlorpyrifos Methyl	5.19	LCS	P	0 - 30
LFB	Cyanazine	24.5	LCS	P	0 - 30
LFB	Demeton	32.1	LCS	F	0 - 30
LFB	Diazinon	2.31	LCS	P	0 - 30
LFB	Dimethenamid	2.12	LCS	P	0 - 30
LFB	Disulfoton	28.6	LCS	P	0 - 30
LFB	Dithiopyr	1.16	LCS	P	0 - 30
LFB	EPTC	21.8	LCS	P	0 - 30
LFB	Ethion	7.62	LCS	P	0 - 30
LFB	Ethoprop	9.55	LCS	P	0 - 30
LFB	Fenamiphos	2.54	LCS	P	0 - 30
LFB	Fipronil	21.2	LCS	P	0 - 30
LFB	Fipronil Desulfinyl	12.8	LCS	P	0 - 30
LFB	Fipronil Sulfide	18.9	LCS	P	0 - 30
LFB	Fipronil Sulfone	8.42	LCS	P	0 - 30
LFB	Flumioxazin	2.02	LCS	P	0 - 30
LFB	Fonofos	19.1	LCS	P	0 - 30
LFB	Hexazinone	18.1	LCS	P	0 - 30
LFB	Imazalil	3.84	LCS	P	0 - 30
LFB	Malathion	2.14	LCS	P	0 - 30
LFB	Metalaxyl	5.64	LCS	P	0 - 30
LFB	Metolachlor	2.33	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P379289

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Metribuzin	0.236	LCS	P	0 - 30
LFB	Mevinphos	13.5	LCS	P	0 - 30
LFB	Molinate	3.45	LCS	P	0 - 30
LFB	Norflurazon	17.2	LCS	P	0 - 30
LFB	Oxadiazon	11.3	LCS	P	0 - 30
LFB	Parathion Ethyl	0.157	LCS	P	0 - 30
LFB	Parathion Methyl	0.529	LCS	P	0 - 30
LFB	Pendimethalin	1.32	LCS	P	0 - 30
LFB	Phorate	17.3	LCS	P	0 - 30
LFB	Prodiamine	10.4	LCS	P	0 - 30
LFB	Prometon	3.91	LCS	P	0 - 30
LFB	Prometryn	1.77	LCS	P	0 - 30
LFB	Pronamide	11.4	LCS	P	0 - 30
LFB	Propiconazole	15.9	LCS	P	0 - 30
LFB	Simazine	0.173	LCS	P	0 - 30
LFB	Sulfentrazone	34.9	LCS	F	0 - 30
LFB	Tebuconazole	53.0	LCS	F	0 - 30
LFB	Terbufos	15.2	LCS	P	0 - 30
LFB	Terbutylazine	2.05	LCS	P	0 - 30

Reference Method: EPA 8276
Batch ID: P379288

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158600	Chlordane	2.32	Spike	P	0 - 30
2158600	Toxaphene	6.90	Spike	P	0 - 30
LFB	Chlordane	45.8	LCS	F	0 - 30
LFB	Toxaphene	7.22	LCS	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P379101

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2157327	3-Hydroxycarbofuran	0.617	Spike	P	0 - 30
2157327	Aldicarb	18.6	Spike	P	0 - 30
2157327	Aldicarb Sulfone	2.41	Spike	P	0 - 30
2157327	Aldicarb Sulfoxide	3.00	Spike	P	0 - 30
2157327	Carbaryl	20.4	Spike	P	0 - 30
2157327	Carbofuran	2.07	Spike	P	0 - 30
2157327	Methiocarb	7.82	Spike	P	0 - 30
2157327	Methomyl	14.5	Spike	P	0 - 30
2157327	Oxamyl	2.01	Spike	P	0 - 30
2157327	Propoxur	5.32	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P379179

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P379179

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158377	Acesulfame K	4.80	Spike	P	0 - 30
2158377	AMPA	7.50	Spike	P	0 - 30
2158377	Endothall	8.73	Spike	P	0 - 30
2158377	Glufosinate	23.2	Spike	P	0 - 30
2158377	Glyphosate	15.1	Spike	P	0 - 30
2158377	Sucralose	5.72	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P379190

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158377	2,4-D	8.66	Spike	P	0 - 30
2158377	Acetaminophen	20.6	Spike	P	0 - 30
2158377	Acetamiprid	31.1	Spike	F	0 - 30
2158377	Afidopyropen	8.04	Spike	P	0 - 30
2158377	Bentazon	21.0	Spike	P	0 - 30
2158377	Benzovindiflupyr	9.23	Spike	P	0 - 30
2158377	Carbamazepine	5.32	Spike	P	0 - 30
2158377	Clothianidin	12.3	Spike	P	0 - 30
2158377	Dinotefuran	0.777	Spike	P	0 - 30
2158377	Diuron	2.51	Spike	P	0 - 30
2158377	Fenuron	8.43	Spike	P	0 - 30
2158377	Fluridone	1.75	Spike	P	0 - 30
2158377	Hydrocodone	19.2	Spike	P	0 - 30
2158377	Ibuprofen	3.35	Spike	P	0 - 30
2158377	Imazapyr	2.90	Spike	P	0 - 30
2158377	Imidacloprid	12.8	Spike	P	0 - 30
2158377	Linuron	4.70	Spike	P	0 - 30
2158377	Mandestrobin	4.07	Spike	P	0 - 30
2158377	MCPP	0.847	Spike	P	0 - 30
2158377	Naproxen	18.7	Spike	P	0 - 30
2158377	Primidone	2.92	Spike	P	0 - 30
2158377	Pyraclostrobin	5.19	Spike	P	0 - 30
2158377	Thiamethoxam	11.2	Spike	P	0 - 30
2158377	Tolfenpyrad	0.759	Spike	P	0 - 30
2158377	Triclopyr	11.8	Spike	P	0 - 30

Reference Method: SM 2120 B-2011
Batch ID: P379207

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158526	Color (true)	1.25	Sample	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P379208

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158529	Color (true)	0.993	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
 Batch ID: P379416

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158297	Total Alkalinity	0.288	Sample	P	0 - 2.8

Reference Method: SM 2540 C-2011
 Batch ID: P379821

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158521	TDS	1.87	Sample	P	0 - 10

Reference Method: SM 2540 D-2011
 Batch ID: P379601

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158521	TSS	8.00	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
 Batch ID: P379421

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158297	Fluoride	0.477	Spike	P	0 - 2.5

Reference Method: SM 5310 B-2011
 Batch ID: P379361

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2158477	Organic Carbon	0.707	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2158561
Field Sample ID: G3SD0157

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbaryl-d7	49.4	P	30 - 160

Lab Sample ID: 2158562
Field Sample ID: G3SD0133

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbaryl-d7	51.3	P	30 - 160

Lab Sample ID: 2158563
Field Sample ID: G3SD0157

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	120	P	30 - 160
EPA 8321B	Acetaminophen-d4	95.7	P	30 - 160
EPA 8321B	Carbamazepine-d10	73.4	P	30 - 160
EPA 8321B	Diuron-d6	66.6	P	30 - 160
EPA 8321B	Endothall-d6	109	P	30 - 160
EPA 8321B	Endothall-d6	109	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	82.9	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	82.9	P	30 - 160
EPA 8321B	Ibuprofen-d3	90.2	P	30 - 160
EPA 8321B	Primidone-d5	90.6	P	30 - 160
EPA 8321B	Sucralose-d6	44.9	P	30 - 160
EPA 8321B	Sucralose-d6	44.9	P	30 - 160

Lab Sample ID: 2158564
Field Sample ID: G3SD0133

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	105	P	30 - 160
EPA 8321B	Acetaminophen-d4	87.2	P	30 - 160
EPA 8321B	Carbamazepine-d10	78.8	P	30 - 160
EPA 8321B	Diuron-d6	67.8	P	30 - 160
EPA 8321B	Endothall-d6	115	P	30 - 160
EPA 8321B	Endothall-d6	115	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	84.4	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	84.4	P	30 - 160
EPA 8321B	Ibuprofen-d3	75.6	P	30 - 160
EPA 8321B	Primidone-d5	92.7	P	30 - 160
EPA 8321B	Sucralose-d6	32.5	P	30 - 160
EPA 8321B	Sucralose-d6	32.5	P	30 - 160

Lab Sample ID: 2158565
Field Sample ID: G3SD0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbaryl-d7	46.4	P	30 - 160

Lab Sample ID: 2158566
Field Sample ID: G3SD0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
------------------	-----------	--------	-----------	----------------

Quality Assurance Report Surrogates

Lab Sample ID: 2158566
Field Sample ID: G3SD0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	117	P	30 - 160
EPA 8321B	Acetaminophen-d4	91.7	P	30 - 160
EPA 8321B	Carbamazepine-d10	79.1	P	30 - 160
EPA 8321B	Diuron-d6	75.1	P	30 - 160
EPA 8321B	Endothall-d6	109	P	30 - 160
EPA 8321B	Endothall-d6	109	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	105	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	105	P	30 - 160
EPA 8321B	Ibuprofen-d3	88.1	P	30 - 160
EPA 8321B	Primidone-d5	92.5	P	30 - 160
EPA 8321B	Sucralose-d6	40.7	P	30 - 160
EPA 8321B	Sucralose-d6	40.7	P	30 - 160

Lab Sample ID: 2158573
Field Sample ID: G3SD0157

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Decachlorobiphenyl	80.9	P	44 - 112
EPA 8270E	Tetrachloro-m-xylene	65.7	P	34 - 106
EPA 8276	Decachlorobiphenyl	73.5	P	26 - 97.0
EPA 8276	PCNB	89.2	P	38 - 141

Lab Sample ID: 2158574
Field Sample ID: G3SD0133

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Decachlorobiphenyl	80.6	P	44 - 112
EPA 8270E	Tetrachloro-m-xylene	61.3	P	34 - 106
EPA 8276	Decachlorobiphenyl	55.7	P	26 - 97.0
EPA 8276	PCNB	84.7	P	38 - 141

Lab Sample ID: 2158575
Field Sample ID: G3SD0157

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Simazine-d10	94.4	P	72 - 135
EPA 8270E	Terbufos-d10	82.4	P	57 - 134

Lab Sample ID: 2158576
Field Sample ID: G3SD0133

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Simazine-d10	116	P	72 - 135
EPA 8270E	Terbufos-d10	109	P	57 - 134

Lab Sample ID: 2158577
Field Sample ID: G3SD0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Decachlorobiphenyl	78.3	P	44 - 112
EPA 8270E	Tetrachloro-m-xylene	59.4	P	34 - 106
EPA 8276	Decachlorobiphenyl	62.3	P	26 - 97.0

Quality Assurance Report Surrogates

Lab Sample ID: 2158577
Field Sample ID: G3SD0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8276	PCNB	76.9	P	38 - 141

Lab Sample ID: 2158578
Field Sample ID: G3SD0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Simazine-d10	99.8	P	72 - 135
EPA 8270E	Terbufos-d10	86.5	P	57 - 134

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A37839

Included Lab Sample IDs: 2158579, 2158580, 2158581, 2158582

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Selenium	104	105	P/P	90 - 110
Selenium	105	106	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A97621

Included Lab Sample IDs: 2158520, 2158521, 2158526, 2158529, 2158532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105
Turbidity	101	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A97622

Included Lab Sample IDs: 2158524, 2158525, 2158528, 2158531, 2158534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	95.3	93.0	P/P	90 - 110
O-Phosphate-P	95.7	95.3	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A97655

Included Lab Sample IDs: 2158520, 2158521, 2158526, 2158529, 2158532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	101	P/P	90 - 115
Color (true)	101	103	P/P	90 - 115
Color (true)	98.6	100	P/P	90 - 115

Reference Method: EPA 8321B

Run ID: A97710

Included Lab Sample IDs: 2158561, 2158562, 2158565

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
3-Hydroxycarbofuran	107	97.4	P/P	60 - 150
3-Hydroxycarbofuran	97.4	106	P/P	60 - 150
Aldicarb	112	106	P/P	60 - 150
Aldicarb	134	112	P/P	60 - 150
Aldicarb Sulfone	107	110	P/P	50 - 150
Aldicarb Sulfone	110	105	P/P	50 - 150
Aldicarb Sulfoxide	107	108	P/P	50 - 150
Aldicarb Sulfoxide	108	100	P/P	50 - 150
Carbaryl	108	110	P/P	60 - 150
Carbaryl	110	103	P/P	60 - 150
Carbofuran	114	113	P/P	50 - 150
Carbofuran	116	114	P/P	50 - 150
Methiocarb	101	97.7	P/P	50 - 150
Methiocarb	97.7	98.3	P/P	50 - 150
Methomyl	102	108	P/P	50 - 150
Methomyl	108	101	P/P	50 - 150
Oxamyl	102	95.1	P/P	50 - 150
Oxamyl	108	102	P/P	50 - 150

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B

Run ID: A97710

Included Lab Sample IDs: 2158561, 2158562, 2158565

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Propoxur	117	84.1	P/P	50 - 150
Propoxur	84.1	102	P/P	50 - 150

Reference Method: SM 5310 B-2011

Run ID: A97718

Included Lab Sample IDs: 2158522, 2158523, 2158527, 2158530, 2158533

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.1	100	P/P	90 - 110
Organic Carbon	100	96.9	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A97739

Included Lab Sample IDs: 2158520, 2158521, 2158526, 2158529, 2158532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110
Total Alkalinity	104	102	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A97739

Included Lab Sample IDs: 2158520, 2158521, 2158526, 2158529, 2158532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A97754

Included Lab Sample IDs: 2158522, 2158523, 2158527, 2158530, 2158533

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	94.6	92.9	P/P	90 - 110
Ammonia-N	98.0	94.6	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A97771

Included Lab Sample IDs: 2158520, 2158521, 2158526, 2158529, 2158532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.2	98.6	P/P	90 - 110
Chloride	98.6	98.3	P/P	90 - 110
Sulfate	97.5	98.4	P/P	90 - 110
Sulfate	98.4	98.7	P/P	90 - 110
Sulfate	99.2	99.1	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A97772

Included Lab Sample IDs: 2158579, 2158580, 2158581, 2158582

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A97772

Included Lab Sample IDs: 2158579, 2158580, 2158581, 2158582

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	90.8	94.7	P/P	90 - 110
Aluminum	94.7	93.7	P/P	90 - 110
Antimony	100	102	P/P	90 - 110
Antimony	102	102	P/P	90 - 110
Arsenic	95.5	97.2	P/P	90 - 110
Arsenic	97.2	97.3	P/P	90 - 110
Boron	100	98.4	P/P	90 - 110
Boron	102	100	P/P	90 - 110
Cadmium	97.5	97.5	P/P	90 - 110
Cadmium	97.6	97.5	P/P	90 - 110
Chromium	93.8	95.4	P/P	90 - 110
Chromium	95.4	94.4	P/P	90 - 110
Copper	92.6	92.9	P/P	90 - 110
Copper	92.9	93.5	P/P	90 - 110
Lead	93.0	93.8	P/P	90 - 110
Lead	93.8	93.0	P/P	90 - 110
Nickel	93.2	94.1	P/P	90 - 110
Nickel	94.1	93.2	P/P	90 - 110
Silver	99.1	99.3	P/P	90 - 110
Silver	99.3	99.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A97781

Included Lab Sample IDs: 2158522, 2158523, 2158527, 2158530, 2158533

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	96.9	P/P	90 - 110
Kjeldahl Nitrogen	96.7	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A97783

Included Lab Sample IDs: 2158522, 2158523, 2158527, 2158530, 2158533

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A97786

Included Lab Sample IDs: 2158579, 2158580, 2158581, 2158582

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	102	101	P/P	90 - 110
Iron	102	101	P/P	90 - 110
Potassium	98.6	97.6	P/P	90 - 110
Sodium	101	100	P/P	90 - 110
Zinc	102	98.4	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E

Run ID: A97815

Included Lab Sample IDs: 2158575, 2158578

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	99.1		P	80 - 120
Alachlor	99.9		P	80 - 120
Ametryn	117		P	80 - 120
Atrazine	96.8		P	80 - 120
Atrazine Desethyl	120		P	80 - 120
Azinphos Methyl	102		P	80 - 120
Butylate	96.4		P	80 - 120
Carbophenothion	98.5		P	80 - 120
Chlorpyrifos Ethyl	97.2		P	80 - 120
Chlorpyrifos Methyl	95.1		P	80 - 120
Cyanazine	103		P	80 - 120
Demeton	99.9		P	80 - 120
Diazinon	104		P	80 - 120
Dimethenamid	100		P	80 - 120
Disulfoton	98.1		P	80 - 120
Dithiopyr	105		P	80 - 120
EPTC	99.2		P	80 - 120
Ethion	94.9		P	80 - 120
Ethoprop	98.5		P	80 - 120
Fenamiphos	97.9		P	80 - 120
Fipronil	97.4		P	80 - 120
Fipronil Desulfinyl	101		P	80 - 120
Fipronil Sulfide	98.4		P	80 - 120
Fipronil Sulfone	95.8		P	80 - 120
Flumioxazin	91.2		P	80 - 120
Fonofos	98.9		P	80 - 120
Hexazinone	101		P	80 - 120
Imazalil	101		P	80 - 120
Malathion	92.3		P	80 - 120
Metalaxyl	100		P	80 - 120
Metolachlor	107		P	80 - 120
Metribuzin	101		P	80 - 120
Mevinphos	85.9		P	80 - 120
Molinate	98.3		P	80 - 120
Norflurazon	104		P	80 - 120
Oxadiazon	109		P	80 - 120
Parathion Ethyl	97.7		P	80 - 120
Parathion Methyl	91.1		P	80 - 120
Pendimethalin	99.2		P	80 - 120
Phorate	100		P	80 - 120
Prodiamine	97.3		P	80 - 120
Prometon	104		P	80 - 120
Prometryn	106		P	80 - 120
Pronamide	97.5		P	80 - 120
Propiconazole	103		P	80 - 120
Sulfentrazone	106		P	80 - 120
Tebuconazole	106		P	80 - 120
Terbufos	99.0		P	80 - 120
Terbuthylazine	101		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B

Run ID: A97817

Included Lab Sample IDs: 2158563, 2158564, 2158566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	60.7	60.5	P/P	60 - 160
Acesulfame K	96.3	67.0	P/P	60 - 160
Endothall	102	96.2	P/P	60 - 160
Endothall	97.1	101	P/P	60 - 160

Reference Method: EPA 8321B

Run ID: A97819

Included Lab Sample IDs: 2158563, 2158564, 2158566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	75.4	71.3	P/P	60 - 160
AMPA	80.1	75.4	P/P	60 - 160
AMPA	81.8	86.0	P/P	60 - 160
Glufosinate	104	98.4	P/P	60 - 160
Glufosinate	109	96.6	P/P	60 - 160
Glufosinate	96.6	96.6	P/P	60 - 160
Glyphosate	103	102	P/P	60 - 160
Glyphosate	108	105	P/P	60 - 160
Glyphosate	113	108	P/P	60 - 160

Reference Method: EPA 8270E

Run ID: A97856

Included Lab Sample IDs: 2158573, 2158574, 2158577

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	99.6		P	80 - 120
Alpha-BHC	101		P	80 - 120
Alpha-Chlordane	99.5		P	80 - 120
Beta-BHC	105		P	80 - 120
Bifenthrin	96.4		P	80 - 120
Carbophenothion	94.2		P	80 - 120
Chlorothalonil	82.3		P	80 - 120
Cis-Nonachlor	98.8		P	80 - 120
Cypermethrin	94.1		P	80 - 120
DDD-p,p'	98.3		P	80 - 120
DDE-p,p'	98.9		P	80 - 120
DDT-p,p'	99.2		P	80 - 120
Delta-BHC	97.9		P	80 - 120
Dicofol	95.8		P	80 - 120
Dieldrin	99.9		P	80 - 120
Endosulfan I	101		P	80 - 120
Endosulfan II	95.2		P	80 - 120
Endosulfan Sulfate	99.6		P	80 - 120
Endrin	99.4		P	80 - 120
Endrin Aldehyde	103		P	80 - 120
Endrin Ketone	101		P	80 - 120
Ethalfuralin	99.9		P	80 - 120
Gamma-BHC	101		P	80 - 120
Gamma-Chlordane	97.6		P	80 - 120
Heptachlor	99.0		P	80 - 120
Heptachlor Epoxide	100		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E

Run ID: A97856

Included Lab Sample IDs: 2158573, 2158574, 2158577

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Hexachlorobenzene	98.5		P	80 - 120
Methoxychlor	98.7		P	80 - 120
Mirex	100		P	80 - 120
Permethrin	94.1		P	80 - 120
Trans-Nonachlor	99.2		P	80 - 120
Trifluralin	99.6		P	80 - 120

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A97880

Included Lab Sample IDs: 2158522, 2158523, 2158527, 2158530, 2158533

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	104	104	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A97908

Included Lab Sample IDs: 2158563, 2158564, 2158566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	107	122	P/P	60 - 160
Sucralose	109	107	P/P	60 - 160

Reference Method: EPA 8276

Run ID: A97915

Included Lab Sample IDs: 2158573, 2158574, 2158577

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	101		P	80 - 120
Chlordane	103		P	80 - 120
Toxaphene	101		P	80 - 120
Toxaphene	58.4*		F	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A97919

Included Lab Sample IDs: 2158579, 2158580, 2158581, 2158582

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Thallium	99.4	99.4	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A97922

Included Lab Sample IDs: 2158579, 2158580, 2158582

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	99.7	100	P/P	90 - 110
Magnesium	102	102	P/P	90 - 110

Quality Assurance Report

Calibration Verification

Reference Method: EPA 8270E

Run ID: A97948

Included Lab Sample IDs: 2158575, 2158576, 2158578

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	96.2		P	80 - 120
Alachlor	98.7		P	80 - 120
Ametryn	98.8		P	80 - 120
Atrazine	95.2		P	80 - 120
Atrazine Desethyl	96.2		P	80 - 120
Azinphos Methyl	97.0		P	80 - 120
Bromacil	98.2		P	80 - 120
Butylate	104		P	80 - 120
Carbophenothion	98.9		P	80 - 120
Chlorpyrifos Ethyl	101		P	80 - 120
Chlorpyrifos Methyl	100		P	80 - 120
Cyanazine	93.2		P	80 - 120
Demeton	99.7		P	80 - 120
Diazinon	96.7		P	80 - 120
Dimethenamid	96.7		P	80 - 120
Disulfoton	98.0		P	80 - 120
Dithiopyr	94.0		P	80 - 120
EPTC	99.9		P	80 - 120
Ethion	98.0		P	80 - 120
Ethoprop	97.7		P	80 - 120
Fenamiphos	102		P	80 - 120
Fipronil	98.4		P	80 - 120
Fipronil Desulfinyl	97.2		P	80 - 120
Fipronil Sulfide	98.9		P	80 - 120
Fipronil Sulfone	97.7		P	80 - 120
Flumioxazin	109		P	80 - 120
Fonofos	99.0		P	80 - 120
Hexazinone	98.6		P	80 - 120
Imazalil	94.1		P	80 - 120
Malathion	98.1		P	80 - 120
Metalaxyl	101		P	80 - 120
Metolachlor	97.4		P	80 - 120
Metribuzin	102		P	80 - 120
Mevinphos	97.6		P	80 - 120
Molinate	103		P	80 - 120
Norflurazon	93.9		P	80 - 120
Oxadiazon	97.7		P	80 - 120
Parathion Ethyl	97.2		P	80 - 120
Parathion Methyl	96.8		P	80 - 120
Pendimethalin	104		P	80 - 120
Phorate	102		P	80 - 120
Prodiamine	96.7		P	80 - 120
Prometon	98.0		P	80 - 120
Prometryn	101		P	80 - 120
Pronamide	92.3		P	80 - 120
Propiconazole	107		P	80 - 120
Simazine	97.4		P	80 - 120
Sulfentrazone	95.5		P	80 - 120
Tebuconazole	97.4		P	80 - 120
Terbufos	99.9		P	80 - 120
Terbutylazine	96.2		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B

Run ID: A97978

Included Lab Sample IDs: 2158563, 2158564, 2158566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	97.2	99.8	P/P	50 - 150
2,4-D	99.8	83.7	P/P	50 - 150
Acetaminophen	96.1	84.8	P/P	60 - 160
Acetaminophen	97.4	96.1	P/P	60 - 160
Acetamiprid	89.7	98.6	P/P	50 - 150
Acetamiprid	98.4	89.7	P/P	50 - 150
Afidopyropen	94.6	91.2	P/P	50 - 150
Afidopyropen	99.0	94.6	P/P	50 - 150
Bentazon	122	110	P/P	50 - 160
Benzovindiflupyr	103	94.3	P/P	50 - 150
Benzovindiflupyr	94.3	104	P/P	50 - 150
Carbamazepine	101	98.5	P/P	60 - 150
Carbamazepine	98.5	99.9	P/P	60 - 150
Clothianidin	93.8	98.8	P/P	60 - 150
Clothianidin	98.8	79.4	P/P	60 - 150
Dinotefuran	103	81.8	P/P	50 - 150
Dinotefuran	81.8	92.7	P/P	50 - 150
Diuron	104	103	P/P	60 - 150
Diuron	93.8	104	P/P	60 - 150
Fenuron	88.4	96.9	P/P	60 - 150
Fenuron	96.9	97.5	P/P	60 - 150
Fluridone	107	109	P/P	60 - 160
Fluridone	109	107	P/P	60 - 160
Hydrocodone	89.3	92.3	P/P	60 - 150
Hydrocodone	92.1	89.3	P/P	60 - 150
Ibuprofen	90.3	91.2	P/P	50 - 160
Ibuprofen	91.2	76.8	P/P	50 - 160
Imazapyr	109	96.0	P/P	50 - 150
Imazapyr	96.0	97.0	P/P	50 - 150
Imidacloprid	101	88.7	P/P	60 - 150
Imidacloprid	88.7	88.8	P/P	60 - 150
Linuron	100	91.6	P/P	60 - 150
Linuron	91.6	97.4	P/P	60 - 150
Mandestrobin	100	101	P/P	50 - 150
Mandestrobin	99.3	100	P/P	50 - 150
MCP	101	104	P/P	50 - 150
MCP	104	98.2	P/P	50 - 150
Naproxen	114	125	P/P	50 - 160
Naproxen	79.6	114	P/P	50 - 160
Primidone	102	91.5	P/P	60 - 150
Primidone	91.5	90.3	P/P	60 - 150
Pyraclostrobin	93.7	98.7	P/P	60 - 150
Pyraclostrobin	94.0	93.7	P/P	60 - 150
Thiamethoxam	104	83.7	P/P	50 - 150
Thiamethoxam	83.7	101	P/P	50 - 150
Tolfenpyrad	92.0	94.5	P/P	60 - 150
Tolfenpyrad	93.2	92.0	P/P	60 - 150
Triclopyr	112	98.1	P/P	50 - 150
Triclopyr	98.1	89.0	P/P	50 - 150

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A97989

Included Lab Sample IDs: 2158581

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	102	99.5	P/P	95 - 105
Magnesium	102	103	P/P	95 - 105

Reference Method: EPA 8321B

Run ID: A98066

Included Lab Sample IDs: 2158563, 2158566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bentazon	109	105	P/P	50 - 160
Bentazon	77.9	130	P/P	50 - 160

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity							0.0	
	Turbidity							0.881	
EPA 200.7 Rev. 4.4	Barium	106		106	105	105			0.640
	Calcium	107		102	109	105			2.88
	Iron	109		110	109	105			3.66
	Magnesium	112		103	116	111			3.50
	Potassium	107		108	108	107			0.422
	Sodium	112		103	106	105			0.981
	Zinc	106		105	105	105			0.495
EPA 200.8 Rev. 5.4	Aluminum	91.9		97.2	96.0	96.9			1.24
	Antimony	104		104	104	107			0.0080
	Arsenic	100		102	101	105			0.201
	Boron	95.9		95.6	95.6	98.1			0.0281
	Cadmium	104		102	101	105			0.446
	Chromium	97.1		99.1	98.6	99.6			0.477
	Copper	95.3		99.0	98.4	97.0			0.656
	Lead	95.4		95.7	95.3	96.7			0.451
	Nickel	94.4		97.3	97.9	95.7			0.588
	Selenium	105		107	106	109			1.73
	Silver	100		100	99.9	101			0.355
	Thallium	99.4		105	106	107			0.647
EPA 300.0 Rev. 2.1	Chloride	96.7		99.6	102			0.696	
	Sulfate	97.9		96.8	101			1.03	
	Sulfate	101		100				0.278	
EPA 350.1 Rev. 2.0	Ammonia-N	97.2		99.1	92.0				4.86
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102		104	100				2.20
EPA 353.2 Rev. 2.0	NO2NO3-N	101		98.2					0.0
EPA 365.1 Rev. 2.0	Total-P	104		103	104				0.615
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.9		95.0	96.8				1.01
	O-Phosphate-P	98.2		92.0	92.8				0.427
EPA 8270E	Acetochlor	99.5	94.5	101	96.3		5.23		4.82
	Alachlor	99.0	94.9	103	97.8		4.30		4.86
	Aldrin	35.3	32.7	44.5	42.2		7.73		5.31
	Alpha-BHC	82.6	74.6	88.4	76.7		10.1		14.2
	Alpha-Chlordane	76.2	71.9	76.3	74.6		5.85		2.20
	Ametryn	103	116	101	98.3		12.1		2.56
	Atrazine	100	101				0.993		0.914
	Atrazine Desethyl	85.0	80.3	60.9	84.6		5.76		32.6
	Azinphos Methyl	133	111	125	122		18.7		2.79
	Beta-BHC	109	98.3	145	133		9.93		9.28
	Bifenthrin	69.9	63.6	66.5	72.3		9.39		8.30
	Bromacil	89.5	99.6	88.2	78.7		10.6		6.53
	Butylate	48.8	60.6	32.5	50.9		21.6		44.1
	Carbophenothion	75.2	74.0	92.9	94.0		1.61		1.15
	Carbophenothion	104	108	110	101		4.56		8.54
	Chlorothalonil	91.4	38.0	93.9	67.1		82.5		33.2
	Chlorpyrifos Ethyl	104	103	123	108		0.380		12.8
	Chlorpyrifos Methyl	97.9	92.9	103	99.9		5.19		3.51
	Cis-Nonachlor	78.4	75.2	78.6	78.1		4.19		0.671
	Cyanazine	127	99.0	106	102		24.5		3.58
	Cypermethrin	88.8	76.5	85.4	83.9		14.9		1.76
	DDD-p,p'	89.8	83.1	90.5	88.1		7.85		2.73
	DDE-p,p'	75.3	72.8	74.6	72.8		3.32		2.49
	DDT-p,p'	69.0	67.7	67.9	66.5		1.90		2.05

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270E	Delta-BHC	113	93.9	140	131	18.7	6.48
	Demeton	78.6	56.8	85.9	87.5	32.1	1.84
	Diazinon	91.9	89.8	97.0	93.7	2.31	3.38
	Dicofol	84.5	79.5	75.8	70.3	6.10	7.42
	Dieldrin	86.9	80.3	91.4	93.7	7.88	2.53
	Dimethenamid	92.4	94.4	100	94.3	2.12	5.93
	Disulfoton	87.6	65.7	90.2	89.5	28.6	0.861
	Dithiopyr	97.4	98.5	99.8	92.7	1.16	6.97
	Endosulfan I	87.9	79.0	89.9	84.1	10.7	6.69
	Endosulfan II	100	84.3	102	94.8	17.1	7.51
	Endosulfan Sulfate	94.9	78.2	100	89.7	19.3	11.3
	Endrin	87.2	81.0	86.9	82.7	7.40	4.98
	Endrin Aldehyde	105	88.3	91.8	81.7	17.0	11.7
	Endrin Ketone	97.4	82.7	107	95.2	16.3	11.4
	EPTC	46.5	57.9	29.1	41.4	21.8	34.9
	Ethalfuralin	69.4	64.4	67.1	61.5	7.52	8.70
	Ethion	98.9	107	121	107	7.62	12.1
	Ethoprop	87.0	79.0	95.7	96.1	9.55	0.412
	Fenamiphos	105	102	100	94.3	2.54	6.01
	Fipronil	87.5	108	109	100	21.2	8.25
	Fipronil Desulfinyl	94.6	107	110	97.4	12.8	12.2
	Fipronil Sulfide	88.8	107	112	98.5	18.9	12.6
	Fipronil Sulfone	95.4	104	108	90.8	8.42	15.0
	Flumioxazin	157	161	177	174	2.02	1.43
	Fonofos	93.4	77.1	93.0	85.8	19.1	8.11
	Gamma-BHC	106	92.3	135	120	13.3	12.2
	Gamma-Chlordane	69.3	67.3	68.0	67.1	2.90	1.30
	Heptachlor	59.6	58.2	55.6	54.3	2.44	2.43
	Heptachlor Epoxide	79.9	73.4	82.7	77.1	8.39	7.11
	Hexachlorobenzene	68.5	63.3	66.4	61.0	7.85	8.48
	Hexazinone	96.5	116			18.1	2.10
	Imazalil	94.4	90.8	66.8	210	3.84	103
	Malathion	109	111	115	104	2.14	7.70
	Metalaxyl	106	101	105	100	5.64	4.64
	Methoxychlor	78.0	75.7	78.3	72.7	3.00	7.44
	Metolachlor	92.3	94.5			2.33	0.666
	Metribuzin	104	104			0.236	5.83
	Mevinphos	79.5	69.5	83.8	86.7	13.5	3.40
	Mirex	58.5	56.5	56.3	57.2	3.53	1.61
	Molinate	66.0	63.7	57.2	68.8	3.45	18.4
	Norflurazon	91.8	109	106	85.4	17.2	14.3
	Oxadiazon	95.2	107	96.3	88.1	11.3	7.41
	Parathion Ethyl	95.4	95.5	108	97.6	0.157	10.3
	Parathion Methyl	95.9	96.4	109	105	0.529	3.53
	Pendimethalin	93.1	94.3	110	101	1.32	8.53
	Permethrin	71.5	65.9	71.8	70.4	8.11	1.99
	Phorate	90.0	75.7	82.6	86.5	17.3	4.55
	Prodiamine	104	93.8	110	76.7	10.4	35.8
	Prometon	94.8	98.6	89.1	93.4	3.91	4.70
	Prometryn	102	103	106	99.8	1.77	6.20
	Pronamide	92.7	104	99.4	108	11.4	7.96
	Propiconazole	99.8	117	112	111	15.9	0.837
	Simazine	101	101	119	107	0.173	10.9
	Sulfentrazone	70.2	99.9	100	90.6	34.9	10.3

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	SMP
EPA 8270E	Tebuconazole	59.8	103	97.4	76.4	53.0	22.2
	Terbufos	92.6	79.5	97.9	98.6	15.2	0.749
	Terbutylazine	96.0	98.0	102	95.1	2.05	6.68
	Trans-Nonachlor	74.5	71.5	74.7	74.3	4.11	0.547
	Trifluralin	69.3	66.3	69.1	64.7	4.42	6.66
EPA 8276	Chlordane	157	98.2	97.3	95.1	45.8	2.32
	Toxaphene	76.9	82.6	76.5	82.0	7.22	6.90
EPA 8321B	2,4-D	133		131	142		8.66
	3-Hydroxycarbofuran	88.6		82.9	83.4		0.617
	Acesulfame K	95.5		79.3	75.2		4.80
	Acetaminophen	88.1		92.5	114		20.6
	Acetamiprid	82.0		78.4	57.3		31.1
	Afidopyropen	71.5		87.0	80.3		8.04
	Aldicarb	74.0		64.9	78.2		18.6
	Aldicarb Sulfone	95.0		103	101		2.41
	Aldicarb Sulfoxide	96.0		44.3	45.6		3.00
	AMPA	98.1		83.7	77.7		7.50
	Bentazon	76.6					21.0
	Benzovindiflupyr	78.1		75.4	82.7		9.23
	Carbamazepine	80.5		66.9	70.5		5.32
	Carbaryl	61.9		61.4	50.0		20.4
	Carbofuran	70.2		69.0	67.6		2.07
	Clothianidin	70.9		92.3	104		12.3
	Dinotefuran	79.0		119	120		0.777
	Diuron	78.7		63.8	65.4		2.51
	Endothall	86.8		106	97.5		8.73
	Fenuron	91.3		77.4	71.1		8.43
	Fluridone	92.4		88.0	90.0		1.75
	Glufosinate	95.0		195	246		23.2
	Glyphosate	93.1		183	213		15.1
	Hydrocodone	82.6		78.4	95.1		19.2
	Ibuprofen	86.7		79.4	82.2		3.35
	Imazapyr	90.1					2.90
	Imidacloprid	89.6		157	138		12.8
	Linuron	73.5		75.0	71.5		4.70
	Mandestrobin	87.7		84.9	88.5		4.07
	MCPP	131		133	134		0.847
	Methiocarb	73.0		63.7	58.9		7.82
	Methomyl	85.6		90.0	77.8		14.5
	Naproxen	83.9		97.3	117		18.7
	Oxamyl	92.9		96.1	98.0		2.01
	Primidone	79.7		85.7	88.2		2.92
	Propoxur	71.7		66.7	62.4		5.32
	Pyraclostrobin	82.3		83.5	88.0		5.19
	Sucralose	113		108	101		5.72
	Thiamethoxam	86.9		108	121		11.2
	Tolfenpyrad	58.6		57.1	57.5		0.759
	Triclopyr	105		115	130		11.8
SM 2120 B-2011	Color (true)	100					1.25
	Color (true)	101					0.993
SM 2320 B-2011	Total Alkalinity	104					0.288
SM 2540 C-2011	TDS						1.87
SM 2540 D-2011	TSS						8.00
SM 4500 F-C-2011	Fluoride	99.0		98.8	99.3		0.477

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	Precision		
				LCS	SMP	MS
SM 5310 B-2011	Organic Carbon	99.0	99.8 101			0.707

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2158520, 2158521, 2158526, 2158529, 2158532
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2158579, 2158580, 2158581, 2158582
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2158579, 2158580, 2158581, 2158582
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2158520, 2158521, 2158526, 2158529, 2158532
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2158520, 2158521, 2158526, 2158529, 2158532
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2158522, 2158523, 2158527, 2158530, 2158533
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2158522, 2158523, 2158527, 2158530, 2158533
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2158522, 2158523, 2158527, 2158530, 2158533
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2158522, 2158523, 2158527, 2158530, 2158533
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2158524, 2158525, 2158528, 2158531, 2158534
EPA 8270E	Organochlorine pesticides in water matrices by GC/MS-SIM	2158573, 2158574, 2158577
EPA 8270E	Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS	2158575, 2158576, 2158578
EPA 8276	Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization	2158573, 2158574, 2158577
EPA 8321B	Carbamates in water matrices by HPLC/MS/MS	2158561, 2158562, 2158565
EPA 8321B	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2158563, 2158564, 2158566
EPA 8321B	Water soluble organic compounds by HPLC/MS/MS	2158563, 2158564, 2158566
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2158520, 2158521, 2158526, 2158529, 2158532
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2158520, 2158521, 2158526, 2158529, 2158532
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2158579, 2158580, 2158581, 2158582
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2158520, 2158521, 2158526, 2158529, 2158532
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2158520, 2158521, 2158526, 2158529, 2158532
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2158520, 2158521, 2158526, 2158529, 2158532
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2158522, 2158523, 2158527, 2158530, 2158533

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	02/18/2020			02/18/2020 16:35	Yen Ta	2158520, 2158521, 2158526, 2158529, 2158532
EPA 200.7 Rev. 4.4	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/24/2020 15:53	Martin Acosta	2158579
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/24/2020 16:01	Martin Acosta	2158580
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/24/2020 16:08	Martin Acosta	2158581
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/24/2020 16:15	Martin Acosta	2158582
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 21:18	Betina Topolski	2158579
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 21:25	Betina Topolski	2158580
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 21:40	Betina Topolski	2158582
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/04/2020 12:19	Betina Topolski	2158581
EPA 200.8 Rev. 5.4	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/25/2020 01:18	Justin Cutchin	2158579
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/25/2020 02:43	Justin Cutchin	2158580
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/25/2020 02:51	Justin Cutchin	2158581
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/25/2020 03:00	Justin Cutchin	2158582
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/26/2020 20:17	Justin Cutchin	2158579
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/26/2020 20:59	Justin Cutchin	2158580
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/26/2020 21:04	Justin Cutchin	2158581
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	02/26/2020 21:09	Justin Cutchin	2158582
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 18:15	Alexander Thompson	2158579
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 18:24	Alexander Thompson	2158580
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 18:34	Alexander Thompson	2158581
	02/18/2020	02/20/2020 15:35	Elliott D. Healy	03/02/2020 18:44	Alexander Thompson	2158582
	02/18/2020			02/26/2020 13:09	Elliott Rodriguez	2158532
	02/18/2020			02/26/2020 14:57	Elliott Rodriguez	2158520
	02/18/2020			02/26/2020 15:10	Elliott Rodriguez	2158521
	02/18/2020			02/26/2020 15:22	Elliott Rodriguez	2158526
EPA 300.0 Rev. 2.1	02/18/2020			02/26/2020 15:34	Elliott Rodriguez	2158529
	02/18/2020			02/28/2020 08:44	Elliott Rodriguez	2158520
	02/18/2020			02/28/2020 08:56	Elliott Rodriguez	2158521
	02/18/2020			02/28/2020 09:08	Elliott Rodriguez	2158529
	02/18/2020			02/21/2020 12:01	Ping Hua	2158522
	02/18/2020			02/21/2020 12:11	Ping Hua	2158523
	02/18/2020			02/21/2020 12:13	Ping Hua	2158527
	02/18/2020			02/21/2020 12:14	Ping Hua	2158530
EPA 350.1 Rev. 2.0	02/18/2020			02/21/2020 12:16	Ping Hua	2158533
	02/18/2020	02/21/2020 09:20	Sima Feizi	02/24/2020 17:05	Jhamieka S. Greenwood	2158527
	02/18/2020	02/21/2020 09:20	Sima Feizi	02/24/2020 17:21	Jhamieka S. Greenwood	2158522
	02/18/2020	02/21/2020 09:20	Sima Feizi	02/24/2020 17:23	Jhamieka S. Greenwood	2158523
	02/18/2020	02/21/2020 09:20	Sima Feizi	02/24/2020 17:24	Jhamieka S. Greenwood	2158530
EPA 351.2 Rev. 2.0	02/18/2020	02/21/2020 09:20	Sima Feizi	02/24/2020 17:26	Jhamieka S. Greenwood	2158533

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 353.2 Rev. 2.0	02/18/2020			02/25/2020 08:48	Beverly Y. Sanders	2158522
	02/18/2020			02/25/2020 08:49	Beverly Y. Sanders	2158523
	02/18/2020			02/25/2020 08:50	Beverly Y. Sanders	2158527
	02/18/2020			02/25/2020 08:52	Beverly Y. Sanders	2158530
	02/18/2020			02/25/2020 08:53	Beverly Y. Sanders	2158533
EPA 365.1 Rev. 2.0	02/18/2020	02/27/2020 13:38	Yen Ta	02/28/2020 09:48	Lipi Saha	2158522
	02/18/2020	02/27/2020 13:38	Yen Ta	02/28/2020 09:49	Lipi Saha	2158523
	02/18/2020	02/27/2020 13:38	Yen Ta	02/28/2020 09:50	Lipi Saha	2158527
	02/18/2020	02/27/2020 13:38	Yen Ta	02/28/2020 09:51	Lipi Saha	2158530
	02/18/2020	02/27/2020 13:38	Yen Ta	02/28/2020 09:52	Lipi Saha	2158533
EPA 365.1 Rev. 2.0 dissolved	02/18/2020			02/18/2020 16:42	Carlos Miranda	2158528
	02/18/2020			02/18/2020 16:43	Carlos Miranda	2158531
	02/18/2020			02/18/2020 16:45	Carlos Miranda	2158524
	02/18/2020			02/18/2020 16:50	Carlos Miranda	2158525
EPA 8270E	02/18/2020	02/19/2020 11:00	Fe-Val Siddell	02/22/2020 00:44	Michael Poppell	2158573
	02/18/2020	02/19/2020 11:00	Fe-Val Siddell	02/22/2020 01:16	Michael Poppell	2158574
	02/18/2020	02/19/2020 11:00	Fe-Val Siddell	02/22/2020 01:47	Michael Poppell	2158577
	02/18/2020	02/21/2020 09:00	Fe-Val Siddell	02/24/2020 17:44	Vandana T. Nagar	2158575
	02/18/2020	02/21/2020 09:00	Fe-Val Siddell	02/24/2020 18:09	Vandana T. Nagar	2158578
	02/18/2020	02/21/2020 09:00	Fe-Val Siddell	02/26/2020 18:02	Vandana T. Nagar	2158575
	02/18/2020	02/21/2020 09:00	Fe-Val Siddell	02/26/2020 18:29	Vandana T. Nagar	2158578
	02/18/2020	02/21/2020 09:00	Fe-Val Siddell	02/26/2020 21:45	Vandana T. Nagar	2158576
EPA 8276	02/18/2020	02/19/2020 11:00	Fe-Val Siddell	02/26/2020 07:18	Julie Wi	2158573
	02/18/2020	02/19/2020 11:00	Fe-Val Siddell	02/26/2020 08:14	Julie Wi	2158574
	02/18/2020	02/19/2020 11:00	Fe-Val Siddell	02/26/2020 10:04	Julie Wi	2158577
EPA 8321B	02/18/2020	02/19/2020 09:00	Hoor Shaik	02/21/2020 03:11	Juyoung Kim	2158561
	02/18/2020	02/19/2020 09:00	Hoor Shaik	02/21/2020 04:20	Juyoung Kim	2158562
	02/18/2020	02/19/2020 09:00	Hoor Shaik	02/21/2020 04:55	Juyoung Kim	2158565
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/20/2020 00:51	Rebecca Ware	2158563
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/20/2020 11:47	Rebecca Ware	2158564
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/20/2020 12:01	Rebecca Ware	2158566
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/20/2020 22:34	Rebecca Ware	2158563
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/20/2020 22:55	Rebecca Ware	2158564
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/21/2020 10:07	Rebecca Ware	2158566
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/26/2020 23:55	Rebecca Ware	2158563
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/27/2020 00:19	Rebecca Ware	2158564
	02/18/2020	02/19/2020 16:00	Rebecca Ware	02/27/2020 00:30	Rebecca Ware	2158566
	02/18/2020	02/20/2020 10:00	Hoor Shaik	03/04/2020 08:14	Juyoung Kim	2158563
	02/18/2020	02/20/2020 10:00	Hoor Shaik	03/04/2020 09:24	Juyoung Kim	2158564

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 8321B	02/18/2020	02/20/2020 10:00	Hoor Shaik	03/04/2020 09:58	Juyoung Kim	2158566
	02/18/2020	02/20/2020 10:00	Hoor Shaik	03/06/2020 02:01	Juyoung Kim	2158563
	02/18/2020	02/20/2020 10:00	Hoor Shaik	03/06/2020 10:06	Juyoung Kim	2158566
SM 2120 B-2011	02/18/2020			02/18/2020 17:36	Madeline Funaro	2158520
	02/18/2020			02/18/2020 17:37	Madeline Funaro	2158521
	02/18/2020			02/18/2020 17:38	Madeline Funaro	2158526
	02/18/2020			02/18/2020 17:46	Madeline Funaro	2158529
	02/18/2020			02/18/2020 18:04	Madeline Funaro	2158532
SM 2320 B-2011	02/18/2020			02/22/2020 20:44	Dale L. Simmons	2158520
	02/18/2020			02/22/2020 20:55	Dale L. Simmons	2158521
	02/18/2020			02/22/2020 21:08	Dale L. Simmons	2158529
	02/18/2020			02/22/2020 21:49	Dale L. Simmons	2158532
	02/18/2020			02/22/2020 22:26	Dale L. Simmons	2158526
SM 2340B	02/18/2020			03/02/2020 21:18	Betina Topolski	2158579
	02/18/2020			03/02/2020 21:25	Betina Topolski	2158580
	02/18/2020			03/02/2020 21:40	Betina Topolski	2158582
	02/18/2020			03/04/2020 12:19	Betina Topolski	2158581
SM 2540 C-2011	02/18/2020			02/21/2020 15:21	Yijie Li	2158520, 2158521, 2158526, 2158529, 2158532
SM 2540 D-2011	02/18/2020			02/21/2020 15:21	Yijie Li	2158520, 2158521, 2158526, 2158529, 2158532
SM 4500 F-C-2011	02/18/2020			02/22/2020 20:44	Dale L. Simmons	2158520
	02/18/2020			02/22/2020 20:55	Dale L. Simmons	2158521
	02/18/2020			02/22/2020 21:08	Dale L. Simmons	2158529
	02/18/2020			02/22/2020 21:49	Dale L. Simmons	2158532
	02/18/2020			02/22/2020 22:26	Dale L. Simmons	2158526
SM 5310 B-2011	02/18/2020			02/20/2020 04:28	Madeline Funaro	2158522
	02/18/2020			02/20/2020 04:40	Madeline Funaro	2158523
	02/18/2020			02/20/2020 04:53	Madeline Funaro	2158527
	02/18/2020			02/20/2020 05:05	Madeline Funaro	2158530
	02/18/2020			02/20/2020 05:43	Madeline Funaro	2158533