

Biological Analysis Report

SO-DIST-2020-07-21-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Overflow Analyses Performed By:
Sanders Laboratories
10090 Bavaria Road
Ft. Myers, FL 33913
DOH Accreditation E85457

Event Description: **2020 SMP : Hatchee Cricks**
Request ID: **RQ-2020-03-16-53**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Chris Hormuth

For additional information please contact
Cheryl Swanson - Administrator
Jacqueline Savage - Microbiology
Diane Power - Bench Biology
Lauren Gottfried - Aquatic Toxicology
Puja Jasrotia, Ph.D. - Molecular Biology
Amelia Rankin - Taxonomy (Invertebrate & Algal)
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8177

Certified by: Puja Jasrotia, Environmental Manager

Date Certified: 20-AUG-2020 10:01



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Molecular and Overflow.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: YellowFever@LittletonRd

Collection Date/Time: 07/20/2020 08:40

Field ID: G3SD0149

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2184370	SOP-PCR-1.0	HF183-qPCR**	210	I	GEU/100 mL	P384145	
	SOP-PCR-1.4	BacR-qPCR**	2.0E+03	U	TSC/100 mL	P384145	
	SOP-PCR-1.5	DG3-qPCR**	1.7E+03	U	TSC/100 mL	P384145	
2184377	Colilert-18	Escherichia Coli-Quanti-Tray	1990		MPN/100 mL		

Sample Location: PowellCreek@Bayshore

Collection Date/Time: 07/20/2020 09:00

Field ID: G3SD0165

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2184371	SOP-PCR-1.0	HF183-qPCR**	140	U	GEU/100 mL	P384933	
	SOP-PCR-1.2	GULL2-qPCR**	8.4E+03	U	TSC/100 mL	P384933	LCS
	SOP-PCR-1.3	GFD-purified-qPCR**	170	U	TSC/100 mL	P385132	LCS, MS
	SOP-PCR-1.4	BacR-qPCR**	1.5E+03	U	TSC/100 mL	P384933	
	SOP-PCR-1.5	DG3-qPCR**	1.3E+03	U	TSC/100 mL	P384933	
2184378	Colilert-18	Escherichia Coli-Quanti-Tray	2420	J	MPN/100 mL		

Ref. Method and Comment:
 Colilert-18: All Quanti-Tray wells were positive. Result may be higher than reported.

Sample Location: DaughtreyCreek@Bayshore

Collection Date/Time: 07/20/2020 09:20

Field ID: G3SD0164

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2184372	SOP-PCR-1.0	HF183-qPCR**	94	U	GEU/100 mL	P384933	
	SOP-PCR-1.4	BacR-qPCR**	3.50E+04		TSC/100 mL	P384933	
	SOP-PCR-1.5	DG3-qPCR**	860	U	TSC/100 mL	P384933	
2184379	Colilert-18	Escherichia Coli-Quanti-Tray	105		MPN/100 mL		

Sample Location: PopashCreek@Pritchett

Collection Date/Time: 07/20/2020 09:40

Field ID: G3SD0079

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2184373	SOP-PCR-1.0	HF183-qPCR**	94	U	GEU/100 mL	P384933	
	SOP-PCR-1.2	GULL2-qPCR**	5.6E+03	U	TSC/100 mL	P384933	LCS
	SOP-PCR-1.3	GFD-purified-qPCR**	110	UJ	TSC/100 mL	P385132	LCS, MS
	SOP-PCR-1.4	BacR-qPCR**	1.0E+03	U	TSC/100 mL	P384933	
	SOP-PCR-1.5	DG3-qPCR**	860	U	TSC/100 mL	P384933	
2184380	Colilert-18	Escherichia Coli-Quanti-Tray	2420		MPN/100 mL		

Ref. Method and Comment:
 SOP-PCR-1.3: Quality control failure(s) observed. See QA report for details.

Quality Assurance Report Method Blank Results

Reference Method: SOP-PCR-1.0
 Batch ID: P384145

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report Method Blank Results

Reference Method: SOP-PCR-1.0
Batch ID: P384145

Component	Result	Code	Units
HF183-qPCR	280	U	GEU/100 mL

Reference Method: SOP-PCR-1.0
Batch ID: P384933

Component	Result	Code	Units
HF183-qPCR	280	U	GEU/100 mL

Reference Method: SOP-PCR-1.2
Batch ID: P384933

Component	Result	Code	Units
GULL2-qPCR	1.7E+04	U	TSC/100 mL

Reference Method: SOP-PCR-1.3
Batch ID: P385132

Component	Result	Code	Units
GFD-purified-qPCR	340	U	TSC/100 mL

Reference Method: SOP-PCR-1.4
Batch ID: P384145

Component	Result	Code	Units
BacR-qPCR	3.0E+03	U	TSC/100 mL

Reference Method: SOP-PCR-1.4
Batch ID: P384933

Component	Result	Code	Units
BacR-qPCR	3.0E+03	U	TSC/100 mL

Reference Method: SOP-PCR-1.5
Batch ID: P384145

Component	Result	Code	Units
DG3-qPCR	2.6E+03	U	TSC/100 mL

Reference Method: SOP-PCR-1.5
Batch ID: P384933

Component	Result	Code	Units
DG3-qPCR	2.6E+03	U	TSC/100 mL

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SOP-PCR-1.0
Batch ID: P384145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SOP-PCR-1.0
Batch ID: P384145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
HF183-qPCR	114	116	P/P	50 - 175

Reference Method: SOP-PCR-1.0
Batch ID: P384933

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
HF183-qPCR	129	117	P/P	50 - 175

Reference Method: SOP-PCR-1.2
Batch ID: P384933

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
GULL2-qPCR	94.7	21.0	P/F	50 - 300

Reference Method: SOP-PCR-1.3
Batch ID: P385132

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
GFD-purified-qPCR	106	12.7	P/F	50 - 175

Reference Method: SOP-PCR-1.4
Batch ID: P384145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
BacR-qPCR	107	117	P/P	50 - 175

Reference Method: SOP-PCR-1.4
Batch ID: P384933

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
BacR-qPCR	98.5	89.2	P/P	50 - 175

Reference Method: SOP-PCR-1.5
Batch ID: P384145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
DG3-qPCR	93.6	93.2	P/P	50 - 175

Reference Method: SOP-PCR-1.5
Batch ID: P384933

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
DG3-qPCR	96.1	168	P/P	50 - 175

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SOP-PCR-1.0
Batch ID: P384145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SOP-PCR-1.0

Batch ID: P384145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2180711	HF183-qPCR	95.9	96.3	P	50 - 175
2180712	HF183-qPCR	96.9		P	50 - 175
2180713	HF183-qPCR	111		P/P	50 - 175
2181430	HF183-qPCR	105		P	50 - 175
2181431	HF183-qPCR	97.4	96.3	P	50 - 175
2181555	HF183-qPCR	90.8		P	50 - 175
2181556	HF183-qPCR	89.3		P	50 - 175
2181933	HF183-qPCR	98.5		P	50 - 175
2182342	HF183-qPCR	104		P	50 - 175
2183113	HF183-qPCR	107		P	50 - 175
2184015	HF183-qPCR	106		P	50 - 175
2184016	HF183-qPCR	104		P	50 - 175
2184370	HF183-qPCR	102		P	50 - 175

Reference Method: SOP-PCR-1.0

Batch ID: P384933

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2184371	HF183-qPCR	103	108	P/P	50 - 175
2184372	HF183-qPCR	96.5		P	50 - 175
2184373	HF183-qPCR	85.3		P	50 - 175
2184576	HF183-qPCR	148		P	50 - 175
2184832	HF183-qPCR	98.3	108	P	50 - 175
2184833	HF183-qPCR	96.6		P	50 - 175
2184834	HF183-qPCR	101		P	50 - 175
2185504	HF183-qPCR	86.7		P	50 - 175
2186533	HF183-qPCR	108		P	50 - 175
2186840	HF183-qPCR	94.8		P	50 - 175
2186923	HF183-qPCR	104		P	50 - 175
2186924	HF183-qPCR	103		P	50 - 175
2187410	HF183-qPCR	131		P	50 - 175

Reference Method: SOP-PCR-1.2

Batch ID: P384933

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2184371	GULL2-qPCR	93.4	93.3	P/P	50 - 300
2184373	GULL2-qPCR	68.8		P	50 - 300
2184576	GULL2-qPCR	89.7		P	50 - 300
2185504	GULL2-qPCR	50.7		P	50 - 300
2186533	GULL2-qPCR	73.8	93.3	P	50 - 300
2186923	GULL2-qPCR	71.7		P	50 - 300
2186924	GULL2-qPCR	52.1		P	50 - 300

Reference Method: SOP-PCR-1.3

Batch ID: P385132

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2181555	GFD-purified-qPCR	94.1	93.3	P	50 - 175
2181556	GFD-purified-qPCR	55.0		P	50 - 175
2183113	GFD-purified-qPCR	106		P	50 - 175
2184015	GFD-purified-qPCR	80.4		P	50 - 175

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SOP-PCR-1.3

Batch ID: P385132

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2184016	GFD-purified-qPCR	89.2	115	P	50 - 175
2184371	GFD-purified-qPCR	94.4		P/P	50 - 175
2184373	GFD-purified-qPCR	39.2		F	50 - 175
2184576	GFD-purified-qPCR	104		P	50 - 175
2185504	GFD-purified-qPCR	80.3		P	50 - 175
2186533	GFD-purified-qPCR	57.6		P	50 - 175
2186923	GFD-purified-qPCR	56.2		P	50 - 175
2186924	GFD-purified-qPCR	58.4		P	50 - 175

Reference Method: SOP-PCR-1.4

Batch ID: P384145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2180711	BacR-qPCR	77.3	87.1	P	50 - 175
2180712	BacR-qPCR	81.1		P	50 - 175
2180713	BacR-qPCR	94.8		P/P	50 - 175
2181430	BacR-qPCR	79.2		P	50 - 175
2181431	BacR-qPCR	75.5		P	50 - 175
2181555	BacR-qPCR	80.0		P	50 - 175
2181556	BacR-qPCR	72.5		P	50 - 175
2181933	BacR-qPCR	77.9		P	50 - 175
2182342	BacR-qPCR	82.7		P	50 - 175
2183113	BacR-qPCR	83.6		P	50 - 175
2184015	BacR-qPCR	86.9		P	50 - 175
2184016	BacR-qPCR	76.2		P	50 - 175
2184370	BacR-qPCR	84.8		P	50 - 175

Reference Method: SOP-PCR-1.4

Batch ID: P384933

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2184371	BacR-qPCR	82.7	86.0	P/P	50 - 175
2184372	BacR-qPCR	70.6		P	50 - 175
2184373	BacR-qPCR	72.7		P	50 - 175
2184576	BacR-qPCR	86.9		P	50 - 175
2184832	BacR-qPCR	70.1		P	50 - 175
2184833	BacR-qPCR	77.8		P	50 - 175
2184834	BacR-qPCR	80.4		P	50 - 175
2185504	BacR-qPCR	59.2		P	50 - 175
2186533	BacR-qPCR	69.3		P	50 - 175
2186840	BacR-qPCR	54.2		P	50 - 175
2186923	BacR-qPCR	79.9		P	50 - 175
2186924	BacR-qPCR	69.3		P	50 - 175
2187410	BacR-qPCR	75.1		P	50 - 175

Reference Method: SOP-PCR-1.5

Batch ID: P384145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2180711	DG3-qPCR	80.8	81.7	P	50 - 175
2180712	DG3-qPCR	82.3		P	50 - 175
2180713	DG3-qPCR	84.0		P/P	50 - 175

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SOP-PCR-1.5
Batch ID: P384145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2181430	DG3-qPCR	80.4		P	50 - 175
2181431	DG3-qPCR	114		P	50 - 175
2181555	DG3-qPCR	80.8		P	50 - 175
2181556	DG3-qPCR	84.3		P	50 - 175
2181933	DG3-qPCR	83.6		P	50 - 175
2182342	DG3-qPCR	82.5		P	50 - 175
2183113	DG3-qPCR	81.8		P	50 - 175
2184015	DG3-qPCR	78.5		P	50 - 175
2184016	DG3-qPCR	99.4		P	50 - 175
2184370	DG3-qPCR	87.1		P	50 - 175

Reference Method: SOP-PCR-1.5
Batch ID: P384933

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2184371	DG3-qPCR	102	106	P/P	50 - 175
2184372	DG3-qPCR	95.1		P	50 - 175
2184373	DG3-qPCR	97.4		P	50 - 175
2184576	DG3-qPCR	82.0		P	50 - 175
2184832	DG3-qPCR	106		P	50 - 175
2184833	DG3-qPCR	82.1		P	50 - 175
2184834	DG3-qPCR	95.5		P	50 - 175
2185504	DG3-qPCR	71.0		P	50 - 175
2186533	DG3-qPCR	96.0		P	50 - 175
2186840	DG3-qPCR	72.7		P	50 - 175
2186923	DG3-qPCR	71.3		P	50 - 175
2186924	DG3-qPCR	90.0		P	50 - 175
2187410	DG3-qPCR	85.2		P	50 - 175

Quality Assurance Report Precision

Reference Method: SOP-PCR-1.0
Batch ID: P384145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2180713	HF183-qPCR	14.3	Spike	P	0 - 25

Reference Method: SOP-PCR-1.0
Batch ID: P384933

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2184371	HF183-qPCR	4.78	Spike	P	0 - 25

Reference Method: SOP-PCR-1.2
Batch ID: P384933

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: SOP-PCR-1.2
Batch ID: P384933

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2184371	GULL2-qPCR	0.102	Spike	P	0 - 25

Reference Method: SOP-PCR-1.3
Batch ID: P385132

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2184371	GFD-purified-qPCR	19.5	Spike	P	0 - 25

Reference Method: SOP-PCR-1.4
Batch ID: P384145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2180713	BacR-qPCR	8.43	Spike	P	0 - 25

Reference Method: SOP-PCR-1.4
Batch ID: P384933

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2184371	BacR-qPCR	3.93	Spike	P	0 - 25

Reference Method: SOP-PCR-1.5
Batch ID: P384145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2180713	DG3-qPCR	2.78	Spike	P	0 - 25

Reference Method: SOP-PCR-1.5
Batch ID: P384933

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2184371	DG3-qPCR	3.07	Spike	P	0 - 25

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2184370
Field Sample ID: G3SD0149

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
SOP-PCR-1.0	SKETA-qPCR	79.7	P	50 - 300
SOP-PCR-1.4	SKETA-qPCR	102	P	50 - 300
SOP-PCR-1.5	SKETA-qPCR	93.2	P	50 - 300

Quality Assurance Report Surrogates

Lab Sample ID: 2184371
Field Sample ID: G3SD0165

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
SOP-PCR-1.0	SKETA-qPCR	99.3	P	50 - 300
SOP-PCR-1.0	SKETA-qPCR	103	P	50 - 300
SOP-PCR-1.2	SKETA-qPCR	103	P	50 - 300
SOP-PCR-1.2	SKETA-qPCR	103	P	50 - 300
SOP-PCR-1.4	SKETA-qPCR	102	P	50 - 300
SOP-PCR-1.4	SKETA-qPCR	103	P	50 - 300
SOP-PCR-1.5	SKETA-qPCR	94.9	P	50 - 300
SOP-PCR-1.5	SKETA-qPCR	95.5	P	50 - 300

Lab Sample ID: 2184372
Field Sample ID: G3SD0164

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
SOP-PCR-1.0	SKETA-qPCR	87.0	P	50 - 300
SOP-PCR-1.4	SKETA-qPCR	88.6	P	50 - 300
SOP-PCR-1.5	SKETA-qPCR	79.6	P	50 - 300

Lab Sample ID: 2184373
Field Sample ID: G3SD0079

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
SOP-PCR-1.0	SKETA-qPCR	71.6	P	50 - 300
SOP-PCR-1.2	SKETA-qPCR	63.0	P	50 - 300
SOP-PCR-1.4	SKETA-qPCR	75.7	P	50 - 300
SOP-PCR-1.5	SKETA-qPCR	56.3	P	50 - 300

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			Precision
				LCS	SMP	MS	
SOP-PCR-1.0	HF183-qPCR	116	114	95.9	96.9	111	14.3
	HF183-qPCR			96.3			
SOP-PCR-1.2	GULL2-qPCR	117	129	103	108	96.5	4.78
	GULL2-qPCR			85.3			
SOP-PCR-1.3	GFD-purified-qPCR	21.0	94.7	93.4	93.3	68.8	0.102
	GFD-purified-qPCR			89.7			
SOP-PCR-1.4	BacR-qPCR	12.7	106	94.1	55.0	106	19.5
	BacR-qPCR			80.4			
SOP-PCR-1.5	DG3-qPCR	117	107	77.3	81.1	94.8	8.43
	DG3-qPCR			87.1			
SOP-PCR-1.5	DG3-qPCR	89.2	98.5	82.7	86.0	70.6	3.93
	DG3-qPCR			72.7			
SOP-PCR-1.5	DG3-qPCR	93.2	93.6	80.8	82.3	84.0	2.78
	DG3-qPCR			81.7			
SOP-PCR-1.5	DG3-qPCR	168	96.1	102	106	95.1	3.07
	DG3-qPCR			97.4			

Reference Method Descriptions

Method	Description	Associated Samples
Colilert-18	E. coli by Colilert-18/QT by Sanders in Fort Myers	2184377, 2184378, 2184379, 2184380

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
SOP-PCR-1.0	Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction	2184370, 2184371, 2184372, 2184373
SOP-PCR-1.2	Detection and quantification of coastal bird specific Catellicoccus marimammalium Gull2 genetic marker with quantitative polymerase chain reaction	2184371, 2184373
SOP-PCR-1.3	Detection and quantification of bird specific Helicobacter GFD genetic marker with quantitative polymerase chain reaction	2184371, 2184373
SOP-PCR-1.4	Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction	2184370, 2184371, 2184372, 2184373
SOP-PCR-1.5	Detection and quantification of canine-specific DG3 Bacteroides genetic marker with quantitative polymerase chain reaction	2184370, 2184371, 2184372, 2184373

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
Colilert-18	07/20/2020	07/20/2020 14:39		07/21/2020 09:14	Sanders-Fort Myers	2184377, 2184378, 2184379, 2184380
SOP-PCR-1.0	07/21/2020	07/21/2020 13:00	Lauren Campbell	07/23/2020 14:27	Lauren Campbell	2184370
	07/21/2020	07/21/2020 13:00	Mailin Lopez	08/04/2020 14:04	Lauren Campbell	2184371, 2184372, 2184373
SOP-PCR-1.2	07/21/2020	07/21/2020 13:00	Mailin Lopez	08/07/2020 11:44	Mailin Lopez	2184371, 2184373
SOP-PCR-1.3	07/21/2020	07/21/2020 13:00	Lauren Campbell	08/07/2020 15:28	Mailin Lopez	2184371, 2184373
SOP-PCR-1.4	07/21/2020	07/21/2020 13:00	Lauren Campbell	07/24/2020 11:56	Lauren Campbell	2184370
	07/21/2020	07/21/2020 13:00	Mailin Lopez	08/05/2020 10:52	Lauren Campbell	2184371, 2184372, 2184373
SOP-PCR-1.5	07/21/2020	07/21/2020 13:00	Lauren Campbell	07/24/2020 13:30	Lauren Campbell	2184370
	07/21/2020	07/21/2020 13:00	Mailin Lopez	08/05/2020 13:58	Lauren Campbell	2184371, 2184372, 2184373