From:	Jeremy Bosso
To:	Bunch, Cheryl; DEP-Overflowmicro; King, Michael
Subject:	microbiology report for RQ-2020-07-27-18 samples collected 07/27/2020
Date:	Tuesday, July 28, 2020 5:24:15 PM
Attachments:	FF_2020_07_27A.txt STF_2020_07_27A.pdf DEP_Reports_2020.xlsx

---

Good afternoon. I've attached the microbiology report for RQ-2020-07-27-18 samples collected 07/27/2020.

Samples analyzed by **Enterolert (QT)** were incubated from 07/27/2020 (14:40) to 07/28/2020 (15:40).

The sample analyzed by **SM 9223B (Colilert-18 QT)** was incubated from 07/27/2020 (15:10) to 07/28/2020 (11:45).

Note: The "J2" qualifier on sample 20-46825 (site G4NW0401) is due to our lab duplicate for that sample being outside of our internal precision acceptance criteria. This is most likely due to random error and because the duplicate results were both such low concentrations (5.2 versus 1.0 MPN/100 mL before dilution factor). We always report the higher number of duplicate analyses.

Thank you,

Jeremy Bosso, MPH
Technical Director, Wetlands Research Laboratory
University of West Florida
11000 University Pkwy
Pensacola, FL 32514
(850) 857-6096

Confidentiality Notice: This e-mail communication and any attachments may contain confidential and privileged information for the use of the designated recipient(s) named above only and is not to be passed on to other persons, unless otherwise stated. If you are not the intended recipient, you are hereby notified that you have received this communication in error and that any review, disclosure, dissemination, distribution or copying of it or its contents is prohibited. If you have received this communication in error, please delete the communication and ***notify the sender that the message was received in error***.