

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 153484 RQ- 2020-07-06-240 Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7/20

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jon Woods	7/15/20	755	22.4	757.8	8.6	8.59	99.7	—	1.03	P/F	L/F
ICV	"	"	805	22.5	757.9	8.6	8.55	98.6	—	—	P/F	L/F
CCV	"	"	1341	21.2	756.2	8.9	8.86	99.8	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Jon Woods	7/15/20	757	191120D	12/2020	1000	1000	P/F	L/F
ICV	"	"	803	2908C81	8/2021	100	102.6	P/F	L/F
CCV	"	"	1344	"	"	100	99.4	P/F	L/F
CCV	"	"	1345	191120D	12/2020	1000	988	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU 7.02	mV	Pass / Fail	Lab / Field
Calibr.	Jon Woods	7/15/20	756	200407B	10/2021	7.01	22.0	-23.9	-23.9	P/F	L/F
Calibr.	"	"	759	191120A	6/2021	4.00	21.7	4.00	145.3	P/F	L/F
Calibr.	"	"	800	191120C	6/2021	10.03	22.0	10.04	-190.0	P/F	L/F
ICV	"	"	802	"	"	10.04	21.3	10.08	-195.0	P/F	L/F
CCV	"	"	1347	191120A	6/2021	4.00	22.7	4.01	145.3	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 166.1, slope from 4 to 7 169.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 7/20

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Jon woods	7/15/20	801	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: