

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: 1534564

RQ: 2020-08-24-40

Project: SMP

Boldly "X" this box if there are qualified data on this page.

Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/7/2020

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jon Woods	10/12/20	842	16.1	755.4	9.4	9.42	99.4	-	1.06	P/F	L/F
ICV	"	"	854	20.2	755.6	9.0	9.00	98.6	-	-	P/F	L/F
CCV	"	"	1400	20.9	755.1	8.9	9.03	101.1	-	-	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Jon Woods	10/12/20	844	191120D	12/2020	1000	1000	P/F	L/F
ICV	"	"	851	2911C93	11/2021	100	100.2	P/F	L/F
CCV	"	"	1403	191120D	12/2020	1000	998	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Jon Woods	10/12/20	845	200407B	10/2021	7.02	19.8	7.02	-44.5	P/F	L/F
Calibr.	"	"	847	191120A	6/2021	4.00	19.6	4.00	128.0	P/F	L/F
Calibr.	"	"	848	191120C	6/2021	10.05	19.9	10.06	-207.9	P/F	L/F
ICV	"	"	850	"	"	10.06	19.4	10.03	-207.5	P/F	L/F
CCV	"	"	1404	191120A	6/2021	4.00	20.1	4.03	129.4	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
If mV are recorded: slope from 7 to 10 163.4, slope from 4 to 7 172.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)
If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/7/20

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Jon Woods	10/12/20	849	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: