

Body: "X" in this box if there are qualified data on this page.

Meter ID: 153484 RQ: 2820-09-14-43 Project: 2820-SMP

Body: "X" in this box if there are qualified data on this page.

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400									
DO	Name	Date	Time	Temp	Baro	D.O.	Chart	Meter	% DO
DEP SOP				°C	mmHg	mg/L			
FT 1500									
Calibr.	Ben Hazen	10/22/20	07:51	20.3	1556	9.0	9.03	99.9	1.120
ICV	"	"	08:05	19.8	758.4	9.1	9.13	99.9	0.101
CCV	"	"	14:00	23.5	757.8	8.5	8.33	102.7	0.101
CCV	"	"							P / F

DO Acceptance criteria from Table 1: ± 0.3 mg/L
Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.									
Spec.	Name	Date	Time	Lot #	Expir.	Standard	Meter	Pass	Lab
Condi.					Date	umhos/cm	Reading	/	
FT 1200									
Calibr.	Ben Hazen	10/22/20	07:54	1911200	12/7/20	1000	101.0	0.101	0.101
ICV	"	"	07:55	2911098	11/2/21	100.0	104.0	0.101	0.101
CCV	"	"	14:03	1911200	12/7/20	1000	97.5	0.101	0.101
CCV	"	"							P / F

Conductivity Acceptance criteria $\pm 5\%$

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range ± 50 ; mV pH 4 Range $\pm 180 \pm 50$;									
pH	Name	Date	Time	Lot #	Expir.	pH	Temp	Meter	Pass
DEP SOP					Date	Buffer	°C	reading	/
FT 1100						SU		SU	
Calibr.	Ben Hazen	10/22/20	07:58	2004078	10/21	7.01	20.0	7.02	40.6
Calibr.	"	"	08:00	1911200	06/21	4.00	19.3	4.00	134.9
Calibr.	"	"	08:02	1911200	06/21	10.06	19.6	10.06	206.0
ICV	"	"	08:03	"	"	10.06	19.5	10.03	206.7
CCV	"	"	14:04	1911200	06/21	4.01	20.5	4.10	126.8
CCV	"	"							P / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range ± 50 ; mV pH 4 Range $\pm 180 \pm 50$;
If mV are recorded: slope from 7 to 10 16.0 ; slope from 4 to 7 17.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)
If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 07/20
If NO, what will be used? Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)				
Name	Date	Time	Calibrated	ICV Value,
		CT(°F)	Value (0.00 or Offset), meters	meters
Pressure mode in air	Ben Hazen	10/22/20	07:44	0.000
Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.				

COMMENTS: