

Chemical Analysis Report

TLH-ROC-2020-09-24-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Run 2**
Request ID: **RQ-2020-07-13-30**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 14-OCT-2020 14:21



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Middle Warrior Creek

Collection Date/Time: 09/24/2020 10:40

Field ID: G1TLHR0007

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2198084	EPA 180.1 Rev. 2.0	Turbidity	0.80		NTU	P387816	
	EPA 300.0 Rev. 2.1	Chloride	4.1		mg Cl/L	P388092	
		Sulfate	0.20	U	mg SO4/L	P388096	
	SM 2120 B-2011	Color (true)	820		PCU	P387817	
	SM 2320 B-2011	Total Alkalinity	9.2		mg CaCO3/L	P388086	
	SM 2540 C-2011	TDS	119	J	mg/L	P388195	
	SM 2540 D-2011	TSS	2	U	mg/L	P388119	
2198085	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P387942	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.019		mg N/L	P387893	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P387979	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P387912	
	EPA 365.1 Rev. 2.0	Total-P	0.075		mg P/L	P387968	MS
2198086	SM 5310 B-2011	Organic Carbon	55		mg C/L	P387982	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.042		mg P/L	P387811	

Ref. Method and Comment:

SM 2540 C-2011: A constant dry weight was not attained.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387816

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387893

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387979

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387912

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387968

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387811

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P387817

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P388086

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P388195

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P388119

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P387942

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P387982

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387893

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387979

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	92.5		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387912

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387968

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	108		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387811

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.4		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P387817

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	96.5		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P388086

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.8		P	93 - 104

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P387942

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P387982

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	102		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197720	Chloride	103		P	90 - 110
2197931	Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197720	Sulfate	98.9		P	90 - 110
2197931	Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387893

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2198190	Ammonia-N	100	98.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387979

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2198099	Kjeldahl Nitrogen	99.2	88.0	P/F	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387912

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197935	NO2NO3-N	92.8	92.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387968

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197883	Total-P	112	113	F/F	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387811

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197930	O-Phosphate-P	97.4	97.1	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P387942

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197928	Fluoride	99.6	100	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P387982

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197863	Organic Carbon	100	99.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387816

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197852	Turbidity	0.725	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197720	Chloride	0.719	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197720	Sulfate	0.320	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387893

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2198190	Ammonia-N	1.89	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387979

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2198099	Kjeldahl Nitrogen	8.13	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387912

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197935	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387968

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197883	Total-P	0.634	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387811

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197930	O-Phosphate-P	0.145	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
 Batch ID: P387817

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197898	Color (true)	0.223	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
 Batch ID: P388086

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2198744	Total Alkalinity	1.35	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
 Batch ID: P388195

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197898	TDS	2.46	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
 Batch ID: P387942

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197928	Fluoride	0.467	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
 Batch ID: P387982

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197863	Organic Carbon	0.401	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A101184

Included Lab Sample IDs: 2198086

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.2	98.1	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A101185

Included Lab Sample IDs: 2198084

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A101186

Included Lab Sample IDs: 2198084

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.5	99.1	P/P	90 - 115

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A101212

Included Lab Sample IDs: 2198085

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.6	99.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A101223

Included Lab Sample IDs: 2198085

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	97.0	97.0	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A101239

Included Lab Sample IDs: 2198084

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	104	104	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A101262

Included Lab Sample IDs: 2198085

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.6	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A101281

Included Lab Sample IDs: 2198084

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	107	107	P/P	90 - 110
Sulfate	105	105	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A101288

Included Lab Sample IDs: 2198085

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	91.3	95.4	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A101307

Included Lab Sample IDs: 2198084

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	95.8	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101309

Included Lab Sample IDs: 2198085

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	104	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					0.725	
EPA 300.0 Rev. 2.1	Chloride	105	103	103		0.719	
	Sulfate	103	98.9	100		0.320	
EPA 350.1 Rev. 2.0	Ammonia-N	98.8	100	98.4			1.89
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	92.5	99.2	88.0			8.13
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	98.5	92.8	92.8			0.0
EPA 365.1 Rev. 2.0	Total-P	108	112	113			0.634
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.4	97.4	97.1			0.145
SM 2120 B-2011	Color (true)	96.5				0.223	
SM 2320 B-2011	Total Alkalinity	98.8				1.35	
SM 2540 C-2011	TDS					2.46	
SM 4500 F-C-2011	Fluoride	101	99.6	100			0.467
SM 5310 B-2011	Organic Carbon	102	100	99.5			0.401

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2198084
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2198084
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2198084
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2198085
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2198085
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2198085
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2198085
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2198086
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2198084
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2198084
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2198084
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2198084
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2198084
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2198085

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/24/2020			09/24/2020 14:51	Yen Ta	2198084
EPA 300.0 Rev. 2.1	09/24/2020			09/30/2020 23:06	Lipi Saha	2198084
EPA 350.1 Rev. 2.0	09/24/2020			09/27/2020 10:46	Ping Hua	2198085
EPA 351.2 Rev. 2.0	09/24/2020	09/29/2020 15:45	David Boose	09/30/2020 12:46	Elliott Rodriguez	2198085
EPA 353.2 Rev. 2.0	09/24/2020			09/28/2020 09:38	Beverly Y. Sanders	2198085
EPA 365.1 Rev. 2.0	09/24/2020	09/29/2020 12:36	Yen Ta	09/30/2020 16:50	Benjamin T. Nordmann	2198085
EPA 365.1 Rev. 2.0 dissolved	09/24/2020			09/24/2020 14:54	Samantha Davila	2198086
SM 2120 B-2011	09/24/2020			09/24/2020 15:57	Benjamin T. Nordmann	2198084
SM 2320 B-2011	09/24/2020			09/30/2020 18:24	Dale L. Simmons	2198084
SM 2540 C-2011	09/24/2020			09/25/2020 18:22	Madeline Gruver	2198084
SM 2540 D-2011	09/24/2020			09/25/2020 18:22	Madeline Gruver	2198084
SM 4500 F-C-2011	09/24/2020			09/25/2020 19:17	Dale L. Simmons	2198084
SM 5310 B-2011	09/24/2020			09/29/2020 06:58	David Boose	2198085