

Chemical Analysis Report

SW-DIST-2020-03-19-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **WBID 1633 - Cont. Mon**
Request ID: **RQ-2020-03-16-55**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Matthew Bearden

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 09-APR-2020 09:33



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: McKay Tidal @ Slow Speed Sign

Collection Date/Time: 03/18/2020 10:50

Field ID: G5SW0173

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2166946	EPA 180.1 Rev. 2.0	Turbidity	6.1		NTU	P381001	
	EPA 300.0 Rev. 2.1	Bromide	63		mg Br/L	P381201	
	SM 2320 B-2011	Total Alkalinity	136		mg CaCO3/L	P381413	
	SM 2540 D-2011	TSS	10		mg/L	P381255	
	SM 4500 F-C-2011	Fluoride	1.1		mg F/L	P381452	
2166947	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P381234	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.51		mg N/L	P381281	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.005	I	mg N/L	P381311	
	EPA 365.1 Rev. 2.0	Total-P	0.054		mg P/L	P381312	
	SM 5310 B-2011	Organic Carbon	3.7		mg C/L	P381091	
2166948	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.011		mg P/L	P381027	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P381001

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P381201

Component	Result	Code	Units
Bromide	0.013	U	mg Br/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P381234

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P381281

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P381311

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P381312

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P381027

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P381413

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-2011
Batch ID: P381255

Component	Result	Code	Units
TSS	3	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-2011
Batch ID: P381452

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P381091

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P381201

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Bromide	104		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P381234

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.7		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P381281

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.4		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P381311

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P381312

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P381027

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	106		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P381413

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 107

Reference Method: SM 4500 F-C-2011
Batch ID: P381452

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.3		P	92 - 104

Reference Method: SM 5310 B-2011
Batch ID: P381091

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	103		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P381201

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166933	Bromide	103	101	P/P	90 - 110
2166937	Bromide	99.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P381234

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166887	Ammonia-N	103	106	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P381281

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166947	Kjeldahl Nitrogen	94.7	95.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P381311

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166878	NO2NO3-N	100	99.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P381312

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166272	Total-P	106	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P381027

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166888	O-Phosphate-P	103	105	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P381452

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2165407	Fluoride	98.4	101	P/P	91 - 105

Reference Method: SM 5310 B-2011
Batch ID: P381091

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2166887	Organic Carbon	99.8	96.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P381001

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166886	Turbidity	2.81	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P381201

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166933	Bromide	0.919	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P381234

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166887	Ammonia-N	2.27	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P381281

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166947	Kjeldahl Nitrogen	0.679	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P381311

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166878	NO2NO3-N	0.468	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P381312

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166272	Total-P	0.0193	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P381027

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166888	O-Phosphate-P	1.62	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P381413

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166937	Total Alkalinity	0.0452	Sample	P	0 - 3.3

Quality Assurance Report Precision

Reference Method: SM 2540 D-2011
Batch ID: P381255

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166886	TSS	3.64	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P381452

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2165407	Fluoride	1.47	Spike	P	0 - 2.0

Reference Method: SM 5310 B-2011
Batch ID: P381091

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2166887	Organic Carbon	2.82	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A98316

Included Lab Sample IDs: 2166946

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	98.5	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A98329

Included Lab Sample IDs: 2166948

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	102	104	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A98371

Included Lab Sample IDs: 2166947

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.4	98.4	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A98406

Included Lab Sample IDs: 2166946

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bromide	102	104	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A98413

Included Lab Sample IDs: 2166947

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A98444

Included Lab Sample IDs: 2166947

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A98451

Included Lab Sample IDs: 2166947

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	95.2	90.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A98480

Included Lab Sample IDs: 2166947

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.5	97.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 2320 B-2011

Run ID: A98490

Included Lab Sample IDs: 2166946

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	102	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A98495

Included Lab Sample IDs: 2166946

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	99.1	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					2.81	
EPA 300.0 Rev. 2.1	Bromide	104	103	101	99.9		0.919
EPA 350.1 Rev. 2.0	Ammonia-N	97.7	103	106			2.27
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.4	94.7	95.5			0.679
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	100	99.9			0.468
EPA 365.1 Rev. 2.0	Total-P	101	106	106			0.0193
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	106	103	105			1.62
SM 2320 B-2011	Total Alkalinity	104				0.0452	
SM 2540 D-2011	TSS					3.64	
SM 4500 F-C-2011	Fluoride	98.3	98.4	101			1.47
SM 5310 B-2011	Organic Carbon	103	99.8	96.7			2.82

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2166946
EPA 300.0 Rev. 2.1	Bromide in aqueous matrices	2166946
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2166947
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2166947
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2166947
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2166947
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2166948
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2166946
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2166946
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2166946
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2166947

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/19/2020			03/19/2020 17:42	Julia Storbeck	2166946
EPA 300.0 Rev. 2.1	03/19/2020			03/24/2020 09:44	Elliott Rodriguez	2166946
EPA 350.1 Rev. 2.0	03/19/2020			03/24/2020 16:27	Ping Hua	2166947
EPA 351.2 Rev. 2.0	03/19/2020	03/25/2020 09:45	Sima Feizi	03/26/2020 09:52	Jhamieka S. Greenwood	2166947
EPA 353.2 Rev. 2.0	03/19/2020			03/26/2020 08:44	Beverly Y. Sanders	2166947
EPA 365.1 Rev. 2.0	03/19/2020	03/26/2020 12:10	Lipi Saha	03/27/2020 10:10	Lipi Saha	2166947
EPA 365.1 Rev. 2.0 dissolved	03/19/2020			03/19/2020 16:01	Carlos Miranda	2166948
SM 2320 B-2011	03/19/2020			03/27/2020 13:59	Dale L. Simmons	2166946
SM 2540 D-2011	03/19/2020			03/23/2020 15:22	Yijie Li	2166946
SM 4500 F-C-2011	03/19/2020			03/27/2020 16:21	Dale L. Simmons	2166946
SM 5310 B-2011	03/19/2020			03/21/2020 07:41	Lauren Lees	2166947