

Chemical Analysis Report

SW-DIST-2020-12-09-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **HAs Wildcat and Horse**
Request ID: **RQ-2020-10-26-02**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Ben Paswater

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 30-DEC-2020 14:09

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Wildcat Slough @ Mosaic S2

Collection Date/Time: 12/08/2020 12:30

Field ID: G3SW0055

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2212027	EPA 180.1 Rev. 2.0	Turbidity	4.7		NTU	P390934	
	EPA 300.0 Rev. 2.1	Chloride	22		mg Cl/L	P391161	
	SM 2120 B-2011	Color (true)	190		PCU	P390947	
	SM 2320 B-2011	Total Alkalinity	32		mg CaCO3/L	P391137	
	SM 2540 C-2011	TDS	272		mg/L	P391518	
	SM 2540 D-2011	TSS	5	I	mg/L	P391523	
	SM 4500 F-C-2011	Fluoride	0.27		mg F/L	P391030	
2212028	EPA 350.1 Rev. 2.0	Ammonia-N	0.015		mg N/L	P391130	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P391555	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.055		mg N/L	P391022	
	EPA 365.1 Rev. 2.0	Total-P	0.46		mg P/L	P391058	
	SM 5310 B-2011	Organic Carbon	25		mg C/L	P391061	
2212029	EPA 300.0 Rev. 2.1 dissolved	Sulfate	120		mg SO4/L	P391259	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.38		mg P/L	P390931	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 300.0 Rev. 2.1 dissolved: Sulfate: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P390934

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P391161

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1 dissolved
Batch ID: P391259

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P391130

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P391555

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P391022

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P391058

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P390931

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P390947

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P391137

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P391518

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P391523

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P391030

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P391061

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P391161

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1 dissolved
Batch ID: P391259

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P391130

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P391555

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P391022

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P391058

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P390931

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P390947

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P391137

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	93 - 104

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P391030

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.6		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P391061

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P391161

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211733	Chloride	96.9		P	90 - 110
2212033	Chloride	95.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1 dissolved
Batch ID: P391259

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2212333	Sulfate	101		P	90 - 110
2212500	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P391130

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211308	Ammonia-N	99.8	99.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P391555

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211964	Kjeldahl Nitrogen	101	95.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P391022

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211954	NO2NO3-N	98.5	95.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P391058

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211737	Total-P	102	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P390931

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211955	O-Phosphate-P	102	104	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P391030

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211963	Fluoride	98.4	96.4	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P391061

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2211954	Organic Carbon	99.5	99.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P390934

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211776	Turbidity	3.21	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P391161

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211733	Chloride	0.195	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P391130

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211308	Ammonia-N	0.145	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P391555

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211964	Kjeldahl Nitrogen	4.34	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P391022

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211954	NO2NO3-N	2.18	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P391058

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211737	Total-P	1.86	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P390931

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211955	O-Phosphate-P	1.17	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P390947

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211953	Color (true)	0.424	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P391137

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2212395	Total Alkalinity	1.88	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P391518

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211720	TDS	2.34	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P391030

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211963	Fluoride	1.43	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P391061

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2211954	Organic Carbon	0.0	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A102455

Included Lab Sample IDs: 2212027

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.1	98.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A102499

Included Lab Sample IDs: 2212029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	102	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A102500

Included Lab Sample IDs: 2212027

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	104	104	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A102508

Included Lab Sample IDs: 2212027

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	104	P/P	85 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A102539

Included Lab Sample IDs: 2212028

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A102541

Included Lab Sample IDs: 2212027

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	101	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A102548

Included Lab Sample IDs: 2212028

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.8	99.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A102578

Included Lab Sample IDs: 2212028

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.7	99.6	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 2320 B-2011

Run ID: A102580

Included Lab Sample IDs: 2212027

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	101	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1 dissolved

Run ID: A102617

Included Lab Sample IDs: 2212029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	100	99.3	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A102697

Included Lab Sample IDs: 2212028

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	101	P/P	85 - 115

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A102729

Included Lab Sample IDs: 2212028

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	105	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					3.21	
EPA 300.0 Rev. 2.1	Chloride	98.7	96.9	95.6		0.195	
EPA 300.0 Rev. 2.1 dissolved	Sulfate	102	101	101			
EPA 350.1 Rev. 2.0	Ammonia-N	99.0	99.8	99.6			0.145
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	101	95.2			4.34
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	98.5	95.0			2.18
EPA 365.1 Rev. 2.0	Total-P	102	102	104			1.86
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	102	104			1.17
SM 2120 B-2011	Color (true)	102				0.424	
SM 2320 B-2011	Total Alkalinity	103				1.88	
SM 2540 C-2011	TDS					2.34	
SM 4500 F-C-2011	Fluoride	95.6	98.4	96.4			1.43
SM 5310 B-2011	Organic Carbon	99.2	99.5	99.5			0.0

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2212027
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2212027
EPA 300.0 Rev. 2.1 dissolved	Sulfate, dissolved, in filtered, aqueous matrices	2212029
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2212028
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2212028
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2212028
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2212028
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2212029
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2212027
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2212027
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2212027
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2212027
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2212027
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2212028

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	12/09/2020			12/09/2020 15:09	Yen Ta	2212027
EPA 300.0 Rev. 2.1	12/09/2020			12/11/2020 17:20	Lipi Saha	2212027
EPA 300.0 Rev. 2.1 dissolved	12/09/2020			12/15/2020 23:00	Lipi Saha	2212029
EPA 350.1 Rev. 2.0	12/09/2020			12/14/2020 17:18	Ping Hua	2212028
EPA 351.2 Rev. 2.0	12/09/2020	12/21/2020 14:46	David Boose	12/22/2020 14:23	Julia Storbeck	2212028
EPA 353.2 Rev. 2.0	12/09/2020			12/11/2020 10:36	Beverly Y. Sanders	2212028
EPA 365.1 Rev. 2.0	12/09/2020	12/11/2020 17:49	Madeline Gruver	12/17/2020 18:12	Benjamin T. Nordmann	2212028
EPA 365.1 Rev. 2.0 dissolved	12/09/2020			12/09/2020 14:49	Blanca Fach	2212029
SM 2120 B-2011	12/09/2020			12/09/2020 17:37	Benjamin T. Nordmann	2212027
SM 2320 B-2011	12/09/2020			12/15/2020 01:47	Dale L. Simmons	2212027
SM 2540 C-2011	12/09/2020			12/14/2020 14:56	Yijie Li	2212027
SM 2540 D-2011	12/09/2020			12/14/2020 15:56	Yijie Li	2212027
SM 4500 F-C-2011	12/09/2020			12/10/2020 20:07	Dale L. Simmons	2212027
SM 5310 B-2011	12/09/2020			12/11/2020 02:38	Madeline Gruver	2212028