

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 12/7/2020
Customer: WAS
RQ: RQ-2020-11-16-08
Collected By: Curtis Musson, Blake Spencer
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Scurlock Creek At Pilgrim Rd	 G3WA0009
Unnamed Creek At Palmview Rd	 G3WA0013

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 12.7.2020
1449 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Unnamed Creek At Palmview Rd **Date:** 12/7/2020 ✓
WIN/Field ID: G3WA0013 ✓ **WBID:** 233A **ENR:** **Latitude:** 30.8122
County: JACKSON **Org ID:** 21FLWQA **Longitude:** -85.4542
Receiving Body of Water: Choctawhatchee - St. Andrew **RQ:** 2020-11-16-08

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Curtis Musson		x		x	x
☒	Blake Spencer	x		x		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☒ Syringe	☐ Pole
Depth measured with: MeterStick	
Secchi Equipment ID:	
Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# <u>EXO #1</u>					
Photos: <u>NA</u>		Flow: <u>FLOW</u>		Tide: _____		Tide Times: High: _____ Low: _____					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other: _____											
SBIO ID Numbers: _____											
Notes: 											

Duplicate Sample

Time Zone: <u>EST</u>	Time: <u>N/A</u>	Field ID: <u>N/A</u>	(WIN ID_DUP)
WIN DMT Activity ID: _____			

Blank

Time Zone: _____	Time: <u>N/A</u>	Field ID: <u>N/A</u>	(Blank Type_DDMMYYYY)
DI Source: <u>N/A</u>		Location: <u>N/A</u>	
DI Type: <u>N/A</u>		Equipment ID: <u>N/A</u>	

LIMS EVENT ID: WAS-SECT-2020-12-07-01 WIN Import ID: 21FLWQA_WAS-SECT-2020-12-07-01_03-05-2021.TXT
Enter Field Parameters, Verify Station ID In LIMS DMT: SBS 12/18/20 DT 3/5/21 QC Station ID, Field Parameters In LIMS DMT: AWM 12/29//2020 ✓
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data In WIN: _____
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2020-11-16-08

Customer: WQAP

Project ID: SMP

Collected By: Curtis Musson, Blake Spencer

Field Report Prepared By: Curtis Musson

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G3WA0013✓

Location: Unnamed Creek At Palmview Rd

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date ✓	Time ✓	DO (mg/L) ✓	DO (%SAT) ✓	Temp. (C°) ✓	pH ✓	Sample Depth (m) ✓	Secchi Disk (m) ✓	Total Depth (m) ✓	Sp COND (µmhos/cm) ✓	Salinity (PPT) ✓
12/7/2020	928 EST	9 J	83 J	11.712	5.17	0.1	0.2 ☒ VOB (S)	0.2	48.9	0.02

Central Lab please use top row for login purposes

EST										
-----	--	--	--	--	--	--	--	--	--	--

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA9354050 Exp: 12-20-20	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD Lot # NA9354060 Exp: 12-20-20	☒ Ice ☒ HNO3*	☒ <2	1	A
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # 1690646 Syringe Lot # 0079900	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	A
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☒ W-E8321-DI ☒ W-E8321-MS	☒ Ice		1	A
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)	☐ Ice			
		☐ Ice			

Name:
Curtis Musson
Blake Spencer

Signature:

Date:

12/7/20

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: Exo #1

RQ: 2020-11-02-05

Project: BMAP & NWBT

Boldly "X" this box if there are quality data on this page

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	12/7/20	0658	20.040	754.1	9.01	9.01	100.3	-	-	P/F	L/F
ICV	Blake Spencer	12/7/20	0703	2.050	754.1	9.01	9.01	100.3	-	-	P/F	L/F
CCV	Blake Spencer	12/7/20	1440	18.135	752.8	9.38	10.44	110.6	-	-	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	12/7/20	0704	191120D	12/2020	1000	1000	P/F	L/F
ICV	Blake Spencer	12/7/20	0705	191120A	12/2020	1000	999	P/F	L/F
CCV	Blake Spencer	12/7/20	1444	2911C93	11/2021	100	103.6	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	12/7/20	0707	200407A	10/2021	7.02	20.01	7.02	-21.2	P/F	L/F
Calibr.	Blake Spencer	12/7/20	0711	200221A	09/2021	4.00	20.00	4.00	152.3	P/F	L/F
Calibr.	Blake Spencer	12/7/20	0716	191120C	06/2021	10.06	20.01	10.06	-199.7	P/F	L/F
ICV	Blake Spencer	12/7/20	0717	191120C	06/2021	10.06	20.12	10.04	-190.0	P/F	L/F
CCV	Blake Spencer	12/7/20	1446	200221A	09/2021	4.00	19.95	4.03	150.1	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Cooler Packing Worksheet for Request: RQ-2020-11-16-08

Little River

Ship Cooler On: 11/12/2020

Customer/Project: WAS-SECT/SMP

Assigned To: SHAIK_A

Requester: Curtis Musson

Priority: 3

850-245-8453

2600 Blair Stone Rd

Tallahassee, FL 32399

Attn: Curtis Musson

Comments:

No preservatives needed.

Group A: Nutrients, E. coli, Metals, Tracers x 6

Packing Notes:

No Fedex Labels

No Trash Bags

No Bubblewrap

Requested Analyses:

Bottle Group: A# of Sites: 6

Container ID: P-1L

Qty: 6

Preservation: ICE

Lot #

00078451

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

W-COLOR

W-F

W-SO4-IC

W-TDS

W-TSS

DescriptionAlkalinity in aqueous matrices as mg CaCO₃/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

True Color measured at 450 nm (not valid for NPDES monitoring)

Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)

Sulfate in aqueous matrices

Total Dissolved Solids in aqueous matrices

Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Container ID: BP-1L

Qty: 6

Preservation: ICE

Lot #

1269688

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-250ML-WO-THI

Qty: 6

Preservation: ICE

Lot # L19247

Description: 250 ml Sterilized no tablet

Analysis

ECOLI-18QT

Description

Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Container ID: BG-500ML

Qty: 6

Preservation: ICE

Lot # 00078033

Description: Brown Glass Bottle 500ML

Affix BLUE dot to container.

Analysis

W-E8321-DI

W-E8321-MS

Description

Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Container ID: P-500ML

Qty: 6

Preservation: HNO3

Lot # 00078385

Description: Plastic Bottle 500 mL

Affix RED nitric acid dot to container.

Analysis

W-HARD

W-ICP

W-ICPMS

Description

Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Container ID: P-500ML

Qty: 6

Preservation: ICE And H2SO4

Lot # 00078385

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 6

Preservation: FILTER-ICE

Lot # 00078144

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Cooler Packed By: NTB Number of Coolers: 2 Date: 11-12-20

Kit must also include:

- ☒ Field Sheets
- ☒ Temperature Control Bottle
- ☒ FedEx Bills, if applicable
- ☒ Plastic Bags/Zip Ties
- ☒ Acid Labels (sets of 24)

If Preservation Included:

ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____

PLEASE RETURN ALL COOLERS FOR RQ-2020-11-16-08