

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 9/28/2020
Customer: WAS
RQ: RQ-2020-09-07-22

Collected By: Blake Spencer, Evelyn Becerra
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Hurricane Creek At Us 90	 G1WA0098
Willachoochee Creek At 161	 G1WA0087
Richlander Creek At Cr 65b	 G1WA0093
Double Branch At Cr159	 G1WA0091
Little River At Highbridge Road	 G1WA0086
Lewis Creek Upstream Of 157	 G1WA0075

RELINQUISHED BY(SIGNATURE): 

DATE/TIME: 9/29/20
16:40

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Little River At Highbridge Road **Date:** 9/28/2020 ✓
WIN/Field ID: G1WA0086 ✓ **WBID:** 424 **ENR:** **Latitude:** 30.5135
County: GADSDEN **Org ID:** 21FLWQA **Longitude:** -84.5232
Receiving Body of Water: Ochlockonee - St. Marks **RQ:** 2020-09-07-22

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Blake Spencer	x		x		
☒	Evelyn Becerra		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☒ Syringe	☐ Pole
Depth measured with: Sonar	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☒ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☐ Normal ☐ Low ☒ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# EXO #1					
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time: N/A	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time: N/A	Field ID: N/A	(Blank Type_DDMMYYYY)
DI Source: N/A		Location: N/A	
DI Type: N/A		Equipment ID: N/A	

LIMS EVENT ID: WAS-SECT-2020-09-28-01 **WIN Import ID:** 21FLWQA_WAS-SECT-2020-09-28-01_03-05-2021.TXT
Enter Field Parameters, Verify Station ID In LIMS DMT: SBS 10/2/2020 **QC Station ID, Field Parameters In LIMS DMT:** AWM 12/29//2020 ✓
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets SBS 10/2/2020 **Migrate Data to WIN:** DT 3/5/21 **Verify Data In WIN:**
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2020-09-07-22

Customer: WQAP

Project ID: SMP

Collected By: Blake Spencer, Evelyn Becerra

Field Report Prepared By: Blake Spencer

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G1WA0086 ✓

Location: Little River At Highbridge Road

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date ✓	Time ✓	DO (mg/L) ✓	DO (%SAT) ✓	Temp. (C°) ✓	pH ✓	Sample Depth (m) ✓	Secchi Disk (m) ✓	Total Depth (m) ✓	Sp COND (µmhos/cm) ✓	Salinity (PPTH) ✓
9/28/2020	14:37 EST	6.45	76	23.6	6.65	0.3	0.5 ☐ VOB (S)	3.3	65.3	0.03

Central Lab please use top row for login purposes

	14:41 EST	6.4	75.6	23.61	6.39	2.8			65.4	0.03
--	--------------	-----	------	-------	------	-----	--	--	------	------

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA9354050 Exp: 12-20-20	☒ Ice ☒ H2SO4*	☒ <2	1	B
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # 9240133 Syringe Lot # 16968882	☒ Ice		1	B
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	B
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	B
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☒ Ice		1	B
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)	☐ Ice			
		☐ Ice			

Name:
Blake Spencer
Evelyn Becerra

Signature:

Date:

9/28/20

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Exo #1

RQ: 2020-09-02-22

Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	9/28/20	713	21.4	753.9	8.79	8.79	99.3	—	—	P/F	L/F
ICV	Blake Spencer	9/28/20	714	21.4	754.0	8.79	8.78	99.3	—	—	P/F	L/F
CCV	Blake Spencer	9/28/20	1714	25.40	752.8	8.15	8.11	98.9	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	9/28/20	0717	191120A	12/20	1000	1000	P/F	L/F
ICV	Blake Spencer	9/28/20	0719	2911C93	11/21	100	103.9	P/F	L/F
CCV	Blake Spencer	9/28/20	1719	191120A	12/20	1000	1000	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	9/28/20	722	200407B	10/21	7.01	22.23	7.01	-19.2	P/F	L/F
Calibr.	Blake Spencer	9/28/20	726	191120A	6/21	4.00	22.40	4.00	155.9	P/F	L/F
Calibr.	Blake Spencer	9/28/20	0728	191120C	6/21	10.03	22.4	10.03	-199.2	P/F	L/F
ICV	Blake Spencer	9/28/20	0731	191120A	6/21	4.00	22.4	4.01	155.4	P/F	L/F
CCV	Blake Spencer	9/28/20	0721	200407B	10/21	7.01	22.6	7.07	-19.9	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Cooler Packing Worksheet for Request: RQ-2020-09-07-22

NWQI - Alligator Creek

Ship Cooler On: 9/3/2020

Customer/Project: WAS-SECT/NWQI

Assigned To: KLUG_K

Requester: Curtis Musson

Priority: 3

850-245-8453
2600 Blair Stone Rd

Tallahassee, FL 32399

Attn: Curtis Musson

Comments:

No preservatives needed.

Group A: Nutrients, Metals, E. coli, Tracers X 2

Group A: Minnow Creek & Alligator Creek Upstream Of SR 273

Group B: Nutrients, E. coli, x 4

Packing Notes:

No FedEx labels

No Trash Bags

No Bubblewrap

Requested Analyses:

Bottle Group: A

of Sites: 2

Container ID: P-1L

Qty: 2 Preservation: ICE

Lot # 00077923

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

W-COLOR

W-F

W-SO4-IC

W-TDS

W-TSS

Description

Alkalinity in aqueous matrices as mg CaCO3/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

True Color measured at 450 nm (not valid for NPDES monitoring)

Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)

Sulfate in aqueous matrices

Total Dissolved Solids in aqueous matrices

Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Container ID: BP-1L

Qty: 2 Preservation: ICE

Lot # 128333G

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-250ML-WO-THI

Qty: 2 Preservation: ICE

Lot # 019303

Description: 250 ml Sterilized no tablet

Analysis

ECOLI-18QT

Description

Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Container ID: BG-500ML

Qty: 2 Preservation: ICE

Lot # 00076950

Description: Brown Glass Bottle 500ML

Affix BLUE dot to container.

Analysis

W-E8321-DI

W-E8321-MS

Description

Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Container ID: P-500ML

Qty: 2 Preservation: HNO3

Lot # 00078041

Description: Plastic Bottle 500 mL

Affix RED nitric acid dot to container.

Analysis

W-HARD

W-ICP

W-ICPMS

Description

Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Container ID: P-500ML

Qty: 2 Preservation: ICE And H2SO4

Lot # 00078041

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 2 Preservation: FILTER-ICE

Lot # 00078144

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group: B

of Sites: 4

Container ID: P-1L

Qty: 4 Preservation: ICE

Lot # 00077923

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

W-COLOR

W-F

W-SO4-IC

W-TDS

W-TSS

Description

Alkalinity in aqueous matrices as mg CaCO3/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

True Color measured at 450 nm (not valid for NPDES monitoring)

Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)

Sulfate in aqueous matrices

Total Dissolved Solids in aqueous matrices

Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Container ID: BP-1L

Qty: 4 Preservation: ICE

Lot # 1283330

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-250ML-WO-THI

Qty: 4 Preservation: ICE

Lot # 119303

Description: 250 ml Sterilized no tablet

Analysis

ECOLI-18QT

Description

Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Container ID: P-500ML

Qty: 4 Preservation: ICE And H2SO4

Lot # 00078041

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 4 Preservation: FILTER-ICE

Lot # 00078144

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Cooler Packed By: HK

Number of Coolers: 2

Date: 8-28-20

Kit must also include:

- ☒ Field Sheets
- ☒ Temperature Control Bottle
- ☐ FedEx Bills, if applicable
- ☐ Plastic Bags/Zip Ties
- ☐ Acid Labels (sets of 24)

If Preservation Included:

ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____

PLEASE RETURN ALL COOLERS FOR RQ-2020-09-07-22