

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

YSI 6021

RQ:

2021-06-21-26

Project:

SMP/Vogel/Von

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/18/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J Purnish	6/23/21	730	22.8	762	8.6	9.6	100.3	—	1.066	P/F	L/F
ICV	J	6/23/21	731	22.8	762	8.6	8.64	100.3	—	1.066	P/F	L/F
CCV	J Purnish	6/23/21	1300	23.8	763	8.5	8.53	101.0	—	1.066	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J Purnish	6/23/21	732	210426B	5/22	1000	1000	P/F	L/F
ICV	J	6/23/21	733	210426A	5/22	100	101.4	P/F	L/F
CCV	J Purnish	6/23/21	1302	210426D	5/22	20000	19908	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J Purnish	6/23/21	734	2112E39	12/22	7.	22.8	7.01	-13.7	P/F	L/F
Calibr.	J	6/23/21	735	2103659	3/23	4.	22.9	4.00	165.3	P/F	L/F
Calibr.	J	6/23/21	736	2103F53	9/22	10.	23.1	10.02	-184.6	P/F	L/F
ICV	J	6/23/21	737	2103659	3/23	4.0	22.9	4.10	158.6	P/F	L/F
CCV	J Purnish	6/23/21	1303	2103F53	9/22	10.0	23.7	10.06	-182.6	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 170.5, slope from 4 to 7 178.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/18/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 5; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J Purnish	6/23/21	745	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/23/2021
Customer: NE-DIST
RQ: RQ-2021-06-21-26

Collected By: Jeremy Parrish, Jessica Lee
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
PIGEON CREEK @ ANDREWS RD	 19010098
PIGEON CREEK AT US 1	 19010099
Creek draining former Hampton lake 100m above rd	 G4NE0292
Little St. Marys at White Oak Rd.	 G4NE0289

WIN ID/Field ID:



MILLS CR @ CR 115

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/23/2021
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: PIGEON CREEK @ ANDREWS RD **Date:** 6/23/2021
WIN/Field ID: 19010098 **WBID:** 2105 **ENR:** - **Latitude:** 30.7286
County: NASSAU **Org ID:** 21FLA **Longitude:** -81.97
Receiving Body of Water: Nassau - St. Marys **RQ:** 2021-06-21-26

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish	x	x	x		
☒	Jessica Lee		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# DESI							
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:							
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:													
SBIO ID Numbers:													
Notes: RECENT RAIN													

Duplicate Sample

Time Zone: EST	Time: _____	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID: _____			

Blank

Time Zone: _____	Time: _____	Field ID: _____	(Blank Type_DDDMMYY)
DI Source: _____	Location: _____		
DI Type: _____	Equipment ID: _____		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: _____ WIN Import ID: _____
Enter Field Parameters, Verify Station ID In LIMS DMT: _____ QC Station ID, Field Parameters In LIMS DMT: BD 07/07/2021
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets *GL* 06/23/2021 Migrate Data to WIN: _____ Verify Data In WIN: _____
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-06-21-26

Customer: WQAP

Project ID: SMP

Collected By: Jeremy Parrish, Jessica Lee

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



19010098

Location: PIGEON CREEK @ ANDREWS RD

Comments: Microbiology sample(s) submitted to contract laboratory

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
6/23/2021	1025 EST	4.3	50.8	23.8	4.45	0.3	0.4 ☐ VOB (S)	0.7	78.3	0.04
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4*	☒ <2	1	A
	Lot # Sa0293040 Exp: 10/19/21				
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD	☒ Ice ☒ HNO3*	☒ <2	1	A
	Lot # Na0344060 Exp: 1/4/22				
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice		1	A
	Filter Lot # ronb29490				
	Syringe Lot #				
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	A
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R	☐ Ice			
	☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)				
Phytoplankton/Periphyton	☐ DTY-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Jeremy Parrish
Jessica Lee

Signature:

06/23/2021

Date:

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

ALS Contact: Mandy Sullivan

Page 1 of 1

[illegible]