

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Desi

RQ- 2021-04-26-12

Project: Nassau West 1

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/15/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Brian Davis	4/28/21	740	24.3	764.1	8.4	8.4	100.5	✓	1.06	(P) F	(L) F
ICV	↓	↓	741	24.3	764.1	8.4	8.41	100.5	✓	1.06	(P) F	(L) F
CCV	Kim Appleby	4/29/21	1527	22.9	763.1	8.6	8.60	100.0	—	1.06	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Brian Davis	4/28/21	743	201222 B	1/22	1000	1000	(P) F	(L) F
ICV	↓	↓	745	201222 A	1/22	100	103.6	(P) F	(L) F
CCV	Kim Appleby	↓	1529	200902C	9/21	1413	1403	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Brian Davis	4/28/21	752	2010F27	10/22	7.00	25.5	7.00	-35.0	(P) F	(L) F
Calibr.			755	2010C31	9/22	4.00	25.5	4.01	146.2	(P) F	(L) F
Calibr.			758	202A7	5/22	10.00	25.5	10.00	-208.1	(P) F	(L) F
ICV			801	2010F27	10/22	7.00	25.4	7.04	-34.5	(P) F	(L) F
CCV	Kim Appleby		1533	2010C31	8/22	4.00	24.5	4.08	140.2	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Brian Davis	4/28/21	802	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 4/28/2021
Customer: NE-DIST
RQ: RQ-2021-04-26-12

Collected By: Brian Davis, Kimberly Appleby
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Place Label Here WIN ID/Field ID:  19010127 Location: UNNAMED STREAM @ CR 121	
Place Label Here WIN ID/Field ID:  19010048 Location: DEEP CREEK AT CR 121	
Place Label Here WIN ID/Field ID:  19010093 Location: ST MARYS R @ STOKES RD	
Place Label Here WIN ID/Field ID:  G4NE0302 Location: Unnamed Drain to Brandy Branch at US301	

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 4/28/2021
1600

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: UNNAMED STREAM @ CR 121 **Date:** 4/28/2021
WIN/Field ID: 19010127 **WBID:** 2178 **ENR:** - **Latitude:** 30.5167
County: NASSAU **Org ID:** 21FLA **Longitude:** -81.99
Receiving Body of Water: Nassau - St. Marys **RQ:** 2021-04-26-12

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Brian Davis	x		x		
☒	Kimberly Appleby		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: YES	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A (WIN ID_DUP)
WIN DMT Activity ID:		

Blank

Time Zone:	Time:	Field ID: (Blank Type_DDDMMYY)
DI Source:	Location:	
DI Type:	Equipment ID:	
DI Container (Name/ID):		DI Container Type:

LIMS EVENT ID: WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets BD 04/28/2021 Migrate Data to WIN: Verify Data In WIN:
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-04-26-12

Customer: WQAP

Project ID: SMP

Collected By: Brian Davis, Kimberly Appleby

Field Report Prepared By: Brian Davis

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



19010127

Location: UNNAMED STREAM @ CR 121

Comments: Microbiology sample(s) submitted to contract laboratory

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
4/28/2021	1000 EST	2.36	25.8	19.6	4.12	0.1	0.2 ☒ VOB (S)	0.2	46.1	0.02
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA0293040 Exp: 10/19/21	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD Lot # NA0344060 Exp: 1/4/22	☒ Ice ☒ HNO3*	☒ <2	1	A
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # R0NB29490 Syringe Lot #	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	A
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Brian Davis
Kimberly Appleby


Signature:

Date:

04/28/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



Page 1 of 1

[illegible]