

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Stitch RQ: 2021-05-10-23 Project: SMP Nassau NW

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 5/17/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Brian D.</u>	<u>6/9/21</u>	<u>812</u>	<u>22.6</u>	<u>765.7</u>	<u>8.7</u>	<u>8.7</u>	<u>100.8</u>	<u>✓</u>	<u>1.02</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>↓</u>	<u>↓</u>	<u>814</u>	<u>22.6</u>	<u>765.7</u>	<u>8.7</u>	<u>8.71</u>	<u>100.8</u>	<u>✓</u>	<u>1.02</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>↓</u>	<u>↓</u>	<u>1501</u>	<u>22.7</u>	<u>763.8</u>	<u>8.7</u>	<u>8.73</u>	<u>101.8</u>	<u>✓</u>	<u>1.02</u>	<u>P/F</u>	<u>L/F</u>
CCV											<u>P/F</u>	<u>L/F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Brian Davis</u>	<u>6/9/21</u>	<u>804</u>	<u>2010F27</u>	<u>10/22</u>	<u>1000</u>	<u>1000</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>↓</u>	<u>↓</u>	<u>808</u>	<u>201222A</u>	<u>1/22</u>	<u>100</u>	<u>103.7</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>↓</u>	<u>↓</u>	<u>1458</u>	<u>2010F27</u>	<u>10/22</u>	<u>1000</u>	<u>1001</u>	<u>P/F</u>	<u>L/F</u>
CCV								<u>P/F</u>	<u>L/F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Brian Davis</u>	<u>6/9/21</u>	<u>819</u>	<u>2012E39</u>	<u>12/22</u>	<u>7.01</u>	<u>22.9</u>	<u>7.01</u>	<u>-22.7</u>	<u>P/F</u>	<u>L/F</u>
Calibr.	<u>↓</u>	<u>↓</u>	<u>824</u>	<u>2103G59</u>	<u>3/23</u>	<u>4.00</u>	<u>23.0</u>	<u>4.00</u>	<u>153.3</u>	<u>P/F</u>	<u>L/F</u>
Calibr.	<u>↓</u>	<u>↓</u>	<u>829</u>	<u>2103F53</u>	<u>9/22</u>	<u>10.02</u>	<u>22.7</u>	<u>10.03</u>	<u>-195.5</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>↓</u>	<u>↓</u>	<u>832</u>	<u>2012E39</u>	<u>12/22</u>	<u>7.01</u>	<u>22.4</u>	<u>7.01</u>	<u>-21.4</u>	<u>P/F</u>	<u>L/F</u>
CCV	<u>↓</u>	<u>↓</u>	<u>1505</u>	<u>2012E39</u>	<u>12/22</u>	<u>4.00</u>	<u>22.4</u>	<u>3.99</u>	<u>153.4</u>	<u>P/F</u>	<u>L/F</u>
CCV				<u>2103G59</u>	<u>3/23</u>					<u>P/F</u>	<u>L/F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 172.8, slope from 4 to 7 176 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? **YES** / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/17/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air	<u>B - D -</u>	<u>6/9/21</u>	<u>834</u>	<u>0.00</u>	<u>0.00</u>	<u>P/F</u>	<u>L/F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/9/2021
Customer: NE-DIST
RQ: RQ-2021-05-10-23

Collected By: Jessica Lee, Brian Davis
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.


Site Location	Field ID/WIN ID
MOCCASIN CREEK @ MOCCASIN CREEK CIR. E.	 19010121
DAUGHTERY BR @ CREWS RD	 19010136

WIN /Field ID:



CALKINS AT TURNER CEMETARY RD

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/9/2021

1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: CALKINS AT TURNER CEMETARY RD **Date:** 6/9/2021
WIN/Field ID: 19010076 **WBID:** 2264 **ENR:** - **Latitude:** 30.328
County: BAKER **Org ID:** 21FLA **Longitude:** -82.23
Receiving Body of Water: Nassau - St. Marys **RQ:** 2021-05-10-23

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jessica Lee		x		x	x
☒	Brian Davis	x		x		
☐						
☐						

Sampling Equipment	
☐ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# STITCH							
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:							
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:													
SBIO ID Numbers:													
Notes:													

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: **WIN Import ID:**
Enter Field Parameters, Verify Station ID In LIMS DMT: **QC Station ID, Field Parameters In LIMS DMT:** BD 07/07/2021
(Initials/Date) (Initials/Date)

Scan/Save Field Sheets **6/9/2021** **Migrate Data to WIN:** **Verify Data In WIN:**
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-05-10-23

Customer: WQAP

Project ID: SMP

Collected By: Jessica Lee, Brian Davis

Field Report Prepared By: Jessica Lee

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



19010076

Location: CALKINS AT TURNER CEMETARY RD

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
6/9/2021	1125 EST	3.93	47	24	7.01	0.25	0.5 ☒ VOB (S)	0.5	121.6	0.06
							Central Lab please use top row for login purposes			
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H2SO4*	☐ <2		
	Lot # Exp:				
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD	☐ Ice ☐ HNO3*	☐ <2		
	Lot # Exp:				
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
	Filter Lot #				
	Syringe Lot #				
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☒ Ice		1	C
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R	☐ Ice			
	☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN				
Phytoplankton/Periphyton	☐ DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			
	☐ DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV ☐				

Name:
Jessica Lee
Brian Davis

Signature:


Jessica Lee

Date:

6/9/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



SR # _____

Page 1 of 1

[illegible]