

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Stitch RQ: 2021-05-24-11 Project: Nassau Lakes

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/17/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Brian Davis	6/14/21	840	20.2	759.2	9.0	9.0	99.9	/	1.03	P/F	L/F
ICV	↓	↓	842	20.3	759.2	9.0	9.04	100.0	/	1.03	P/F	L/F
CCV	Kim Appleby	↓	1649	21.2	758.6	8.9	8.88	99.9	/	1.03	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Brian Davis	6/14/21	833	201222B	1/22	1000	1000	P/F	L/F
ICV	↓	↓	835	201222A	1/22	100	102.0	P/F	L/F
CCV	Kim Appleby	6/14/21	1652	200902L	9/21	141.13	141.0	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot # BP 2012E29	Expir. Date 12/22	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Brian Davis	6/14/21	846	201222B	1/22	7.02	19.9	7.02	-22.4	P/F	L/F
Calibr.	↓	↓	849	2103655	3/23	4.00	19.9	4.00	152.1	P/F	L/F
Calibr.	↓	↓	852	2103F53	9/22	10.05	19.8	10.06	-195.3	P/F	L/F
ICV	↓	↓	856	2012E29	12/22	7.02	19.8	7.04	-22.4	P/F	L/F
CCV	Kim Appleby	↓	1655	7103659	3/23	4.00	20.2	4.01	151.4	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.9, slope from 4 to 7 -174.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/17/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Brian Davis	6/14/21	858	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/14/2021
Customer: NE-DIST
RQ: RQ-2021-05-24-11

Collected By: Kimberly Appleby, Brian Davis
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB


Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
OCEAN POND CENTER	 21010011
PALESTINE LAKE CENTER	 19010044

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/14/2021
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: OCEAN POND CENTER

Date: 6/14/2021

WIN/Field ID: 21010011

WBID: 2339

ENR: -

Latitude: 30.2276

County: BAKER

Org ID: 21FLA

Longitude: -82.44

Receiving Body of Water: Nassau - St. Marys

RQ: 2021-05-24-11

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒ Kimberly Appleby		☒		☒	☒
☒ Brian Davis		☒	☒		
☐					
☐					

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☒ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# STITCH	
Photos: NO	Flow: NA	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: **WIN Import ID:**

Enter Field Parameters, Verify Station ID in LIMS DMT: **QC Station ID, Field Parameters in LIMS DMT:** BD 07/07/2021

(Initials/Date) (Initials/Date)

Scan/Save Field Sheets BD 06/15/2021 **Migrate Data to WIN:** **Verify Data in WIN:**

(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-05-24-11

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Brian Davis

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



21010011

Location: OCEAN POND CENTER

Comments:

R Qualified

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
6/14/2021	1325 EST	7.33	96.5	29.8	5.28	0.3	0.9 ☐ VOB (S)	2.9	41.1	0.02
	1327	6.88	88.5	28.3	5.2	2.0	Central Lab please use top row for login purposes		41.1	0.02
	1330 EST	6.79	88.5	28.3	5.2	2.4			41.1	0.02

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA0353110 Exp: 1641920400000	☒ Ice ☒ H2SO4*	☒ <2	1	B
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD Lot # NA0344060 Exp: 1/4/22	☒ Ice ☒ HNO3*	☒ <2	1	B
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # RONB38756 Syringe Lot #	☒ Ice		1	B
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	B
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	B
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☐ Ice			
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Kimberly Appleby
Brian Davis

Signature:

Date:

06/15/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution