

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSI 600

RQ: 2021-04-26-13

Project: Palm Valley

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/17/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Parnish	6/7/21	925	16.3	764.6	9.87	9.8	100.6	—	1.034	(P) / (F)	(L) / (F)
ICV	J	6/7/21	926	16.5	764.6	9.8	9.8	100.7	—	1.034	(P) / (F)	(L) / (F)
CCV	J. Lee	1	1336	23.4	764.0	8.4	8.46	99.4	—	1.034	(P) / (F)	(L) / (F)
CCV				23.4							P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Parnish	6/7/21	729	201222B	1/22	1000	1000	(P) / (F)	(L) / (F)
ICV	J		730	2009020	9/21	1413	1409	(P) / (F)	(L) / (F)
CCV	J. Lee		1346	210426E	5/22	50000	5000	(P) / (F)	(L) / (F)
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Parnish	6/7/21	731	2012E39	12/22	7.	16.8	7.00	30.0	(P) / (F)	(L) / (F)
Calibr.	J		733	2103659	3/23	4.	16.9	4.00	148.2	(P) / (F)	(L) / (F)
Calibr.	J		735	2103F53	9/22	10.	16.6	10.00	143.0	(P) / (F)	(L) / (F)
ICV	J		738	2103659	3/23	4.	16.9	4.00	145.3	(P) / (F)	(L) / (F)
CCV	J. Lee		1350	2103F53	9/22	10.06	10.5	9.99	-15.2	(P) / (F)	(L) / (F)
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 223.0, slope from 4 to 7 178.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/17/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Well 3; Date of last verification: 5/18/21

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J. Parnish	6/7/21	747	0.00	0.00	(P) / (F)	(L) / (F)

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/7/2021
Customer: NE-DIST
RQ: RQ-2021-04-26-13

Collected By: Jeremy Parrish, Jessica Lee
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Marshall Creek 0.5 miles above mouth	 G5NE0595
CAPO CREEK 300M ABOVE MOUTH	 27010019
CAPO CREEK 100m above mouth	 G5NE0596
DEEP CREEK @ END OF PINE IS RD	 27010072
DEEP CREEK 1ST NE BEND FROM MOUTH	 27010069

WIN ID/Field ID:



27010165

DEEP CREEK 270M ABOVE ICW

RELINQUISHED BY(SIGNATURE):

DATE/TIME: 6/7/2021 1600
EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

**Rec'd/Inspected
Date:**



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: CAPO CREEK 300M ABOVE MOUTH

Date: 6/7/2021

WIN/Field ID: 27010019

WBID: 2435

ENR: ENRT1

Latitude: 30.0336

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.35

Receiving Body of Water: Upper East Coast

RQ- 2021-04-26-13

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish	x		x		
☒	Jessica Lee		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Sonar	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☒ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# BRUCE	
Photos: NO		Flow: FLOW		Tide: FALLING		Tide Times: High: 930 Low: 1530	
Biological Samples Collected:		☒ None	☐ HA	☐ SCI	☐ RPS	☐ Algae ID	☐ LVS ☐ LVI Other:
SBIO ID Numbers:							
Notes:							

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMYYYY)
DI Source:		Location:	
DI Type:		Equipment ID:	
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID:	WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT:	QC Station ID, Field Parameters In LIMS DMT: BD 07/07/2021
(Initials/Date)	(Initials/Date)
Scan/Save Field Sheets	Migrate Data to WIN: JP 7/9/2021
(Initials/Date)	(Initials/Date)
	Verify Data In WIN:
	(Initials/Date)



RQ-2021-04-26-13

Customer: WQAP

Project ID: SMP

Collected By: Jeremy Parrish, Jessica Lee

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



27010019

Location: CAPO CREEK 300M ABOVE MOUTH

Comments: Microbiology samples submitted to contract laboratory

Matrix: Marine

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
6/7/2021	1045 EST	5.28	82.1	28.2	7.47	0.3	0.4 ☐ VOB (S)	1.4	51933	34.08
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # Sa0293040 Exp: 10/19/21	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # Ronb29490 Syringe Lot #	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☒ Ice		1	A
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☒ F Coli ☒ Enterococci ☒ Contract Lab	☒ Ice		2	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C PKD-QN-BV	☐ Ice			

Name:
Jeremy Parrish
Jessica Lee

Signature:

Jessica Lee

Date:

06/07/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

SR #

A1 S Contact- Mandy Sullivan

9143 Philips Highway, Suite 200 Jacksonville, FL 32256 / Ph(904) 739-2277 / FAX (904) 739-2011

Page 1 of 1[illegible]