

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

51156

RQ:

2021-02-22-07

Project:

Palm Valley

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/15/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Brian Davis	2/25/21	725	20.9	765.9	9.0	✓	100.8	✓	1.03	P/F	L/F
ICV			726	21.0	766.0	9.0	8.99	100.8	✓	1.03	P/F	L/F
CCV			1457	22.9	764.4	8.6	8.68	101.1	✓	1.03	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Brian Davis	2/25/21	729	201222B	1/22	1000	1000	P/F	L/F
ICV			731	201222C	1/22	50,000	49,869	P/F	L/F
CCV			1458			1413	1414	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Brian Davis	2/25/21	734	201222B	1/22	7.01	22.5	7.01	-15.5	P/F	L/F
Calibr.			736	2912B71	12/21	4.00	22.4	4.00	163.0	P/F	L/F
Calibr.			739	2006E28	12/21	10.03	22.4	10.03	-187.8	P/F	L/F
ICV			745	2006F52	6/22	7.01	22.3	7.04	-15.2	P/F	L/F
CCV			1504	2006E28	12/21	7.01	25.0	10.10	-195.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Brian Davis	2/25/21	748	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: <u>2/25/2021</u> Customer: <u>NE-DIST</u> RQ: <u>RQ-2021-02-22-17</u>	Collected By: <u>Kelly McDonald, Brian Davis</u> Sampling Agency: <u>FDEP</u> Project ID: <u>SMP</u>
--	---

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Marshall Creek 0.5 miles above mouth	 G5NE0595
CAPO CREEK 300M ABOVE MOUTH	 27010019
CAPO CREEK 100m above mouth	 G5NE0596
DEEP CREEK @ END OF PINE IS RD	 27010072
DEEP CREEK 1ST NE BEND FROM MOUTH	 27010069
DEEP CREEK 270M ABOVE ICW	 27010165

RELINQUISHED BY(SIGNATURE): *Kelly A McDonald*

DATE/TIME: 2/25/2021 1600

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
☒ Yes / ☐ No

WIN Station Name: Marshall Creek 0.5 miles above mouth **Date:** 2/25/2021

WIN/Field ID: G5NE0595 **WBID:** 2442 **ENR:** - **Latitude:** 30.0254

County: ST. JOHNS **Org ID:** 21FLA **Longitude:** -81.3709

Receiving Body of Water: Upper East Coast **RQ:** 2021-02-22-17

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kelly McDonald	x		x		
☒	Brian Davis		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other _____	☒ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# STITCH					
Photos: NO		Flow: FLOW		Tide: FALLING		Tide Times: High: 0915 Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time: _____	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID: _____			

Blank

Time Zone: _____	Time: _____	Field ID: _____	(Blank Type_DDDMMYY)
DI Source: _____	Location: _____		
DI Type: _____	Equipment ID: _____		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: _____ WIN Import ID: _____

Enter Field Parameters, Verify Station ID In LIMS DMT: _____ QC Station ID, Field Parameters In LIMS DMT: _____

(Initials/Date) (Initials/Date)

Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data In WIN: _____

(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-02-22-17

Customer: WQAP

Project ID: SMP

Collected By: Kelly McDonald, Brian Davis

Field Report Prepared By: Kelly McDonald

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G5NE0595

Location: Marshall Creek 0.5 miles above mouth

Comments:

Matrix: Marine

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
2/25/2021	940 EST	6.25	68.7	16.8	7.3	0.3	0.45 ☐ VOB (S)	1.2	17666	10.61
							Central Lab please use top row for login purposes			
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4*	☒ <2	1	A
	Lot # Sa0293040 Exp: 10.19.21				
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD	☐ Ice ☐ HNO3*	☐ <2		
	Lot # Exp:				
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice		1	A
	Filter Lot # R0cb4867				
	Syringe Lot #				
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☒ Ice		1	A
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☒ F Coli ☒ Enterococci ☐ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R	☐ Ice			
	☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN				
Phytoplankton/Periphyton	☐ DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			
	☐ DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV				

Name:
Kelly McDonald
Brian Davis

Signature:

Kelly A McDonald

Date:

2/25/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

[illegible]