

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Vesi

RQ- 2021-04-26-15

Project: Unnamed Run

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/15/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blair Davis	4/27/21	840	23.6	765.2	8.5	✓	100.7	✓	1.06	P/F	L/F
ICV		5/1	841	23.6	765.2	8.5	8.55	100.7	✓	1.06	P/F	L/F
CCV	Kelly McDonald	↓	1431	23.0	764.6	8.6	8.51	99.9	✓	1.06	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Blair Davis	4/27/21	843	201222B	1/22	1000	1000	P/F	L/F
ICV		↓	845	201222C	1/22	50,000	49,571	P/F	L/F
CCV	Kelly McDonald	↓	1433	200407B	04/2021	20,000	19,350	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blair Davis	4/27/21	847	2010F52	10/22	7.00	24.7	7.00	-36.1	P/F	L/F
Calibr.		↓	849	2010C31	9/22	4.00	24.8	4.01	143.7	P/F	L/F
Calibr.		↓	851	2012797	5/22	10.00	24.8	10.00	-209.8	P/F	L/F
ICV		↓	855	2010F52	10/22	7.00	24.8	7.04	-36.1	P/F	L/F
CCV	Kelly McDonald	↓	1438	2012797	5/22	10.00	25.3	10.12	-217.8	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/15/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Blair Davis	4/27/21	854	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 4/27/2021
Customer: NE-DIST
RQ: RQ-2021-04-26-15

Collected By: Kelly McDonald, Brian Davis
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Unnamed Drain South Entrance	 G5NE0591
Unnamed Drain South Entrance Duplicate	 Field ID: G5NE0591 DUP
Field Blank	 Field ID: <u>FB 04262021</u>

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 4/27/2021

1115

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Unnamed Drain South Entrance **Date:** 4/27/2021
WIN/Field ID: G5NE0591 **WBID:** 2573 **ENR:** - **Latitude:** 29.6958
County: ST. JOHNS **Org ID:** 21FLA **Longitude:** -81.2336
Receiving Body of Water: Upper East Coast **RQ:** 2021-04-26-15

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kelly McDonald	x		x		
☒	Brian Davis		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other _____	☒ Gasoline Motor

Water Level: ☐ Normal ☐ Low ☒ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# DESI					
Photos: NO		Flow: FLOW		Tide: FALLING		Tide Times: High: 0900 Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time: 1120	Field ID: G5NE0591 DUP	(WIN ID_DUP)
WIN DMT Activity ID: _____			

Blank

Time Zone: EST	Time: 1125	Field ID: FB_04262021	(Blank Type_DDDMMYYYY)
DI Source: Neroc		Location: Unnamed drain south entrance	
DI Type: FB		Equipment ID: Ortho 3	
DI Container (Name/ID): Carboy		DI Container Type: Carboy	

LIMS EVENT ID: _____ WIN Import ID: _____
Enter Field Parameters, Verify Station ID In LIMS DMT: _____ QC Station ID, Field Parameters In LIMS DMT: _____
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets BD 04/27/21 Migrate Data to WIN: _____ Verify Data In WIN: _____
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-04-26-15

Customer: WQAP

Project ID: SMP

Collected By: Kelly McDonald, Brian Davis

Field Report Prepared By: Kelly McDonald

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G5NE0591

Location: Unnamed Drain South Entrance

Comments:

Matrix: Marine

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
4/27/2021	1115 EST	7.41	105	23	7.9	0.3	1 ☐ VOB (S)	1.45	50231	32.91
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # Sa0293040 Exp: 10.19.21	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # R0nb29490 Syringe Lot #	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☒ Ice		1	A
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☒ F Coli ☒ Enterococci ☐ Contract Lab	☒ Ice		2	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-OL-C PEN-OL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C PKD-QN-BV	☐ Ice			

Name:
Kelly McDonald
Brian Davis

Signature:

Date:

04/27/21

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



RQ- 2021-04-26-15

Customer: WQAP

Project ID: SMP

Collected By: Kelly McDonald, Brian Davis

Field Report Prepared By: Kelly McDonald, Brian Davis

Sampling Agency: FDEP

DUPLICATE SAMPLE



Location: Unnamed Drain South Entrance

Field ID: G5NE0591 DUP

Comments:

Matrix: Marine

(Check one)

☒ GRAB☐ COMPOSITE

Date

Time

4/27/2021

1120

EST

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # Exp:	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # R0nb29490 Syringe Lot #	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☒ Ice		1	A
Microbiology (Contract Lab Bottle)	☐ E Coli ☒ F Coli ☒ Enterococci ☐ Contract Lab	☒ Ice		2	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD See Notes	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDIN	☐ Ice			
Name: Kelly McDonald Brian Davis		Signature: <i>[Signature]</i>		Date: 04/27/21	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



RQ- 2021-04-26-15

Collected By: Kelly McDonald, Brian Davis

Customer: WQAP

Field Report Prepared By: Kelly McDonald, Brian Davis

Project ID: SMP

Sampling Agency: FDEP

BLANK



Location: Unnamed Drain South Entrance

Field ID: FB_04262021

Comments:

Matrix: FB

(Check one) ☐ COMPOSITE ☒ GRAB

Date

Time

DI Source: Neroc

4/27/2021

1125

EST

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	☐ <2	1	B
	Lot # _____ Exp: _____	10.19.21			
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ H2SO4	☐ <2		
	Lot # _____ Exp: _____				
Filtered Nutrients (P125ML)	☒ W-PO4-F Filter ☒ Field Filtered 0.45 µm	☒ Ice		1	B
	Filter Lot # R0nb29490				
	Syringe Lot # _____				
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-IC / TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☒ Ice		1	B
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R	☐ Ice			
	☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN				

Name:
Kelly McDonald
Brian Davis

Signature:

Date:

04/27/21

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

SR #

9143 Phillips Highway, Suite 200 Jacksonville FL 32256 / Ph(904) 739-2277 / FAX(904) 739-2011

Page 1 of 1

[illegible]