

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Desi

RQ- 2021-06-21-15

Project: SMP-Pellicer Boat

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/18/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Lee	6/22/21	0737	21.5	760.4	8.8	8.81	100.1	-	1.06	P/F	L/F
ICV			0738	21.6	760.4	8.8	8.81	100.0	-	1.06	P/F	L/F
CCV			1253	23.7	760.6	8.5	8.48	100.1	-	1.06	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J. Lee	6/22/21	0744	20426B	5/22	1000	1000	P/F	L/F
ICV			0749	210426A	5/22	100	101.0	P/F	L/F
CCV			1259	210426D	5/22	20000	20057	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Lee	6/22/21	0751	2012E39	12/22	7.02	20.7	7.02	-14.1	P/F	L/F
Calibr.			0753	2103G59	3/22	4.00	20.7	4.00	165.1	P/F	L/F
Calibr.			0754	2103F53	9/22	10.04	21.1	10.04	-183.6	P/F	L/F
ICV			0756	2103G59	3/22	4.00	20.8	4.00	163.2	P/F	L/F
CCV			1305	2103F53	9/22	10.03	21.8	10.09	-188.8	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.5, slope from 4 to 7 179.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/27/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J. Lee	6/22/21	0757	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/22/2021
Customer: NE-DIST
RQ: RQ-2021-06-21-15

Collected By: 6/22/2021 1600
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
PELLICER CR AT US 1	 27010016
Pellicer 1.3 mi downstream of I95	 G5NE0593
PELLICER CR. AT FAVER DYKES BT.RMP.	 27010965

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/22/2021
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: PELLICER CR AT US 1

Date: 6/22/2021

WIN/Field ID: 27010016

WBID: 2580B

ENR: ENRT4

Latitude: 29.6513

County: FLAGLER

Org ID: 21FLA

Longitude: -81.29

Receiving Body of Water: Upper East Coast

RQ: 2021-06-21-15

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jessica Lee		x		x	x
☒	Brian Davis	x		x		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☒ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# DESI	
Photos: NA		Flow: FLOW		Tide: RISING		Tide Times: High: Low:	
Biological Samples Collected:		☒ None	☐ HA	☐ SCI	☐ RPS	☐ Algae ID	☐ LVS ☐ LVI Other:
SBIO ID Numbers:							
Notes:							

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYY)
DI Source:		Location:	
DI Type:		Equipment ID:	
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT: QC Station ID, Field Parameters in LIMS DMT: BD 07/07/2021

Scan/Save Field Sheets 06/22/2021 Migrate Data to WIN: Verify Data in WIN:



RQ-2021-06-21-15

Customer: WQAP

Project ID: SMP

Collected By: Jessica Lee, Brian Davis

Field Report Prepared By: Jessica Lee

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



27010016

Location: PELLICER CR AT US 1

Comments: Microbiology sample(s) submitted to contract laboratory

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
6/22/2021	935 EST	4.78	57.8	24.8	7.01	0.3	1.1 ☐ VOB (S)	2	310	0.15
	937 EST	4.66	56.2	24.7	6.99	1.5	Central Lab please use top row for login purposes		308.6	0.15

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # Exp:	☐ Ice ☐ H2SO4*	☐ <2		
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter Filter Lot # Syringe Lot #	☐ Ice			
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☒ F Coli ☒ Enterococci ☒ Contract Lab	☒ Ice		2	C
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	☐ DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C ☐ DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Jessica Lee
Brian Davis

Signature:

Jessica Lee

Date:

06/22/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

