

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Pro DSS Bruce

RQ: 2021-06-07-21

Project: SMP Palm Coast

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/17/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	6/16/21	749	22.4	759.7	8.7	8.7	99.9	-	1.035	(P) F	(L) F
ICV			751	22.4	759.7	8.7	8.69	100.0	-	1.035	(P) F	(L) F
CCV	K. McDonald		1409	24.9	759.16	8.3	8.24	99.7	-	1.035	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	6/16/21	752	201222B	1/2022	1000	1000	(P) F	(L) F
ICV			756	201222A	1/2022	100	104.103.9	(P) F	(L) F
CCV	K. McDonald		1410	210426D	05/22	20,000	20015	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	6/16/21	806	2012E39	12/22	7.00	22.0	7.00	-34.30	(P) F	(L) F
Calibr.			807	2103G59	3/23	4.00	21.9	4.00	137.9	(P) F	(L) F
Calibr.			809	2103F53	9/22	10.00	22.2	10.00	-201.7	(P) F	(L) F
ICV			811	2103G59	3/23	4.00	22.0	3.97	170.5	(P) F	(L) F
CCV	K. McDonald		1412	2103F53	9/22	10.02	22.9	10.00	-206.8	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/17/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	6/16/21	811	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/16/2021
Customer: NE-DIST
RQ: RQ-2021-06-07-29

Collected By: Kimberly Appleby, Kelly McDonald
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
BULL CREEK DITCH downstream of East ditch	 G2NE1212
BULL CREEK DITCHES at farm entrance	 G2NE1214
Bull Creek East ditch at Water Oak	 G2NE1213

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/16/2021
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
☒ Yes / ☐ No

WIN Station Name: Bull Creek East ditch at Water Oak **Date:** 6/16/2021
WIN/Field ID: G2NE1213 **WBID:** 2615 **ENR:** - **Latitude:** 29.4449
County: FLAGLER **Org ID:** 21FLA **Longitude:** -81.43
Receiving Body of Water: Lower St. Johns **RQ:** 2021-06-07-29

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	x		x		
☒	Kelly McDonald		x		x	X
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☒ Van Dorn He-man
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☒ Other Bridge	☐ Gasoline Motor

Water Level: ☐ Normal ☐ Low ☒ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# BRUCE							
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:							
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:													
SBIO ID Numbers:													
Notes: CODE R FOR RECENT RAIN. 2.2 INCHES IN PAST 48 HRS. MORE TURBID THEN USUAL. WEATHER: OVERCAST, 80S. A COUPLE OF DEAD FISH IN WATER AT SAMPLE LOCATION.													

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYY)
DI Source:		Location:	
DI Type:		Equipment ID:	
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID: BD 07/07/2021
Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT: (Initials/Date) (Initials/Date)
Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN: (Initials/Date) (Initials/Date)



RQ-2021-06-07-29

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Kelly McDonald

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G2NE1213

Location: Bull Creek East ditch at Water Oak

Comments: Microbiology sample(s) submitted to contract laboratory

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
6/16/2021	1100 EST	4.69	56.6	24.7	7.08	0.1	0.21 ☒ VOB (S)	0.21	1536	0.77
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # _____ Exp: _____	☐ Ice ☐ H2SO4*	☐ <2		
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # _____ Exp: _____	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter Filter Lot # _____ Syringe Lot # _____	☐ Ice			
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Kimberly Appleby
Kelly McDonald

Signature:

Kelly A. McDonald

Date:

6/16/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

SR # _____

ALS Contact: Mandy Sullivan

9143 Philips Highway, Suite 200 Jacksonville, FL 32256 / PH(904) 739-2277 / FAX (904) 739-2011

Page 1 of 1[illegible]