

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSI 5726

RQ: 2021-05-31-23

Project: SMP Gainesville

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/17/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Barometer mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Panish	6/16/21	722	21.7	760	8.8	8.8	100.0	—	1.030	P/F	L/F
ICV	J	6/16/21	723	21.8	760	8.8	8.79	100.1	—	1.030	P/F	L/F
CCV	J. Lee	1	1514	23.5	7589	8.5	8.55	100.2	—	1.030	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Panish	6/16/21	724	20222B	1/22	1000	1000	P/F	L/F
ICV	J	2	727	201222A	1/22	100	102.0	P/F	L/F
CCV	J. Lee	1	1519	200902C	9/21	1413	1433	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Panish	6/16/21	724	2012E39	12/22	7.	21.7	7.02	-31.5	P/F	L/F
Calibr.	J		726	203659	3/23	4.	21.7	4.00	149.4	P/F	L/F
Calibr.	J		730	2103F53	9/22	10.	21.9	10.00	-146.5	P/F	L/F
ICV	J		731	2103659	3/23	4.	21.7	4.09	142.3	P/F	L/F
CCV	J. Lee	1	1521	2012103F53	9/22	10.02	23.1	9.97	-148.0	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 180.9 slope from 4 to 7 180.9 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)
 If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/17/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J. Pan	6/16/21	735	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/16/2021
Customer: NE-DIST
RQ: RQ-2021-05-31-23

Collected By: Jeremy Parrish, Jessica Lee
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Lochloosa creek at Fish Camp Rd.	 G3NE0015
Alligator Creek at 301 bypass West of Starke	 G1NE0935
Calf Creek at SE 27th st.	 G1NE0931
Calf Creek 0.5mi upstream of SE27th	 G1NE0934
ALLIGATOR CREEK end of Short rd.	 G1NE0941

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/16/2021
1608 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

**Rec'd/Inspected
Date:**

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 6/23/2021
Customer: NE-DIST
RQ: RQ-2021-05-31-23
Collected By: Jeremy Parrish, Jessica Lee
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
ALLIGATOR CREEK Bradford rd.	 G1NE0942

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 6/23/2021
1600 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

**Rec'd/Inspected
Date:**



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Calf Creek at SE 27th st. **Date:** 6/16/2021
WIN/Field ID: G1NE0931 **WBID:** 2720 **ENR:** - **Latitude:** 29.6296
County: ALACHUA **Org ID:** 21FLA **Longitude:** -82.29
Receiving Body of Water: Ocklawaha **RQ:** 2021-05-31-23

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish	x		x		
☒	Jessica Lee		x			
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# STITCH					
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time: _____	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID: _____			

Blank

Time Zone: _____	Time: _____	Field ID: _____	(Blank Type_DDDMMYY)
DI Source: _____		Location: _____	
DI Type: _____		Equipment ID: _____	
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: _____ WIN Import ID: _____
Enter Field Parameters, Verify Station ID In LIMS DMT: _____ QC Station ID, Field Parameters In LIMS DMT: bd BD 07/07/2021
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets _____ Migrate Data to WIN: _____ Verify Data In WIN: _____
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-05-31-23

Customer: WQAP

Project ID: SMP

Collected By: Jeremy Parrish, Jessica Lee

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G1NE0931

Location: Calf Creek at SE 27th st.

Comments: Microbiology sample(s) submitted to contract laboratory

R code recent rain

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
6/16/2021	1135 EST	7.66	89.9	23.3	7.31	0.06	0.12 ☒ VOB (S)	0.12	153	0.07
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # Exp:	☐ Ice ☐ H2SO4*	☐ <2		
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter Filter Lot # Syringe Lot #	☐ Ice			
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	B
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C PKD-QN-BV	☐ Ice			

Name:
Jeremy Parrish
Jessica Lee

Signature:

Jeremy Parrish

Date:

6/16/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

9143 Philips Highway, Suite 200 Jacksonville, FL 32256 / Ph(904) 739-2277 / FAX (904) 739-2011

Page. 1 of 1

SR #

A.I.S. Contact: Mandy Sullivan

[illegible]