

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: BRUCE

RQ: 2021-03-29-09

Project: Swannock North

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Brian Davis	4/6/21	836	21.7	765.3	8.9	8.9	100.7	✓	1.04	P/F	L/F
ICV			838	21.9	765.4	8.8	8.82	100.8	✓	1.04	P/F	L/F
CCV			1647	22.7	762.8	8.7	8.61	99.8	✓	1.04	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Brian Davis	4/6/21	842	201222B	1/22	1000	1000	P/F	L/F
ICV			845	201222A	1/22	100	103.6	P/F	L/F
CCV			1651	201222B	1/22	1000	997	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Brian Davis	4/6/21	855	2010F27	10/22	7.01	22.4	7.01	-9.0	P/F	L/F
Calibr.			901	2010C31	9/22	4.00	22.4	4.00	168.9	P/F	L/F
Calibr.			905	2012797	5/22	10.03	22.3	10.03	-181.5	P/F	L/F
ICV			909	2010F27	10/22	7.01	22.3	7.05	-9.0	P/F	L/F
CCV			1655	2010C31	9/22	4.00	22.5	4.11	161.9	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Brian Davis	4/6/21	913	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: <u>4/6/2021</u> Customer: <u>NE-DIST</u> RQ: <u>RQ-2021-03-29-09</u>	Collected By: <u>Kimberly Applebee, Brian Davis</u> Sampling Agency: <u>FDEP</u> Project ID: <u>SMP</u>
---	--

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
DEEP CREEK @ HYWY 441	 21010051
DEEP CREEK at farm	 G1NE0940
SWIFT CREEK @ US #41	 21020039
SWIFT CREEK abv. SE 154 ave.	 G1NE0939
CAMP BRANCH AT CR 25A	 21010054
Rocky Creek at CR136	 21010016
LITTLE CREEK AT US 441	 21010033

RELINQUISHED BY(SIGNATURE):

Kimberly Appleby

DATE/TIME: 4/6/2021

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:	Rec'd/Inspected Date:
----------------------------	------------------------------



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: CAMP BRANCH AT CR 25A **Date:** 4/6/2021
WIN/Field ID: 21010054 **WBID:** 3401 **ENR:** - **Latitude:** 30.3864
County: HAMILTON **Org ID:** 21FLA **Longitude:** -82.8774
Receiving Body of Water: Suwannee **RQ:** 2021-03-29-09

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Applebee	x		x		
☒	Brian Davis		x		x	
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID#					
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A (WIN ID_DUP)
WIN DMT Activity ID:		

Blank

Time Zone:	Time:	Field ID: (Blank Type_DDDMMYY)
DI Source:	Location:	
DI Type:	Equipment ID:	
DI Container (Name/ID):		DI Container Type:

LIMS EVENT ID: WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN:
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-03-29-09

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Applebee, Brian Davis

Field Report Prepared By: Kimberly Applebee

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



21010054

Location: CAMP BRANCH AT CR 25A

Comments: Microbiology sample(s) submitted to contract laboratory

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
4/6/2021	1355 EST	9.09	94.2	17.1	7.67	0.05	0.1 ☒ VOB (S)	0.1	171.9	0.08
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # _____ Exp: _____	☐ Ice ☐ H2SO4*	☐ <2		
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # _____ Exp: _____	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter Filter Lot # _____ Syringe Lot # _____	☐ Ice			
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	B
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Kimberly Applebee
Brian Davis

Signature:

Kimberly Applebee

Date:

4/6/2021

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

SR 1

ALS Contact: Mandy Sullivan

Page 1 of 1

[illegible]