

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSF 9551

RQ: 2021-11-08-49

Project: Pellier Trails

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 8/16/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J Pansh	11/8/21	730	18.7	765	9.9	9.4	100.6	—	1.069	P/F	L/F
ICV	L	11/8/21	731	18.7	765	9.4	9.4	100.7	—	1.069	P/F	L/F
CCV	Kim Appleby	11/8/21	1713	19.9	765.4	9.2	9.20	100.9	—	1.069	P/F	L/F
CCV			1613								P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J Pansh	11/8/21	733	210917B	9/22	1000	1000	P/F	L/F
ICV	L		736	210426A	5/22	100	102	P/F	L/F
CCV	Kim Appleby		1515	210917B	9/22	1000	1000	P/F	L/F
CCV			1615				1000	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J Pansh	11/8/21	739	2107FS6	7/23	7.	19.8	7.00	-26.7	P/F	L/F
Calibr.	L		741	2107640	7/23	4.	20.1	4.00	153	P/F	L/F
Calibr.	L		743	2012E08	6/22	10.	20.1	10.00	198	P/F	L/F
ICV	L		745	2107640	7/23	4.	20.0	4.02	150	P/F	L/F
CCV	Kim Appleby		1618	2012E08	6/22	10.04	20.8	9.96	-193.5	P/F	L/F
CCV	L		1621	2107FS6	7/23	7.02	21.2	6.98	-20.9	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 106, slope from 4 to 7 89.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/16/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J Pansh	11/8/21	747	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: