SMP

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

2021-11-05-49 Mavue Pelican mb

Meter ID: YSI Stitch RQ-2021-10-25-63 Project: Pelican Boat

Boldly "X" this box if there are qualified data on this page.

Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 11/10/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	JParish	11/22/21	805	22.3	761	8.7	8.7	100.1	—	1.019	(P) / F	L / F
ICV	"	11/22/21	806	22.3	761	8.7	8.72	100.2	—	1.019	(P) / F	(L) / F
CCV	JParish	11/22/21	1405	22.5	758	8.6	8.63	99.6	—	—	(P) / F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L. **Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	JParish	11/22/21	807	21091713	9/22	1000	1000	(P) / F	(L) / F
ICV	JParish	11/22/21	810	21042614	5/22	100	98.5	(P) / F	(L) F
CCV	JParish	11/22/21	1407	21091713	9/22	1000	996	(P) / F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria ± 5%

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	JParish	11/22/21	815	21091725	9/23	7.	22.8	7.01	-37	(P) / F	(L) F
Calibr.	"	"	816	21076401	7/23	4.	22.9	4.00	143	(P) / F	(L) F
Calibr.	"	"	817	20126081	6/22	10.	22.8	10.03	-204.5	(P) / F	(L) / F
ICV	"	"	818	21076401	7/23	4.	22.8	9.96	-203	(P) / F	(L) / F
CCV	JParish	11/22/21	1408	20126081	6/22	10.	22.3	10.16	-215	(P) / F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50; mV pH 4 Range +180 ± 50; mV pH 10 Range -180 ± 50;
If mV are recorded: slope from 7 to 10 167.5, slope from 4 to 7 180 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)
If **YES**, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 11/10/21
If **NO**, what will be used? (circle one) **Secchi Disk Line / Sonar** Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	JParish	11/22/21	819	0.00	0.00	(P) / F	(L) / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria ± 5 % or ± 0.05m, whichever is greater.

COMMENTS: