From:	Mandy Sullivan
To:	DEP-Overflowmicro
Cc:	Sprague, Kara; Baker, Abigail L
Subject:	RE: [EXTERNAL] - RE: Lab Report and EDD for J2100294 (RQ-2021-01-25-08)
Date:	Wednesday, February 10, 2021 10:59:13 AM
Attachments:	image002.png image003.png image004.png image005.png image006.png

---

Hi Kara,

I received this from our technical manager, let me know how we should proceed:

While I do see that marine samples need to be diluted 10x per the FDEP Tech Specs, I do not see a need to J qualify the data for this sample because I do not believe the data quality to be impacted. The Enterolert media clearly indicates that high salt concentrations impact the results for Enterococcus if the sample is not diluted 10x. The Colilert media has no such indication listed.

I can flag the data if you would like, but I do not think it is strictly necessary.

Thanks!

Mandy M. Sullivan
Project Manager, Environmental
USA

T +1 904 739 2277 **D** +1 904 394 4406
F +1 904 739 2011 **M** +1 904 463 9118
mandy.sullivan@alsglobal.com
9143 Philips Highway, Suite 120
Jacksonville, Florida 32256

Subscribe to Webinar Wednesdays

Right Solutions · Right Partner
alsglobal.com | **How was your ALS experience?**

From: DEP-Overflowmicro [mailto:DEP-Overflowmicro@dep.state.fl.us]
Sent: February 9, 2021 15:19
To: Mandy Sullivan <Mandy.Sullivan@alsglobal.com>
Cc: Sprague, Kara <Kara.Sprague@FloridaDEP.gov>; Baker, Abigail L <Abigail.L.Baker@FloridaDEP.gov>
Subject: [EXTERNAL] - RE: Lab Report and EDD for J2100294 (RQ-2021-01-25-08)

CAUTION: This email originated from outside of ALS. Do not click links or open attachments unless you recognize the sender and are sure content is relevant to you.

Hi Mandy,

Can you take a look at the result for field ID G5NE0582? I can see where the sampler initially

indicated it as fresh but then let you all know it should have been marine. The sample was already set up so a dilution wasn't done. The RFP from the contract requires a dilution be done on marine Quanti-Tray samples. I know you all didn't know until after the fact and it was due to sampler error but we would J-qualify this sample. Let me know what you think.

Thank you,

Kara

Kara Sprague
Environmental Specialist II
FDEP Laboratory - Biology Program
Florida Department of Environmental Protection
Kara.Sprague@Floridadep.gov
Office: 850.245.8826

From: Mandy Sullivan <Mandy.Sullivan@alsglobal.com>
Sent: Friday, January 29, 2021 5:52 PM
To: Tull, Carina <Carina.Tull@FloridaDEP.gov>; DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>; Parrish, Jeremy M. <Jeremy.M.Parrish@FloridaDEP.gov>; Savage, Jacqueline <Jacqueline.Savage@dep.state.fl.us>
Subject: Lab Report and EDD for J2100294 (RQ-2021-01-25-08)

Mandy M. Sullivan
Project Manager, Environmental
USA

T +1 904 739 2277 D +1 904 394 4406
F +1 904 739 2011 M +1 904 463 9118
mandy.sullivan@alsglobal.com
9143 Philips Highway, Suite 120
Jacksonville, Florida 32256

Subscribe to Webinar Wednesdays

Right Solutions · Right Partner
alsglobal.com | **How was your ALS experience?**