

Login Checklist

RQ- **2021 - 03 - 22 - 13**

Login Station ID: **# 3**

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: **03-26-2021 / 09:45**

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
0.9C				5		X			9293 8008 3576

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes **X** No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes _____ No **X**

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA **X**

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes **X** No _____

Caps on tight?

Yes **X** No _____

Sufficient Sample Volume:

Yes **X** No _____

If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes _____ No **X** Init: **ABS/**

Chlorine detected? (One container checked per Field ID)

Yes _____ No _____ Init: **ABS/**

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes **X** No _____ NA _____ Init: **ABS/**

**W-1-4-DIOX (Green dot on container) preserved to pH<4.0?

Yes _____ No _____ NA **X** Init: **ABS/**

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: **ABS/**

Date: **03-26-2021**

Event contains NCRs (Y/N)? **N**

NE-

Event ID:

CEN - DIST - 2021 - 03 - 26 - 0

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IRTG # 3 EXP:10/28/2021
pH Test Strips 0 – 6	213316 BV EXP:05/15/21
pH Paper 0 – 3	N/A
pH Paper 0 – 13	231019 EXP:11/01/2021
Chlorine Test Strips	0261 2023 / 05

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: ABS

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 3/25/2021
Customer: NE-DIST
RQ: RQ-2021-03-22-13

Collected By: Kimberly Applebee, Brian Davis
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
MOULTRIE CR @ US 1	 27010055
RED HOUSE BR @ LEWIS SPEEDWAY	 27010059
RED HOUSE BR @ HEFFERON RD	 27010076
Mill Creek at Powerline bridge	 G2NE1216

RELINQUISHED BY(SIGNATURE):

DATE/TIME:

3/25/21

Kimberly Appleby

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: A.SHAIK

Rec'd/Inspected Date: 03-26-2021

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Pesi

RQ: 2021-01-22-13

Project: St. Augustine

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/15/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Brian Davis</u>	<u>3/25/21</u>	<u>841</u>	<u>22.2</u>	<u>762.5</u>	<u>8.7</u>	<u>8.75</u>	<u>100.3</u>	<u>✓</u>	<u>1.06</u>	<u>(P) F</u>	<u>(L) F</u>
ICV	<u>↓</u>	<u>↓</u>	<u>843</u>	<u>22.1</u>	<u>762.7</u>	<u>8.8</u>	<u>8.75</u>	<u>100.3</u>	<u>✓</u>	<u>1.06</u>	<u>(P) F</u>	<u>(L) F</u>
CCV	<u>↓</u>	<u>↓</u>	<u>1503</u>	<u>23.1</u>	<u>760.9</u>	<u>8.6</u>	<u>8.56</u>	<u>99.8</u>	<u>✓</u>	<u>1.06</u>	<u>(P) F</u>	<u>(L) F</u>
CCV											<u>P / F</u>	<u>L / F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Brian Davis</u>	<u>3/25/21</u>	<u>845</u>	<u>201222B</u>	<u>1/22</u>	<u>1000</u>	<u>1000</u>	<u>(P) F</u>	<u>(L) F</u>
ICV			<u>847</u>	<u>200407B</u>	<u>4/21</u>	<u>20,000</u>	<u>19621</u>	<u>(P) F</u>	<u>(L) F</u>
CCV			<u>1505</u>	<u>201222C</u>	<u>1/22</u>	<u>50,000</u>	<u>49826</u>	<u>(P) F</u>	<u>(L) F</u>
CCV								<u>P / F</u>	<u>L / F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Brian Davis</u>	<u>3/25/21</u>	<u>853</u>	<u>2010F87</u>	<u>10/22</u>	<u>7.01</u>	<u>23.2</u>	<u>7.01</u>	<u>-52.7</u>	<u>(P) F</u>	<u>(L) F</u>
Calibr.			<u>900</u>	<u>2010C31</u>	<u>9/22</u>	<u>4.00</u>	<u>23.1</u>	<u>4.00</u>	<u>145.8</u>	<u>(P) F</u>	<u>(L) F</u>
Calibr.			<u>903</u>	<u>2012797</u>	<u>5/2022</u>	<u>10.02</u>	<u>23.2</u>	<u>10.02</u>	<u>-20.3</u>	<u>(P) F</u>	<u>(L) F</u>
ICV			<u>906</u>	<u>2010F87</u>	<u>10/22</u>	<u>7.01</u>	<u>23.2</u>	<u>7.04</u>	<u>-22.5</u>	<u>(P) F</u>	<u>(L) F</u>
CCV			<u>1507</u>	<u>2012797</u>	<u>5/22</u>	<u>10.02</u>	<u>23.3</u>	<u>10.02</u>	<u>-20.3</u>	<u>(P) F</u>	<u>(L) F</u>
CCV										<u>P / F</u>	<u>L / F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____ slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>Brian Davis</u>	<u>3/25/21</u>	<u>908</u>	<u>0.00</u>	<u>0.00</u>	<u>(P) F</u>	<u>(L) F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: MOULTRIE CR @ US 1

Date: 3/25/2021

WIN/Field ID: 27010055

WBID: 2493B

ENR: -

Latitude: 29.8219

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.3224

Receiving Body of Water: Upper East Coast

RQ- 2021-03-22-13

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Check
☒	Brian Davis		☒		☒	
☒	Kimberly Applebee	☒		☒	☒	☒
☐						
☐						

Sampling Equipment	
☐ Sample Container	☐ Carboy
☐ Dipper	☒ Van Dorn He-Man
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☒ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: No	Flow: FLOW	Tide: FALLING	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WAS UNABLE TO GET AN ACCURATE SECCHI MEASUREMENT DUE TO STRONG CURRENTS			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2021-03-22-13


Customer: WQAP

Project ID: SMP

Collected By: Brian Davis, Kimberly Applebee

Field Report Prepared By: Brian Davis

Sampling Agency: FDEP

Place Label Here						Comments: Microbiology sample(s) submitted to contract laboratory				
WIN ID/Field ID:  27010055										
Location: MOULTRIE CR @ US 1										
Matrix: Marine						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/25/2021	1105 EST	6.44	78.9	20.3	7.46	0.3	0 ☐ VOB (S)	1.132	20350	17.44
							Central Lab please use top row for login purposes			
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	A
		Lot # SA0293040		Exp:		10/19/21				
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☐ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	A
		Lot # NA0344060		Exp:		1/4/22				
Filtered Nutrients (P125ML)		☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice			1	A
		Filter Lot # R0NB35407								
		Syringe Lot #								
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☒ Ice			1	A
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☐ E Coli ☐ F Coli ☒ Enterococci ☒ Contract Lab				☒ Ice			1	A
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice				
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI W-E8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-SQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-DIELDRIN				☐ Ice				
Phytoplankton/Periphyton		DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C				☐ Ice				
Name: Brian Davis Kimberly Applebee						Signature: 			Date: 3/25/21	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: RED HOUSE BR @ LEWIS SPEEDWAY

Date: 3/25/2021

WIN/Field ID: 27010059

WBID: 2472

ENR: -

Latitude: 29.9293

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.3427

Receiving Body of Water: Upper East Coast

RQ: 2021-03-22-13

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Check
☒	Kimberly Applebee	☒	☒	☒	☒	☒
☒	Brian Davis	☒	☒	☒	☒	☒
☐		☐	☐	☐	☐	☐
☐		☐	☐	☐	☐	☐

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded	Meter ID# DESI		
Photos: No	Flow: FLOW	Tide: FALLING	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI	Other:		
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN_ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):	DI Container Type:		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2021-03-22-13

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Applebee, Brian Davis

Field Report Prepared By: Kimberly Applebee

Sampling Agency: FDEP

Place Label Here						Comments: Microbiology sample(s) submitted to contract laboratory				
WIN ID/Field ID:  27010059										
Location: RED HOUSE BR @ LEWIS SPEEDWAY										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/25/2021	1225 EST	7.03	78.4	20.6	7.64	0.15	0.32 ☒ VOB (S)	0.32	560	0.27
	EST						Central Lab please use top row for login purposes			
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #			Exp:					
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #			Exp:					
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab				☒ Ice			1	B
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice				
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI W-E8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-SQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-DIELDRIIN				☐ Ice				
Phytoplankton/Periphyton		DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C				☐ Ice				
Name: Kimberly Applebee Brian Davis						Signature: 			Date: 3/25/21	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: RED HOUSE BR @ HEFFERON RD

Date: 3/25/2021

WIN/Field ID: 27010076

WBID: 2472

ENR: -

Latitude: 29.927

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.3482

Receiving Body of Water: Upper East Coast

RQ- 2021-03-22-13

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Applebee	☒	☒	☒		
☒	Brian Davis		☒	☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: No	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_ODMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2021-03-22-13

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Applebee, Brian Davis

Field Report Prepared By: Kimberly Applebee

Sampling Agency: FDEP

Place Label Here						Comments: Microbiology sample(s) submitted to contract laboratory				
WIN ID/Field ID:  27010076										
Location: RED HOUSE BR @ HEFFERON RD										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/25/2021	1240 EST	6.94	77.8	20.8	7.54	0.2	0.36 ☒ VOB (S)	0.36	555	0.27
	EST						Central Lab please use top row for login purposes			
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #			Exp:					
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #			Exp:					
Filtered Nutrients (P125ML)		☐ W-PC4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-C / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab				☒ Ice			1	B
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice				
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI W-E8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-SQ-R W-CAR8-AA (5 ml of MCAA** Buffer Solution) W-DIELDRIIN				☐ Ice				
Phytoplankton/Periphyton		DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C				☐ Ice				
Name: Kimberly Applebee Brian Davis						Signature: 			Date: 3/25/21	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Mill Creek at Powerline bridge

Date: 3/25/2021

WIN/Field ID: G2NE1216

WBID: 2460

ENR: -

Latitude: 29.9411

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.5235

Receiving Body of Water: Lower St. Johns

RQ- 2021-03-22-13

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Check
☒	Kimberly Applebee		x		x	x
☒	Brian Davis		x		x	
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: MeterStick	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: YES	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):	DI Container Type:		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2021-03-22-13

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Applebee, Brian Davis

Field Report Prepared By: Kimberly Applebee

Sampling Agency: FDEP

Place Label Here						Comments: Microbiology sample(s) submitted to contract laboratory				
WIN ID/Field ID: 										
G2NE1216										
Location: Mill Creek at Powerline bridge										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/25/2021	1345 EST	3.73	41.1	19.8	7.06	0.3	0.9 ☐ VOB (S)	1.36	422.2	0.2
	EST						Central Lab please use top row for login purposes			
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #	Exp:							
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #	Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab				☒ Ice			1	B
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice				
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI WE8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-SQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-DIELDIN				☐ Ice				
Phytoplankton/Periphyton		DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-BV DTM-QN-C PRN-QN-C				☐ Ice				
Name: Kimberly Applebee Brian Davis						Signature: 			Date: 3/25/21	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

Page 1 of 1

of

Client Name						Project Number					
FDEP-NE ROC						RQ-2021-03-22-13					
Report To						Report OC					
Dep-Overflowmicro@dep.state.fl.us						V. Perovich					
Company Address											
8800 Baymeadows Way W											
Jacksonville, FL 32256											
Phone #						FAX #					
904-256-1700											
Sampler's Signature						Sampler's Printed Name					
B-P						D. Perovich					
CLIENT SAMPLE ID			LAB ID			SAMPLING DATE TIME			Matrix		
27010055						4/25/21 1105			SW		
27010059						1225			1		
27010076						1240			1		
27030474						1315			1		
62NE1216						1345			1		
ANALYSIS REQUESTED (Include Method Number and Container Preservative)											
NUMBER OF CONTAINERS											
E Coli											
Enterocoli											
Preservative Key											
0 None											
1 HCL											
2 HN03											
3 H2SO4											
4 NaOH											
5 Zn Acetate											
6 MeOH											
7 NaHSO4											
8 ICE											
REMARKS											
Sc/1											
Flow											
↑											

Date of Request: 23-FEB-2021
 Created By: PARRISH_J On: 23-FEB-2021
 Modified By: JASROTIA_P On: 02-MAR-2021
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: St.Augustine

Send Coolers To:

Phone: 904-256-1693
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: WATTS_J On: 01-MAR-2021
 Biology Request Reviewed By: JASROTIA_P On: 02-MAR-2021
 Sampling Kit Required: YES Ship on: 12-MAR-2021
 Sampling Kit Shipped: YES On: 08-MAR-2021
 Sampling Kit Packed By: SHAIK_A
 Preservatives Needed: NO
 Date to Receive Samples: 22-MAR-2021
 Logged In By: SHAIK_A On: 26-MAR-2021

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf-Salt) Nutrients, Metals, Bacteria
 Group E:(Surf.Fresh) Bacteria

NE-ALS-ECQ
 NE-ALS-ENQ

Packing Notes:

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-BR-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Bromide in aqueous matrices
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
NE-ALS-ENQ	Template: DEFAULT	(ASTM D650399) Enterococci by Enterolert/QT by ALS in Jacksonville
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO ₃
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Manganese		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3
 W-ICPMS Template: (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Aluminum
 Antimony
 Arsenic
 Beryllium
 Cadmium
 Chromium
 Copper
 Lead
 Nickel
 Selenium
 Silver
 Thallium

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4
 W-NH3 Template: DEFAULT (EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
 W-NO2NO3 Template: DEFAULT (EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
 W-S-A-TP Template: DEFAULT (EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
 W-TKN Template: DEFAULT (EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
 W-TOC Template: DEFAULT (SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE
 W-PO4-F Template: DEFAULT (EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group B (Water) With 4 Samples:

Bottle Type: P-250ML-WI-THI Number of Bottles: 4 Preserved With: ICE
 NE-ALS-ECQ Template: DEFAULT (SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville