

RQ-2021-12-06-35

Login Checklist

Login Station ID: #2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 12/2/2021 10:45

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
2.8				6		✓			929380120059

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☐ No ☐

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☐ No ☐Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☒ No ☐ Init: IM

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☒ Init: IM -12/2/21

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ☐ No ☐ NA ☒ Init: IM

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes ☐ No ☐ NA ☒ Init: IM

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: IM

Date: 12/2/2021

Event contains NCRs (Y/N)? ☒

Event ID:

St. Augustine Moses makeup
NE-DIST-2021-12-02-03 (SMP)

Date Received: 02-DEC-2021 10:45

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #2 exp:10/19/22
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	219019 exp: 7/1/22
Chlorine Test Strips	0070 exp: 02/2023

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check (W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 12/1/2021 **Collected By:** Kimberly Appleby, Katrin Villinger
Customer: NE-DIST **Sampling Agency:** FDEP
RQ: RQ-2021-12-06-35 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
RED HOUSE BR @ LEWIS SPEEDWAY	 27010059
MOULTRIE CR @ US 1	 27010055
MOSES CREEK @ CR 206	 27010051
MOSES CR AT US 1	 27010050


RELINQUISHED BY(SIGNATURE):



DATE/TIME: 12/1/2021
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:



Rec'd/Inspected Date:

12/2/21

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: **DESI**

RQ: **2021-12-06-35**

Project: **SMP St. Augustine**

Make-up

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification **11/10/21**

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Katrin Villinger	12/1/21	720	20.8	766.3	9.0	—	100.8	—	1.06762	(P) F	(L) F
ICV	↓	↓	725	20.8	766.4	9.0	9.04	101.1	—	1.06762	(P) F	(L) F
CCV	Kim Appleby	↓	1515	21.2	764.3	8.9	8.97	101.1	—	1.06762	(P) F	(L) F
CCV	↓	↓									P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Katrin Villinger	12/1/21	730	210917B	9/22	1000	1000	(P) F	(L) F
ICV	↓	↓	735	21122A	12/22	15000	14984	(P) F	(L) F
CCV	Kim Appleby	↓	1516	21122B	12/22	1413	1406	(P) F	(L) F
CCV	↓	↓	1518	210917L	9/22	50000	49770	(P) F	(L) F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Katrin Villinger	12/1/21	740	210912S	9/23	7.01	22.6	7.01	-22.54	(P) F	(L) F
Calibr.	↓	↓	750	4076410	7/23	4.00	22.4	4.00	154.4	(P) F	(L) F
Calibr.	↓	↓	755	2012E08	6/22	10.03	22.4	10.03	-194.7	(P) F	(L) F
ICV	↓	↓	800	210912S	9/23	7.01	22.6	7.05	-23.4	(P) F	(L) F
CCV	Kim Appleby	↓	1524	2012E08	6/22	10.02	23.3	10.06	-194.4	(P) F	(L) F
CCV	↓	↓								P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 **171.9**, slope from 4 to 7 **177.2** (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? **YES** / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: **11/10/21**

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Brian Davis	11/30/21	913	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: *calibrated for depth previous day (11/30/21).*



RQ-2021-12-06-35


Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Katrin Villinger

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  27010059										
Location: RED HOUSE BR @ LEWIS SPEEDWAY										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
12/1/2021	940 EST	7.95	77.7	14.3	7.56	0.3	0.568 ☒ VOB (S)	0.568	604	0.29
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #		Exp:						
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #		Exp:						
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC		Formalin Lot #		☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab				☒ Ice			1	B
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice				
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI W-E8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-TQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution)				☐ Ice				
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-BV				☐ Ice				
Name: Kimberly Appleby, Katrin Villinger						Signature: 			Date: 12/1/2021	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PLAN FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: RED HOUSE BR @ LEWIS SPEEDWAY

Date: 12/1/2021

WIN/Field ID: 27010059

WBID: 2472

ENR: -

Latitude: 29.9293

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.34

Receiving Body of Water: Upper East Coast

RQ: 2021-12-06-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	☒	☒	☒	☒	
☒	Katrin Villinger			☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☒ Other Bridge	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# DESI	
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:	
Biological Samples Collected:		☒ None	☐ HA	☐ SCI	☐ RPS	☐ Algae ID	☐ LVS ☐ LVI Other:
SBIO ID Numbers:							
Notes: Weather: sunny							

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):	DI Container Type:		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT: QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets: 12/1/2021 Migrate Data to WIN: Verify Data in WIN:



RQ-2021-12-06-35


Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Katrin Villinger

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here						Comments:					
WIN ID/Field ID:  27010055											
Location: MOULTRIE CR @ US 1											
Matrix: Marine						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)	
12/1/2021	1020 EST	7.09	81.8	15	7.6	0.3	NA (see comments) ☐ VOB (S)	1.458	38725	24.67	
	EST										
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2			
		Lot #		Exp:							
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2			
		Lot #		Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #									
		Syringe Lot #									
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice					
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-C / W-COLOR / W-SO4-IC / W-TDS				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice					
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC		Formalin Lot #		☐ Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☐ E Coli ☐ F Coli ☒ Enterococci				☒ Ice			1	A	
		☐ Contract Lab									
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3				☐ Ice					
		☐ PCR-GULL2 ☐ PCR-GFD ☐									
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice					
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI		W-E8321-MS		W-PSNP-TQ		W-CT-CL-R		☐ Ice	
		W-PCL-TQ-R		W-CAR8-AA (5 ml of MCAA** Buffer Solution)							
Phytoplankton/Periphyton				DTY-QN-C						☐ Ice	
				PKD-QN-BV							
Name: Kimberly Appleby, Katrin Villinger					Signature: 					Date: 12/1/2021	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PLAN FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: MOULTRIE CR @ US 1

Date: 12/1/2021

WIN/Field ID: 27010055

WBID: 2493B

ENR:

Latitude: 29.821908

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.322447

Receiving Body of Water: -

RQ- 2021-12-06-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	☒		☒		
☒	Katrin Villinger		☒		☒	
☐						
☐						

Sampling Equipment	
☐ Sample Container	☐ Carboy
☐ Dipper	☒ Van Dorn <u>He-Man</u>
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☒ Other <u>Bridge</u>	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: NO	Flow: Flowing	Tide: FALLING	Tide Times: High: 550 Low: 1300
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: UNABLE TO MEASURE SECCHI DUE TO SWIFT CURRENT. WEATHER: SUNNY			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):	DI Container Type:		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets KV 12/1/2021 Migrate Data to WIN: Verify Data In WIN:



RQ-2021-12-06-35

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Katrin Villinger

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID: 										
27010051										
Location: MOSES CREEK @ CR 206										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
12/1/2021	1100 EST	5.15	49	13	7.33	0.05	0.1 ☒ VOB (S)	0.1	619	0.3
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #		Exp:						
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #		Exp:						
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC		☐ Formalin Lot #		☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice			1	C
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3				☐ Ice				
		☐ PCR-GULL2 ☐ PCR-GFD ☐								
Tracers (BG-500ML)		☒ W-E8321-DI ☒ W-E8321-MS				☒ Ice			1	C
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI		WE8321-MS		W-PSNP-TQ		W-CT-CL-R		☐ Ice
Phytoplankton/Periphyton		W-PCL-TQ-R		W-CARB-AA (5 ml of MCAA** Buffer Solution)						
						DTY-QN-C		PKD-QN-BV		☐ Ice
Name: Kimberly Appleby, Katrin Villinger						Signature: 			Date: 12/1/2021	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PLAN FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: MOSES CREEK @ CR 206

Date: 12/1/2021

WIN/Field ID: 27010051

WBID: 2535B

ENR: -

Latitude: 29.7563

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.32

Receiving Body of Water: Upper East Coast

RQ: 2021-12-06-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby		☒		☒	
☒	Katrin Villinger		☒	☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Secchi	
Secchi Equipment ID: D4	
Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☐ Normal ☒ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# DESI	
Photos: NO		Flow: FLOW		Tide:		Tide Times: High: Low:	
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:							
SBIO ID Numbers:							
Notes: Weather: sunny							

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets *KV* 12/1/2021 Migrate Data to WIN: Verify Data In WIN:

(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-12-06-35

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Katrin Villinger

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  27010050										
Location: MOSES CR AT US 1										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
12/1/2021	1130 EST	7.16	69.1	13.6	7.32	0.1	0.218 ☒ VOB (S)	0.218	484.6	0.24
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4*		☐ <2		
		Lot #		Exp:						
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3*		☐ <2		
		Lot #		Exp:						
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #								
		Syringe Lot #								
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC		Formalin Lot #		☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ Contract Lab		☐ F Coli ☐ Enterococci		☒ Ice			1	C
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-GULL2		☐ PCR-BacR ☐ PCR-GFD ☐ PCR-DG3		☐ Ice				
Tracers (BG-500ML)		☒ W-E8321-DI		☒ W-E8321-MS		☒ Ice			1	C
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ W-PCL-TQ-R		☐ WE8321-MS ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution)		☐ Ice				
Phytoplankton/Periphyton				☐ DTY-QN-C ☐ PKD-QN-BV		☐ Ice				
Name: Kimberly Appleby, Katrin Villinger					Signature: 				Date: 12/1/2021	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PLAN FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: MOSES CR AT US 1

Date: 12/1/2021

WIN/Field ID: 27010050

WBID: 2535B

ENR:

Latitude: 29.774583

County: ST. JOHNS

Org ID: 21FLA

Longitude: -81.316139

Receiving Body of Water: -

RQ: 2021-12-06-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	☒		☒		
☒	Katrin Villinger		☒	☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: NO	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: SUNNY			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):	DI Container Type:		

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets *KV* 12/1/2021 (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data in WIN: (Initials/Date)

Date of Request: 18-NOV-2021
 Created By: PARRISH_J On: 18-NOV-2021
 Modified By: JASROTIA_P On: 19-NOV-2021
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: St.Augustine Moses makeup

Send Coolers To:

Phone: 904-256-1693
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: A
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 19-NOV-2021
 Biology Request Reviewed By: JASROTIA_P On: 19-NOV-2021
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped:
 Sampling Kit Packed By:
 Preservatives Needed: NO
 Date to Receive Samples: 06-DEC-2021
 Logged In By:

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf-Salt) Bacteria
 Group B:(Surf.Fresh) Bacteria
 Group C:(Fresh)Bacteria Tracers

Packing Notes:

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI Number of Bottles: 1 Preserved With: ICE
 ENT-24-QT Template: DEFAULT (Enterolert/QT) Enterococci by Enterolert Quanti-Tray method - non-chlorinated water systems

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI Number of Bottles: 1 Preserved With: ICE
 ECOLI-18QT Template: DEFAULT (SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI Number of Bottles: 1 Preserved With: ICE
 ECOLI-18QT Template: DEFAULT (SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Type: BG-500ML Number of Bottles: 1 Preserved With: ICE
 W-E8321-DI Template: DEFAULT (EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
 W-E8321-MS Template: SMP (EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS