

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Craw Daddy

RQ- 2521-02-08-19
2021-03-22-48

Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 3/22/20 NA 3/5/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	3/22/21	0824	21.86	753.8	8.87	8.7	100.0	/	1.05	⓪/F	⓪/F
ICV	↓	↓	↓	21.6	753.8	8.7	8.70	98.8	/	/	⓪/F	⓪/F
CCV	Natalie Ayala	3/22/21	1534	23.9	753.2	8.4	8.46	100.3	/	/	⓪/F	⓪/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	3/22/21	0828	201026A	11/21	1000	1000	⓪/F	⓪/F
ICV	↓	↓	↓	210127A	02/22	100	104.0	⓪/F	⓪/F
CCV	Natalie Ayala	3/22/21	1540	210127C	02/22	50000	49778	⓪/F	⓪/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	3/22/21	0840	202114	5/22	7.02	21.1	7.02	-52.5	⓪/F	⓪/F
Calibr.	↓	↓	↓	200052	2/22	4.00	21.1	4.00	126.0	⓪/F	⓪/F
Calibr.	↓	↓	↓	203539	6/22	10.05	21.1	10.05	-228.5	P/F	L/F
ICV	↓	↓	↓	203539	6/22	10.05	21.1	10.03	-228.2	⓪/F	⓪/F
CCV	Natalie Ayala	3/22/21	1541	200052	02/22	4.00	21.9	4.03	124.7	⓪/F	⓪/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 176, slope from 4 to 7 178.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 3/5/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Knotthead; Date of last verification: 12/4/20

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Natalie Ayala	3/22/21	0813	0.272	0.273	⓪/F	⓪/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Shrimphrap

RQ: 2021-03-15-64

Project: SMD

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 3/5/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	3/22/21	8:28	21.7	753.8	8.7	8.70	100.0	✓	1.108	P/F	D/F
ICV	↓	↓	↓	21.7	753.8	8.7	8.72	99.2	✓	✓	P/F	D/F
CCV	Urantes	3/22/21	14:00	22.7	753.5	8.6	9.27	107.4	✓	✓	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	3/22/21	0835	201026A	1/21	1000	1000	P/F	D/F
ICV	↓	↓	↓	2010127A	02/22	100	103.7	P/F	D/F
CCV	Urantes	3/22/21	14:03	2010127A	02/22	100	104.7	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Natalie Ayala	3/22/21	8:40	202114	5/22	7.02	21.1	7.02	-59.0	P/F	D/F
Calibr.	↓	↓	↓	200052	5/22	4.00	21.2	4.00	120.2	P/F	D/F
Calibr.	↓	↓	↓	203539	06/22	10.05	21.1	10.05	-233.0	P/F	D/F
ICV	↓	↓	↓	203539	06/22	10.05	21.1	10.03	-232.9	P/F	D/F
CCV	Urantes	3/22/21	14:05	200052	2/22	4.00	21.7	4.08	115.20	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 174, slope from 4 to 7 179.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 3/5/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Light; Date of last verification: 12/4/20

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Natalie Ayala	3/22/21	0851	0.272	0.271	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

DO Failed Post Cal *