

Chemical Analysis Report

CEN-DIST-2021-04-22-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Adelaide/Lotus/Asher/Florida**
Request ID: **RQ-2021-02-15-08**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 12-MAY-2021 11:19

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Hourglass Lake @ Center

Collection Date/Time: 04/21/2021 12:14

Field ID: G4CE0178

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2241491	EPA 180.1 Rev. 2.0	Turbidity	4.8		NTU	P397006	
	EPA 300.0 Rev. 2.1	Chloride	22		mg Cl/L	P397125	
		Sulfate	17		mg SO4/L	P397128	
	SM 2120 B-2011	Color (true)	19		PCU	P397002	
	SM 2320 B-2011	Total Alkalinity	52		mg CaCO3/L	P397193	
	SM 2540 C-2011	TDS	127		mg/L	P397416	
	SM 2540 D-2011	TSS	7	I	mg/L	P397164	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P397194	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P397513	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P397111	
2241492	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P397400	
	EPA 365.1 Rev. 2.0	Total-P	0.090		mg P/L	P397095	
	SM 5310 B-2011	Organic Carbon	6.5		mg C/L	P397241	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	I	mg P/L	P396999	
	dissolved						

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P397006

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397125

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397128

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397513

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397111

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397400

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397095

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396999

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P397002

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P397193

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P397416

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P397164

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P397194

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P397241

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397125

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397128

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397513

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397111

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.3		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397400

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397095

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396999

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.6		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P397002

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.7		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P397193

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	100		P	93 - 104

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P397194

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.8		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P397241

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.1		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397125

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241460	Chloride	105	105	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397128

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241460	Sulfate	100	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397513

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241365	Ammonia-N	98.4	98.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397111

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241365	Kjeldahl Nitrogen	103	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397400

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241492	NO2NO3-N	102	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397095

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241492	Total-P	108	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396999

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241462	O-Phosphate-P	99.1	98.9	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P397194

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241721	Fluoride	100	98.1	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P397241

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2241322	Organic Carbon	97.2	97.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P397006

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241426	Turbidity	8.31	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397125

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241460	Chloride	0.0233	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P397128

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241460	Sulfate	0.200	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397513

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241365	Ammonia-N	0.406	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397111

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241365	Kjeldahl Nitrogen	3.04	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397400

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241492	NO2NO3-N	1.46	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397095

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241492	Total-P	2.53	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396999

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241462	O-Phosphate-P	0.202	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P397002

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241515	Color (true)	2.35	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P397193

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241721	Total Alkalinity	0.560	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P397416

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241389	TDS	4.52	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P397194

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241721	Fluoride	1.41	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P397241

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2241322	Organic Carbon	0.226	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A104869

Included Lab Sample IDs: 2241491

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.8	100	P/P	90 - 110
Sulfate	98.5	98.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A104915

Included Lab Sample IDs: 2241493

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.5	100	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A104916

Included Lab Sample IDs: 2241491

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	100	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A104918

Included Lab Sample IDs: 2241491

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	98.0	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A104985

Included Lab Sample IDs: 2241492

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	103	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A105004

Included Lab Sample IDs: 2241491

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.0	100	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A105004

Included Lab Sample IDs: 2241491

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	101	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A105019

Included Lab Sample IDs: 2241492

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.0	95.4	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A105021

Included Lab Sample IDs: 2241492

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	96.3	96.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A105082

Included Lab Sample IDs: 2241492

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A105117

Included Lab Sample IDs: 2241492

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.9	97.5	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					8.31	
EPA 300.0 Rev. 2.1	Chloride	105	105	105			0.0233
	Sulfate	102	100	101			0.200
EPA 350.1 Rev. 2.0	Ammonia-N	96.8	98.4	98.8			0.406
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.3	103	100			3.04
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	102	104			1.46
EPA 365.1 Rev. 2.0	Total-P	105	108	105			2.53
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.6	99.1	98.9			0.202
SM 2120 B-2011	Color (true)	99.7				2.35	
SM 2320 B-2011	Total Alkalinity	100				0.560	
SM 2540 C-2011	TDS					4.52	
SM 4500 F-C-2011	Fluoride	96.8	100	98.1			1.41
SM 5310 B-2011	Organic Carbon	96.1	97.2	97.6			0.226

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2241491
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2241491
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2241491
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2241492
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2241492
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2241492
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2241492
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2241493
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2241491
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2241491
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2241491
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2241491
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2241491
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2241492

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	04/22/2021			04/22/2021 17:06	Yen Ta	2241491
EPA 300.0 Rev. 2.1	04/22/2021			04/24/2021 05:03	Lipi Saha	2241491
EPA 350.1 Rev. 2.0	04/22/2021			05/01/2021 12:23	Ping Hua	2241492
EPA 351.2 Rev. 2.0	04/22/2021	04/25/2021 14:48	Benjamin T. Nordmann	04/27/2021 14:28	Virginia P. Brown	2241492
EPA 353.2 Rev. 2.0	04/22/2021			04/29/2021 10:36	Beverly Y. Sanders	2241492
EPA 365.1 Rev. 2.0	04/22/2021	04/24/2021 18:49	Madeline Gruver	04/26/2021 14:11	Shengyu Ding	2241492
EPA 365.1 Rev. 2.0 dissolved	04/22/2021			04/22/2021 15:20	Patrick M. Roche	2241493
SM 2120 B-2011	04/22/2021			04/22/2021 15:04	Madeline Gruver	2241491
SM 2320 B-2011	04/22/2021			04/26/2021 22:00	Dale L. Simmons	2241491
SM 2540 C-2011	04/22/2021			04/23/2021 15:31	Yijie Li	2241491
SM 2540 D-2011	04/22/2021			04/23/2021 15:31	Yijie Li	2241491
SM 4500 F-C-2011	04/22/2021			04/26/2021 22:00	Dale L. Simmons	2241491
SM 5310 B-2011	04/22/2021			04/26/2021 21:55	Lauren Lees	2241492