

Chemical Analysis Report

CEN-DIST-2021-03-19-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lake Emerald / Lake Giles**
Request ID: **RQ-2021-03-08-09**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 12-APR-2021 13:59



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Terrace @ Center

Collection Date/Time: 03/18/2021 12:19

Field ID: G4CE0153

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2232665	EPA 180.1 Rev. 2.0	Turbidity	4.1	A	NTU	P395218	
	EPA 300.0 Rev. 2.1	Chloride	21		mg Cl/L	P395466	
		Sulfate	26		mg SO4/L	P395468	
	SM 2120 B-2011	Color (true)	16	A	PCU	P395235	
	SM 2320 B-2011	Total Alkalinity	35		mg CaCO3/L	P395328	
	SM 2540 C-2011	TDS	118		mg/L	P395619	
	SM 2540 D-2011	TSS	4	I	mg/L	P395623	
	SM 4500 F-C-2011	Fluoride	0.047	I	mg F/L	P395332	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P395763	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.95		mg N/L	P395534	
2232666	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P395336	
	EPA 365.1 Rev. 2.0	Total-P	0.037		mg P/L	P395326	
	SM 5310 B-2011	Organic Carbon	6.3		mg C/L	P395581	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	U	mg P/L	P395241	
	dissolved						

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P395218

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395466

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395468

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395763

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395534

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395336

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395326

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395241

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P395235

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P395328

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P395619

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P395623

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P395332

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P395581

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P395466

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P395468

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P395763

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P395534

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P395336

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P395326

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	109		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P395241

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2120 B-2011
 Batch ID: P395235

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-2011
 Batch ID: P395328

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	93 - 104

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P395332

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.2		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P395581

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395466

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232577	Chloride	103		P	90 - 110
2232746	Chloride	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395468

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232577	Sulfate	101		P	90 - 110
2232746	Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395763

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232575	Ammonia-N	101	99.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395534

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232747	Kjeldahl Nitrogen	107	99.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395336

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232641	NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395326

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232641	Total-P	109	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395241

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232643	O-Phosphate-P	100	101	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P395332

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232751	Fluoride	98.3	98.3	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P395581

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2232673	Organic Carbon	95.6	95.8	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P395218

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232665	Turbidity	5.87	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395466

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232577	Chloride	0.743	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395468

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232577	Sulfate	0.0839	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395763

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232575	Ammonia-N	1.80	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395534

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232747	Kjeldahl Nitrogen	5.82	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395336

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232641	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395326

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232641	Total-P	1.87	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395241

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232643	O-Phosphate-P	1.19	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P395235

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232665	Color (true)	4.31	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P395328

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232751	Total Alkalinity	0.0288	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P395619

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232437	TDS	1.37	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P395332

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232751	Fluoride	0.0	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P395581

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2232673	Organic Carbon	0.0820	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A104168

Included Lab Sample IDs: 2232665

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	102	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A104175

Included Lab Sample IDs: 2232665

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.6	99.4	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A104176

Included Lab Sample IDs: 2232667

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.8	100	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A104210

Included Lab Sample IDs: 2232665

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	97.2	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A104210

Included Lab Sample IDs: 2232665

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	99.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A104215

Included Lab Sample IDs: 2232666

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A104216

Included Lab Sample IDs: 2232665

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.5	99.7	P/P	90 - 110
Sulfate	99.4	99.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A104247

Included Lab Sample IDs: 2232666

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.5	98.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-2011

Run ID: A104316

Included Lab Sample IDs: 2232666

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.6	97.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A104381

Included Lab Sample IDs: 2232666

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.5	97.3	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A104401

Included Lab Sample IDs: 2232666

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.0	99.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					5.87	
EPA 300.0 Rev. 2.1	Chloride	103	103	104		0.743	
	Sulfate	102	101	102		0.0839	
EPA 350.1 Rev. 2.0	Ammonia-N	98.8	101	99.2			1.80
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	107	99.3			5.82
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	102	102			0.0
EPA 365.1 Rev. 2.0	Total-P	109	109	107			1.87
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	100	101			1.19
SM 2120 B-2011	Color (true)	101				4.31	
SM 2320 B-2011	Total Alkalinity	101				0.0288	
SM 2540 C-2011	TDS					1.37	
SM 4500 F-C-2011	Fluoride	99.2	98.3	98.3			0.0
SM 5310 B-2011	Organic Carbon	95.0	95.6	95.8			0.0820

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2232665
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2232665
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2232665
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2232666
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2232666
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2232666
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2232666
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2232667
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2232665
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2232665
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2232665
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2232665
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2232665
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2232666

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/19/2021			03/19/2021 14:00	Yen Ta	2232665
EPA 300.0 Rev. 2.1	03/19/2021			03/24/2021 21:48	Lipi Saha	2232665
EPA 350.1 Rev. 2.0	03/19/2021			03/31/2021 17:28	Ping Hua	2232666
EPA 351.2 Rev. 2.0	03/19/2021	03/26/2021 13:02	Benjamin T. Nordmann	03/30/2021 15:46	Virginia P. Brown	2232666
EPA 353.2 Rev. 2.0	03/19/2021			03/23/2021 09:33	Beverly Y. Sanders	2232666
EPA 365.1 Rev. 2.0	03/19/2021	03/23/2021 09:29	Yen Ta	03/24/2021 11:37	Shengyu Ding	2232666
EPA 365.1 Rev. 2.0 dissolved	03/19/2021			03/19/2021 15:22	Blanca Fach	2232667
SM 2120 B-2011	03/19/2021			03/19/2021 14:53	Benjamin T. Nordmann	2232665
SM 2320 B-2011	03/19/2021			03/22/2021 19:48	Dale L. Simmons	2232665
SM 2540 C-2011	03/19/2021			03/22/2021 14:26	Yijie Li	2232665
SM 2540 D-2011	03/19/2021			03/22/2021 14:26	Yijie Li	2232665
SM 4500 F-C-2011	03/19/2021			03/22/2021 19:48	Dale L. Simmons	2232665
SM 5310 B-2011	03/19/2021			03/27/2021 05:20	Madeline Gruver	2232666