

Chemical Analysis Report

CEN-DIST-2021-03-26-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Bugg Spring**
Request ID: **RQ-2021-03-15-08**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 20-APR-2021 13:41

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Bugg Spring @ Vent

Collection Date/Time: 03/25/2021 12:07

Field ID: G1CE0109

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2233845	EPA 180.1 Rev. 2.0	Turbidity	0.50		NTU	P395558	
	EPA 300.0 Rev. 2.1	Chloride	16	A	mg Cl/L	P395841	
		Sulfate	8.0	A	mg SO4/L	P395842	
	SM 2120 B-2011	Color (true)	2.7	I	PCU	P395561	
	SM 2320 B-2011	Total Alkalinity	138		mg CaCO3/L	P395640	
	SM 2540 C-2011	TDS	180		mg/L	P396070	
	SM 2540 D-2011	TSS	2	U	mg/L	P396074	
	SM 4500 F-C-2011	Fluoride	0.055	I	mg F/L	P395642	
2233846	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P395914	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.15	I	mg N/L	P395965	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.57		mg N/L	P395649	
	EPA 365.1 Rev. 2.0	Total-P	0.082		mg P/L	P395731	
	SM 5310 B-2011	Organic Carbon	1.2		mg C/L	P395712	
2233847	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.076		mg P/L	P395549	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P395558

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395841

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395842

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395914

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395965

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395649

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395731

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395549

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P395561

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P395640

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P396070

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P396074

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P395642

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P395712

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395841

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395842

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395914

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395965

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395649

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395731

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395549

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.4		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P395561

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P395640

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.9		P	93 - 104

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P395642

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.3		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P395712

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395841

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233845	Chloride	99.9		P	90 - 110
2234385	Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395842

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233845	Sulfate	98.1		P	90 - 110
2234385	Sulfate	99.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395914

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233846	Ammonia-N	96.0	97.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395965

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233825	Kjeldahl Nitrogen	106	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395649

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233535	NO2NO3-N	93.8	94.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395731

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233804	Total-P	106	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395549

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233861	O-Phosphate-P	99.5	98.6	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P395642

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233850	Fluoride	95.1	95.9	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P395712

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233610	Organic Carbon	95.4	97.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P395558

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233856	Turbidity	6.00	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P395841

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233845	Chloride	0.356	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P395842

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233845	Sulfate	0.416	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P395914

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233846	Ammonia-N	1.22	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P395965

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233825	Kjeldahl Nitrogen	3.96	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P395649

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233535	NO2NO3-N	0.491	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P395731

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233804	Total-P	2.62	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P395549

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233861	O-Phosphate-P	0.909	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P395561

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233798	Color (true)	1.63	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P395640

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233850	Total Alkalinity	0.969	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P396070

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233708	TDS	3.81	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P395642

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233850	Fluoride	0.457	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P395712

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233610	Organic Carbon	1.27	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A104301

Included Lab Sample IDs: 2233847

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	98.8	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A104305

Included Lab Sample IDs: 2233845

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	102	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A104306

Included Lab Sample IDs: 2233845

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	101	P/P	90 - 115

Reference Method: SM 2320 B-2011

Run ID: A104346

Included Lab Sample IDs: 2233845

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.8	108	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A104346

Included Lab Sample IDs: 2233845

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.2	96.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A104356

Included Lab Sample IDs: 2233846

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	100	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A104385

Included Lab Sample IDs: 2233846

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.0	97.0	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A104389

Included Lab Sample IDs: 2233845

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.9	98.8	P/P	90 - 110
Sulfate	97.4	97.7	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A104412

Included Lab Sample IDs: 2233846

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	104	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A104465

Included Lab Sample IDs: 2233846

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.2	97.9	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A104509

Included Lab Sample IDs: 2233846

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	108	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					6.00	
EPA 300.0 Rev. 2.1	Chloride	100	99.9	100		0.356	
	Sulfate	99.3	98.1	99.9		0.416	
EPA 350.1 Rev. 2.0	Ammonia-N	96.4	96.0	97.4			1.22
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	100	106	102			3.96
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	93.8	94.2			0.491
EPA 365.1 Rev. 2.0	Total-P	104	106	104			2.62
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.4	99.5	98.6			0.909
SM 2120 B-2011	Color (true)	101				1.63	
SM 2320 B-2011	Total Alkalinity	98.9				0.969	
SM 2540 C-2011	TDS					3.81	
SM 4500 F-C-2011	Fluoride	97.3	95.1	95.9			0.457
SM 5310 B-2011	Organic Carbon	95.4	95.4	97.4			1.27

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2233845
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2233845
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2233845
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2233846
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2233846
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2233846
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2233846
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2233847
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2233845
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2233845
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2233845
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2233845
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2233845
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2233846

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/26/2021			03/26/2021 14:46	Yen Ta	2233845
EPA 300.0 Rev. 2.1	03/26/2021			04/01/2021 11:53	Lipi Saha	2233845
EPA 350.1 Rev. 2.0	03/26/2021			04/03/2021 11:59	Ping Hua	2233846
EPA 351.2 Rev. 2.0	03/26/2021	04/05/2021 09:18	Yen Ta	04/06/2021 11:43	Virginia P. Brown	2233846
EPA 353.2 Rev. 2.0	03/26/2021			03/30/2021 09:58	Beverly Y. Sanders	2233846
EPA 365.1 Rev. 2.0	03/26/2021	03/31/2021 13:21	Yen Ta	03/31/2021 17:57	Shengyu Ding	2233846
EPA 365.1 Rev. 2.0 dissolved	03/26/2021			03/26/2021 14:15	Lipi Saha	2233847
SM 2120 B-2011	03/26/2021			03/26/2021 15:49	Benjamin T. Nordmann	2233845
SM 2320 B-2011	03/26/2021			03/29/2021 19:08	Dale L. Simmons	2233845
SM 2540 C-2011	03/26/2021			03/30/2021 14:36	Yijie Li	2233845
SM 2540 D-2011	03/26/2021			03/30/2021 14:36	Yijie Li	2233845
SM 4500 F-C-2011	03/26/2021			03/29/2021 19:08	Dale L. Simmons	2233845
SM 5310 B-2011	03/26/2021			03/31/2021 06:48	Madeline Gruver	2233846