

Chemical Analysis Report

CEN-DIST-2021-09-02-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lake Downey**
Request ID: **RQ-2021-06-07-12**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 28-SEP-2021 15:58



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: SJR Upstream of Lake Hellen Blazes

Collection Date/Time: 09/01/2021 12:30

Field ID: G3CE0018

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2273273	EPA 180.1 Rev. 2.0	Turbidity	3.0	A	NTU	P403333	
	EPA 300.0 Rev. 2.1	Chloride	110		mg Cl/L	P403499	
		Sulfate	22		mg SO4/L	P403500	
	SM 2120 B-2011	Color (true)	220		PCU	P403335	
	SM 2320 B-2011	Total Alkalinity	102		mg CaCO3/L	P403437	
	SM 2540 C-2011	TDS	382		mg/L	P403573	
	SM 2540 D-2011	TSS	3	I	mg/L	P403569	
	SM 4500 F-C-2011	Fluoride	0.26		mg F/L	P403438	
2273274	EPA 350.1 Rev. 2.0	Ammonia-N	0.023		mg N/L	P403421	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.8		mg N/L	P403509	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P403349	
	EPA 365.1 Rev. 2.0	Total-P	0.14		mg P/L	P403427	
	SM 5310 B-2011	Organic Carbon	30		mg C/L	P403549	
2273275	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.050		mg P/L	P403337	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0: Total-P: Batch matrix spike recoveries and relative percent difference are unavailable because of analytical difficulties.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P403333

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403499

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403500

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403421

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403509

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403349

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P403427

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P403337

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P403335

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P403437

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P403573

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P403569

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P403438

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P403549

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403499

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403500

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403421

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403509

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403349

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P403427

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P403337

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.6		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P403335

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.1		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P403437

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	94 - 106

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P403438

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.5		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P403549

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403499

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273152	Chloride	99.6		P	90 - 110
2273276	Chloride	99.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403500

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273152	Sulfate	100		P	90 - 110
2273276	Sulfate	99.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403421

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273358	Ammonia-N	98.2	96.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403349

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273335	NO2NO3-N	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P403337

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273268	O-Phosphate-P	100	99.2	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P403438

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273302	Fluoride	97.6	100	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P403549

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2273256	Organic Carbon	95.6	95.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P403333

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273273	Turbidity	12.4	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403499

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273152	Chloride	0.969	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403421

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273358	Ammonia-N	1.02	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403509

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273208	Kjeldahl Nitrogen	2.92	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403349

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273335	NO2NO3-N	0.496	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P403337

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273268	O-Phosphate-P	0.495	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P403335

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273302	Color (true)	0.772	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P403437

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273302	Total Alkalinity	1.28	Sample	P	0 - 5.2

Quality Assurance Report Precision

Reference Method: SM 2540 C-2011
Batch ID: P403573

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273264	TDS	1.01	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P403438

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273302	Fluoride	1.90	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P403549

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2273256	Organic Carbon	0.137	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A107623

Included Lab Sample IDs: 2273273

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A107624

Included Lab Sample IDs: 2273273

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	101	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A107625

Included Lab Sample IDs: 2273275

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.4	97.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A107639

Included Lab Sample IDs: 2273274

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A107663

Included Lab Sample IDs: 2273274

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	98.6	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A107669

Included Lab Sample IDs: 2273273

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	104	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A107669

Included Lab Sample IDs: 2273273

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A107673

Included Lab Sample IDs: 2273274

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A107690

Included Lab Sample IDs: 2273273

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	101	P/P	90 - 110
Sulfate	99.6	99.7	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A107704

Included Lab Sample IDs: 2273274

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.2	98.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A107739

Included Lab Sample IDs: 2273274

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	105	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					12.4	
EPA 300.0 Rev. 2.1	Chloride	100	99.6	99.5		0.969	
	Sulfate	100	100	99.9			
EPA 350.1 Rev. 2.0	Ammonia-N	98.8	98.2	96.8			1.02
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105					2.92
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	103	100	101			0.496
EPA 365.1 Rev. 2.0	Total-P	105					
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.6	100	99.2			0.495
SM 2120 B-2011	Color (true)	99.1				0.772	
SM 2320 B-2011	Total Alkalinity	103				1.28	
SM 2540 C-2011	TDS					1.01	
SM 4500 F-C-2011	Fluoride	99.5	97.6	100			1.90
SM 5310 B-2011	Organic Carbon	94.8	95.6	95.4			0.137

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2273273
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2273273
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2273273
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2273274
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2273274
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2273274
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2273274
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2273275
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2273273
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2273273
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2273273
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2273273
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2273273
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2273274

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/02/2021			09/02/2021 13:29	Sarah A Nixon	2273273
EPA 300.0 Rev. 2.1	09/02/2021			09/08/2021 17:32	Roberta Tatti	2273273
EPA 350.1 Rev. 2.0	09/02/2021			09/07/2021 13:02	Ping Hua	2273274
EPA 351.2 Rev. 2.0	09/02/2021	09/09/2021 12:30	Benjamin T. Nordmann	09/10/2021 11:34	Virginia P. Brown	2273274
EPA 353.2 Rev. 2.0	09/02/2021			09/03/2021 10:08	Beverly Y. Sanders	2273274
EPA 365.1 Rev. 2.0	09/02/2021	09/07/2021 15:50	Simonne.V. Calhoun	09/08/2021 10:56	Shengyu Ding	2273274
EPA 365.1 Rev. 2.0 dissolved	09/02/2021			09/02/2021 16:51	James L Waggerby	2273275
SM 2120 B-2011	09/02/2021			09/02/2021 15:19	Sarah A Nixon	2273273
SM 2320 B-2011	09/02/2021			09/07/2021 20:54	Sarah A Nixon	2273273
SM 2540 C-2011	09/02/2021			09/03/2021 14:31	Yijie Li	2273273
SM 2540 D-2011	09/02/2021			09/03/2021 14:31	Yijie Li	2273273
SM 4500 F-C-2011	09/02/2021			09/07/2021 20:54	Sarah A Nixon	2273273
SM 5310 B-2011	09/02/2021			09/09/2021 03:42	Madeline Gruver	2273274