

Chemical Analysis Report

CEN-DIST-2021-08-20-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Bugg Spring/Apopka Spring**
Request ID: **RQ-2021-07-19-47**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Matt Bearden

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 16-SEP-2021 17:00



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Bugg Spring @ vent

Collection Date/Time: 08/19/2021 10:15

Field ID: G1CE0109

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2271140	EPA 180.1 Rev. 2.0	Turbidity	0.35		NTU	P402776	
	EPA 300.0 Rev. 2.1	Chloride	18		mg Cl/L	P403164	
		Sulfate	9.6		mg SO4/L	P403166	
	SM 2120 B-2011	Color (true)	3.2	I	PCU	P402778	
	SM 2320 B-2011	Total Alkalinity	143	A	mg CaCO3/L	P402851	
	SM 2540 C-2011	TDS	187		mg/L	P403218	
	SM 2540 D-2011	TSS	2	U	mg/L	P403200	
	SM 4500 F-C-2011	Fluoride	0.055	I	mg F/L	P402854	
2271142	EPA 350.1 Rev. 2.0	Ammonia-N	0.01		mg N/L	P403081	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.17	I	mg N/L	P403261	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.55		mg N/L	P403070	
	EPA 365.1 Rev. 2.0	Total-P	0.070		mg P/L	P402893	
	SM 5310 B-2011	Organic Carbon	1.3		mg C/L	P403372	
2271144	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.061		mg P/L	P402771	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Apopka Spring near boil

Collection Date/Time: 08/19/2021 10:07

Field ID: G1CE0113

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2271141	EPA 180.1 Rev. 2.0	Turbidity	5.8		NTU	P402776	
	EPA 300.0 Rev. 2.1	Chloride	28		mg Cl/L	P403164	
		Sulfate	11		mg SO4/L	P403166	
	SM 2120 B-2011	Color (true)	33	A	PCU	P402779	
	SM 2320 B-2011	Total Alkalinity	110		mg CaCO3/L	P402851	
	SM 2540 C-2011	TDS	192	A	mg/L	P403218	
	SM 2540 D-2011	TSS	13		mg/L	P403200	
	SM 4500 F-C-2011	Fluoride	0.20		mg F/L	P402854	
2271143	EPA 350.1 Rev. 2.0	Ammonia-N	0.093		mg N/L	P403081	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.0		mg N/L	P403261	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.084		mg N/L	P403070	
	EPA 365.1 Rev. 2.0	Total-P	0.050		mg P/L	P402893	
	SM 5310 B-2011	Organic Carbon	15		mg C/L	P403372	
2271145	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P402771	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P402776

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403164

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403166

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403081

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403261

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403070

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402893

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402771

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P402778

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2120 B-2011
Batch ID: P402779

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P402851

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P403218

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P403200

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P402854

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P403372

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403164

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403166

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	104		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403081

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403261

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.6		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403070

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402893

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402771

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	107		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P402778

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2120 B-2011
Batch ID: P402779

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	103		P	85 - 115

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-2011
Batch ID: P402851

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P402854

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P403372

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403164

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271098	Chloride	102		P	90 - 110
2271402	Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403166

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271098	Sulfate	101		P	90 - 110
2271402	Sulfate	98.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403081

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271142	Ammonia-N	94.2	94.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403261

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271142	Kjeldahl Nitrogen	99.4	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403070

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271099	NO2NO3-N	99.0	98.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402893

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271053	Total-P	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402771

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271145	O-Phosphate-P	95.6	97.1	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P402854

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271140	Fluoride	99.9	100	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P403372

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2271056	Organic Carbon	97.0	96.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P402776

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271052	Turbidity	0.0	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403164

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271098	Chloride	0.864	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P403166

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271098	Sulfate	1.14	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P403081

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271142	Ammonia-N	0.527	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P403261

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271142	Kjeldahl Nitrogen	4.09	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P403070

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271099	NO2NO3-N	0.506	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402893

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271053	Total-P	0.336	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402771

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271145	O-Phosphate-P	1.56	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
 Batch ID: P402778

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2270864	Color (true)	1.02	Sample	P	0 - 20

Reference Method: SM 2120 B-2011
 Batch ID: P402779

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271141	Color (true)	2.02	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
 Batch ID: P402851

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271140	Total Alkalinity	1.71	Sample	P	0 - 5.2

Reference Method: SM 2540 C-2011
 Batch ID: P403218

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271141	TDS	3.64	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
 Batch ID: P402854

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271140	Fluoride	0.472	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
 Batch ID: P403372

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2271056	Organic Carbon	0.421	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A107370

Included Lab Sample IDs: 2271144, 2271145

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	102	P/P	90 - 110
O-Phosphate-P	102	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A107371

Included Lab Sample IDs: 2271140, 2271141

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	102	102	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A107372

Included Lab Sample IDs: 2271140, 2271141

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 115
Color (true)	102	102	P/P	90 - 115

Reference Method: SM 2320 B-2011

Run ID: A107408

Included Lab Sample IDs: 2271140, 2271141

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110
Total Alkalinity	103	101	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A107408

Included Lab Sample IDs: 2271140, 2271141

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	101	P/P	90 - 110
Fluoride	103	103	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A107449

Included Lab Sample IDs: 2271140, 2271141

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.5	99.4	P/P	90 - 110
Sulfate	99.7	99.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A107480

Included Lab Sample IDs: 2271142, 2271143

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A107506

Included Lab Sample IDs: 2271142, 2271143

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	99.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A107511

Included Lab Sample IDs: 2271142, 2271143

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.8	97.7	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A107619

Included Lab Sample IDs: 2271142, 2271143

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.8	98.4	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A107644

Included Lab Sample IDs: 2271142, 2271143

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.0	98.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					0.0	
EPA 300.0 Rev. 2.1	Chloride	102	102	100		0.864	
	Sulfate	104	98.3	101		1.14	
EPA 350.1 Rev. 2.0	Ammonia-N	97.4	94.2	94.8			0.527
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.6	99.4	104			4.09
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.5	99.0	98.5			0.506
EPA 365.1 Rev. 2.0	Total-P	106	102	103			0.336
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	107	95.6	97.1			1.56
SM 2120 B-2011	Color (true)	101				1.02	
	Color (true)	103				2.02	
SM 2320 B-2011	Total Alkalinity	101				1.71	
SM 2540 C-2011	TDS					3.64	
SM 4500 F-C-2011	Fluoride	101	99.9	100			0.472
SM 5310 B-2011	Organic Carbon	96.8	97.0	96.2			0.421

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2271140, 2271141
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2271140, 2271141
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2271140, 2271141
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2271142, 2271143
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2271142, 2271143
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2271142, 2271143
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2271142, 2271143
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2271144, 2271145
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2271140, 2271141
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2271140, 2271141
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2271140, 2271141
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2271140, 2271141
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2271140, 2271141
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2271142, 2271143

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/20/2021			08/20/2021 16:40	Sarah A Nixon	2271140, 2271141
EPA 300.0 Rev. 2.1	08/20/2021			08/27/2021 21:03	Lipi Saha	2271140
	08/20/2021			08/27/2021 21:15	Lipi Saha	2271141
EPA 350.1 Rev. 2.0	08/20/2021			08/27/2021 13:47	Ping Hua	2271142
	08/20/2021			08/27/2021 13:59	Ping Hua	2271143
EPA 351.2 Rev. 2.0	08/20/2021	09/01/2021 12:09	Benjamin T. Nordmann	09/02/2021 12:32	Alexandra J Mattheus	2271142
	08/20/2021	09/01/2021 12:09	Benjamin T. Nordmann	09/02/2021 12:37	Alexandra J Mattheus	2271143
EPA 353.2 Rev. 2.0	08/20/2021			08/27/2021 10:40	Beverly Y. Sanders	2271142
	08/20/2021			08/27/2021 10:41	Beverly Y. Sanders	2271143
EPA 365.1 Rev. 2.0	08/20/2021	08/24/2021 14:32	Simonne.V. Calhoun	08/25/2021 18:23	Shengyu Ding	2271142
	08/20/2021	08/24/2021 14:32	Simonne.V. Calhoun	08/25/2021 18:24	Shengyu Ding	2271143
EPA 365.1 Rev. 2.0 dissolved	08/20/2021			08/20/2021 15:26	James L Waggeberby	2271145
	08/20/2021			08/20/2021 15:37	James L Waggeberby	2271144
SM 2120 B-2011	08/20/2021			08/20/2021 18:24	Sarah A Nixon	2271140
	08/20/2021			08/20/2021 18:29	Sarah A Nixon	2271141
SM 2320 B-2011	08/20/2021			08/23/2021 22:52	Sarah A Nixon	2271140
	08/20/2021			08/24/2021 01:47	Sarah A Nixon	2271141
SM 2540 C-2011	08/20/2021			08/23/2021 15:36	Yijie Li	2271140, 2271141
SM 2540 D-2011	08/20/2021			08/23/2021 15:36	Yijie Li	2271140, 2271141
SM 4500 F-C-2011	08/20/2021			08/23/2021 22:39	Sarah A Nixon	2271140
	08/20/2021			08/24/2021 01:47	Sarah A Nixon	2271141
SM 5310 B-2011	08/20/2021			09/03/2021 01:12	Touraj Touran	2271142
	08/20/2021			09/03/2021 01:27	Touraj Touran	2271143