

Boldly "X" this
box if there are
qualified data on
this page.

Thus

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Date of last Temperature Verification 02/11/21

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & galvanic Sensors: DO gain Range 25-75. Rapid-Pulse Sensors: DO gain Range 0.7 to 1.4; DO Charge Range 25-75. DO Acceptance criteria from Table ± 0.3 mg/L.

Conductivity Acceptance criteria $\pm 5\%$

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If NO, what will be used? (circle one) **Secchi Disk Line / Sonar** Unique ID: 33126; Date of last verification: _____

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or $\pm 0.05m$, whichever is greater.

COMMENTS: