

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: My Boys Blue

RQ: 2021-09-13-51

Project: SMP: Continuous

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 08/16/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	William Miller	09/16/21	0720	23.33	759.5	8.5	8.50	99.7		1.061	(P) F	(L) F
ICV	William Miller	09/16/21	0723	23.31	759.5	8.5	8.53	100.0			(P) F	(L) F
CCV	William Miller	09/21/21	1320	27.61	761.4	7.9	8.05	102.1			(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	William Miller	09/16/21	0724	2011243	11/22	50000	50000	(P) F	(L) F
ICV	William Miller	09/16/21	0730	1009989	08/22	2500	2570	(P) F	(L) F
CCV	William Miller	09/21/21	1322	1010077	10/22	10,000		P / F	L / F
CCV	William Miller	09/21/21	1325	1009989	08/22	50,000	50,120	(P) F	(L) F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	William Miller	09/16/21	0734	2104575	04/23	7.00	23.8	7.00		(P) F	(L) F
Calibr.	William Miller	09/16/21	0739	2107246	07/23	4.01	23.6	4.01		(P) F	(L) F
Calibr.	William Miller	09/16/21	0742	1106221	12/22	10.02	23.5	10.02		(P) F	(L) F
ICV	William Miller	09/16/21	0746	2107246	07/23	4.01	23.49	3.99		(P) F	(L) F
CCV	William Miller	09/21/21	1339	1106221	12/22	10.02	27.61	10.05		(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____ slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) **Secchi Disk** Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: