

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: AD New Seabins

NO: 2021-09-13-47

Project: SMP: Cont.

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 08/16/21

DO FDEP SOP FT 1400	Name	Date	Time CET	Temp °C	Baro- meter mmHg	D.O. Gain mg/L	Meter DO mg/L	%DO	Probe Status	Probe Serial	Pass Fail	Lab
Calibr.	William Miller	09/23/21	0720	22.68	757.4	8.6	8.60	99.6		1.092	(P/F) (L/F)	
ICV	William Miller	09/23/21	0723	22.68	757.1	8.6	8.60	99.6			(P/F) (L/F)	
CCV	Gary Snow	09/28/21	1350	23.5	761.2	8.5	8.64	101.6			(P/F) (L/F)	
CCV											P/F L/F	

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Gain FDEP 1400	Name	Date	Time CET	Temp °C	Baro- meter mmHg	Exp. Date	Standard units/cm	Meter Reading	Pass Fail	Lab
Calibr.	William Miller	09/23/21	0724	20.12	43	11/22	50.000	50.000	(P/F) (L/F)	
ICV	William Miller	09/23/21	0726	1009	89	8/22	2500	2585	(P/F) (L/F)	
CCV	Gary Snow	9/28/21	1355	4108	72	7/23	100.0	101.6	(P/F) (L/F)	
CCV									P/F L/F	

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP 550 FDEP 1400	Name	Date	Time CET	Temp °C	Baro- meter mmHg	Exp. Date	pH Buffer SU	Temp °C	Meter Reading	Pass Fail	Lab
Calibr.	William Miller	09/23/21	0731	2104	575	4/23	7.00	23.6	7.00	(P/F) (L/F)	
Calibr.	William Miller	09/23/21	0734	2107	416	07/23	4.01	23.5	4.01	(P/F) (L/F)	
Calibr.	William Miller	09/23/21	0735	1106	221	12/22	10.03	23.4	10.03	(P/F) (L/F)	
ICV	William Miller	09/23/21	0741	2107	416	07/23	11.01	23.4	4.00	(P/F) (L/F)	
CCV	Gary Snow	09/28/21	1358	1106	221	12/22	10.00	23.64	10.07	(P/F) (L/F)	
CCV										P/F L/F	

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CET	Calibrated Value (0.00 or Offset) meters	ICV Value meters	Pass Fail	Lab
Pressure mode in air						P/F L/F	

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: