

Chemical Analysis Report

SO-DIST-2021-05-18-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **2021 SMP : Wilderness Waterway**
Request ID: **RQ-2021-04-19-57**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Chris Hormuth

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 07-JUN-2021 15:34



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: ENRE10 Lopez River

Collection Date/Time: 05/17/2021 11:30

Field ID: G5SD0092

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2247829	EPA 180.1 Rev. 2.0	Turbidity	2.2		NTU	P398380	
	EPA 300.0 Rev. 2.1	Bromide	79		mg Br/L	P398642	
	SM 2320 B-2011	Total Alkalinity	169		mg CaCO3/L	P398525	
	SM 2540 D-2011	TSS	4	I	mg/L	P398608	
	SM 4500 F-C-2011	Fluoride	1.2		mg F/L	P398530	
2247831	EPA 350.1 Rev. 2.0	Ammonia-N	0.010		mg N/L	P398597	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.45		mg N/L	P398424	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P398444	
	EPA 365.1 Rev. 2.0	Total-P	0.033		mg P/L	P398541	
	SM 5310 B-2011	Organic Carbon	6.7		mg C/L	P398808	
2247833	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.015		mg P/L	P398369	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-2011: Batch relative percent difference is unavailable because of analytical difficulties.

Sample Location: ENRE9 Lopez River

Collection Date/Time: 05/17/2021 11:55

Field ID: G5SD0093

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2247830	EPA 180.1 Rev. 2.0	Turbidity	3.5		NTU	P398380	
	EPA 300.0 Rev. 2.1	Bromide	75		mg Br/L	P398642	
	SM 2320 B-2011	Total Alkalinity	177		mg CaCO3/L	P398525	
	SM 2540 D-2011	TSS	6	I	mg/L	P398608	
	SM 4500 F-C-2011	Fluoride	1.2		mg F/L	P398530	
2247832	EPA 350.1 Rev. 2.0	Ammonia-N	0.014		mg N/L	P398928	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.63		mg N/L	P398616	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P398444	
	EPA 365.1 Rev. 2.0	Total-P	0.041		mg P/L	P398541	
	SM 5310 B-2011	Organic Carbon	9.2		mg C/L	P398808	
2247834	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.018		mg P/L	P398369	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-2011: Batch relative percent difference is unavailable because of analytical difficulties.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P398380

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398642

Component	Result	Code	Units
Bromide	0.013	U	mg Br/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398597

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398928

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398424

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398616

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P398444

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P398541

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P398369

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P398525

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-2011
Batch ID: P398608

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P398530

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P398808

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398642

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Bromide	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398597

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398928

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398424

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398616

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P398444

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P398541

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P398369

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.6		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P398525

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	99.3		P	94 - 106

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P398530

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.0		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P398808

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398642

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247185	Bromide	102	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398597

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2249182	Ammonia-N	100	99.6	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398928

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2250204	Ammonia-N	102	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398424

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247079	Kjeldahl Nitrogen	107	107	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398616

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247832	Kjeldahl Nitrogen	99.5	107	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P398444

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247831	NO2NO3-N	106	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P398541

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247466	Total-P	104	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P398369

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247940	O-Phosphate-P	95.8	98.0	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P398530

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247702	Fluoride	101		P	92 - 107

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P398808

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2247022	Organic Carbon	98.5	98.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P398380

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247927	Turbidity	2.49	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398642

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247185	Bromide	2.24	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398597

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2249182	Ammonia-N	0.561	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398928

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2250204	Ammonia-N	0.375	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398424

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247079	Kjeldahl Nitrogen	0.374	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P398616

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247832	Kjeldahl Nitrogen	5.84	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P398444

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247831	NO2NO3-N	0.939	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P398541

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247466	Total-P	1.84	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P398369

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247940	O-Phosphate-P	1.90	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P398525

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2246890	Total Alkalinity	2.02	Sample	P	0 - 5.2

Reference Method: SM 5310 B-2011
Batch ID: P398808

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2247022	Organic Carbon	0.234	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A105506

Included Lab Sample IDs: 2247833, 2247834

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.2	99.5	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A105510

Included Lab Sample IDs: 2247829, 2247830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.0	99.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A105533

Included Lab Sample IDs: 2247831, 2247832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	106	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A105567

Included Lab Sample IDs: 2247829, 2247830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.8	101	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A105567

Included Lab Sample IDs: 2247829, 2247830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	99.5	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A105586

Included Lab Sample IDs: 2247829, 2247830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Bromide	100	106	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A105595

Included Lab Sample IDs: 2247831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.5	98.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A105597

Included Lab Sample IDs: 2247831, 2247832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A105619

Included Lab Sample IDs: 2247832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.2	105	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A105623

Included Lab Sample IDs: 2247831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	99.3	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A105673

Included Lab Sample IDs: 2247831, 2247832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	102	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A105732

Included Lab Sample IDs: 2247832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.1	98.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					2.49	
EPA 300.0 Rev. 2.1	Bromide	101	102	100			2.24
EPA 350.1 Rev. 2.0	Ammonia-N	98.6	100	99.6			0.561
	Ammonia-N	97.4	102	102			0.375
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	107	107			0.374
	Kjeldahl Nitrogen	100	99.5	107			5.84
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	103	106	107			0.939
EPA 365.1 Rev. 2.0	Total-P	105	104	106			1.84
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.6	95.8	98.0			1.90
SM 2320 B-2011	Total Alkalinity	99.3				2.02	
SM 4500 F-C-2011	Fluoride	99.0	101				
SM 5310 B-2011	Organic Carbon	101	98.5	98.7			0.234

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2247829, 2247830
EPA 300.0 Rev. 2.1	Bromide in aqueous matrices	2247829, 2247830
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2247831, 2247832
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2247831, 2247832
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2247831, 2247832
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2247831, 2247832
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2247833, 2247834
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2247829, 2247830
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2247829, 2247830
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2247829, 2247830
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2247831, 2247832

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/18/2021			05/18/2021 15:36	Sarah A Nixon	2247829, 2247830
EPA 300.0 Rev. 2.1	05/18/2021			05/21/2021 17:03	Lipi Saha	2247829
	05/18/2021			05/21/2021 17:22	Lipi Saha	2247830
EPA 350.1 Rev. 2.0	05/18/2021			05/21/2021 12:37	Roberta Tatti	2247831
	05/18/2021			05/27/2021 13:41	Ping Hua	2247832
EPA 351.2 Rev. 2.0	05/18/2021	05/19/2021 14:18	Benjamin T. Nordmann	05/24/2021 14:40	Virginia P. Brown	2247831
	05/18/2021	05/21/2021 13:15	Benjamin T. Nordmann	05/24/2021 11:45	Virginia P. Brown	2247832
EPA 353.2 Rev. 2.0	05/18/2021			05/19/2021 15:40	Touraj Touran	2247831
	05/18/2021			05/19/2021 15:46	Touraj Touran	2247832
EPA 365.1 Rev. 2.0	05/18/2021	05/20/2021 14:05	Yen Ta	05/21/2021 14:28	Shengyu Ding	2247831
	05/18/2021	05/20/2021 14:05	Yen Ta	05/21/2021 14:29	Shengyu Ding	2247832
EPA 365.1 Rev. 2.0 dissolved	05/18/2021			05/18/2021 15:30	Lipi Saha	2247833
	05/18/2021			05/18/2021 15:31	Lipi Saha	2247834
SM 2320 B-2011	05/18/2021			05/20/2021 04:45	Dale L. Simmons	2247829
	05/18/2021			05/20/2021 05:00	Dale L. Simmons	2247830
SM 2540 D-2011	05/18/2021			05/20/2021 15:22	Yijie Li	2247829, 2247830
SM 4500 F-C-2011	05/18/2021			05/20/2021 12:28	Dale L. Simmons	2247829
	05/18/2021			05/20/2021 12:33	Dale L. Simmons	2247830
SM 5310 B-2011	05/18/2021			05/25/2021 13:19	Madeline Gruver	2247831
	05/18/2021			05/25/2021 13:32	Madeline Gruver	2247832