

Non-Conformance BackLog Report

POWER_D

Printed: 12/15/2021 3:39:12 PM

Page 1 of 1

<u>NCR ID:</u>	<u>NCR Type:</u>	<u>Group ID:</u>	<u>Authorized By:</u>	<u>Date Authorized:</u>
9168	PREPARATION	BENCHBIO	Not Yet Authorised	Not Yet Authorised

Category: Sample Homogeneity and Integrity

Show on Final Report: Yes

Initated By: POWER_D Date Initated: 12/15/2021
Standard method for filtering not followed. No blank filtered on 10/28/2021.

Resolved By: POWER_D Date Resolved: 15-DEC-2021
Sample will be "J" qualified. Customers will be notified.

Links to the following LIMS data:

<u>Event ID:</u>	<u>Job ID:</u>	<u>Sample ID:</u>	<u>Test ID:</u>
AMBIENT-2021-10-28-03	TLH-2021-10-28-56	2286655	No Link
NW-DIST-2021-10-28-01	TLH-2021-10-28-19	2286503	No Link
SO-DIST-2021-10-28-01	TLH-2021-10-28-30	2286543	No Link
SW-DIST-2021-10-28-01	TLH-2021-10-28-40	2286583	No Link
SW-DIST-2021-10-28-01	TLH-2021-10-28-40	2286584	No Link
WQAP-2021-10-28-01	TLH-2021-10-28-36	2286561	No Link
WQAP-2021-10-28-01	TLH-2021-10-28-36	2286562	No Link
WQAP-2021-10-28-01	TLH-2021-10-28-36	2286563	No Link
WQAP-2021-10-28-01	TLH-2021-10-28-36	2286564	No Link
