

Chemical Analysis Report

NW-DIST-2021-06-22-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Twin Lakes LVI**
Request ID: **RQ-2021-06-21-27**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Myranda Butler

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 15-JUL-2021 11:28

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Coleman Lake

Collection Date/Time: 06/21/2021 12:55

Field ID: G3NW0216

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2256189	EPA 180.1 Rev. 2.0	Turbidity	4.9		NTU	P400060	
	EPA 300.0 Rev. 2.1	Chloride	25		mg Cl/L	P400506	
		Sulfate	8.3		mg SO4/L	P400507	
	SM 2120 B-2011	Color (true)	49		PCU	P400067	
	SM 2320 B-2011	Total Alkalinity	59		mg CaCO3/L	P400372	
	SM 2540 C-2011	TDS	104		mg/L	P400470	
	SM 2540 D-2011	TSS	4	I	mg/L	P400243	
	SM 4500 F-C-2011	Fluoride	0.12		mg F/L	P400375	
2256191	EPA 350.1 Rev. 2.0	Ammonia-N	0.49		mg N/L	P400409	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.8		mg N/L	P400389	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.078		mg N/L	P400095	
	EPA 365.1 Rev. 2.0	Total-P	0.32		mg P/L	P400299	
	SM 5310 B-2011	Organic Carbon	7.8		mg C/L	P400361	
2256193	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.20		mg P/L	P400063	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Kell-Aire Lake

Collection Date/Time: 06/21/2021 13:10

Field ID: G3NW0217

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2256190	EPA 180.1 Rev. 2.0	Turbidity	12		NTU	P400060	
	EPA 300.0 Rev. 2.1	Chloride	24		mg Cl/L	P400506	
		Sulfate	7.9		mg SO ₄ /L	P400507	
		Color (true)	53		PCU	P400067	
	SM 2320 B-2011	Total Alkalinity	58		mg CaCO ₃ /L	P400372	
	SM 2540 C-2011	TDS	127	A	mg/L	P400470	
	SM 2540 D-2011	TSS	8	IA	mg/L	P400243	
	SM 4500 F-C-2011	Fluoride	0.12		mg F/L	P400375	
2256192	EPA 350.1 Rev. 2.0	Ammonia-N	0.17		mg N/L	P400409	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.3		mg N/L	P400389	
	EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	0.029		mg N/L	P400095	
	EPA 365.1 Rev. 2.0	Total-P	0.21		mg P/L	P400299	
	SM 5310 B-2011	Organic Carbon	7.9		mg C/L	P400361	
2256194	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.025		mg P/L	P400063	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P400060

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P400506

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P400507

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P400409

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P400389

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P400095

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P400299

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P400063

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P400067

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P400372

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P400470

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P400243

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P400375

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P400361

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P400506

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P400507

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P400409

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P400389

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	96.6		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P400095

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P400299

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P400063

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2120 B-2011
Batch ID: P400067

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P400372

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	94 - 106

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P400375

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P400361

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P400506

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2255816	Chloride	98.6		P	90 - 110
2256031	Chloride	96.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P400507

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2255816	Sulfate	95.8		P	90 - 110
2256031	Sulfate	97.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P400409

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2256173	Ammonia-N	99.6	98.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P400389

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2256115	Kjeldahl Nitrogen	103	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P400095

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2256115	NO2NO3-N	95.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P400299

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2256173	Total-P	103	103	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P400375

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2256032	Fluoride	99.6	99.6	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P400361

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2255871	Organic Carbon	96.4	97.0	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P400060

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256135	Turbidity	1.88	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P400506

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256031	Chloride	0.439	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P400507

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256031	Sulfate	0.675	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P400409

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256173	Ammonia-N	0.442	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P400389

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256115	Kjeldahl Nitrogen	1.40	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P400095

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256115	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P400299

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256173	Total-P	0.0447	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P400063

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256193	O-Phosphate-P	0.307	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P400067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256118	Color (true)	7.99	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P400372

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256032	Total Alkalinity	0.267	Sample	P	0 - 5.2

Reference Method: SM 2540 C-2011
Batch ID: P400470

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256190	TDS	1.57	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P400375

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2256032	Fluoride	0.0	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P400361

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2255871	Organic Carbon	0.441	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A106220

Included Lab Sample IDs: 2256189, 2256190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	99.0	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A106221

Included Lab Sample IDs: 2256193, 2256194

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.2	97.7	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A106222

Included Lab Sample IDs: 2256189, 2256190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.8	99.3	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A106232

Included Lab Sample IDs: 2256191, 2256192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A106347

Included Lab Sample IDs: 2256191, 2256192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	103	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A106360

Included Lab Sample IDs: 2256191, 2256192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.1	97.2	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A106363

Included Lab Sample IDs: 2256189, 2256190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.2	104	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A106363

Included Lab Sample IDs: 2256189, 2256190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A106372

Included Lab Sample IDs: 2256191, 2256192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.5	98.1	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A106423

Included Lab Sample IDs: 2256189, 2256190

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	101	P/P	90 - 110
Sulfate	99.6	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A106442

Included Lab Sample IDs: 2256191, 2256192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	96.3	99.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					1.88	
EPA 300.0 Rev. 2.1	Chloride	102	96.3	98.6		0.439	
	Sulfate	96.9	97.6	95.8		0.675	
EPA 350.1 Rev. 2.0	Ammonia-N	101	99.6	98.8			0.442
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	96.6	103	105			1.40
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.5	95.0				0.0
EPA 365.1 Rev. 2.0	Total-P	102	103	103			0.0447
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101					0.307
SM 2120 B-2011	Color (true)	100				7.99	
SM 2320 B-2011	Total Alkalinity	102				0.267	
SM 2540 C-2011	TDS					1.57	
SM 4500 F-C-2011	Fluoride	101	99.6	99.6			0.0
SM 5310 B-2011	Organic Carbon	96.0	96.4	97.0			0.441

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2256189, 2256190
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2256189, 2256190
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2256189, 2256190
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2256191, 2256192
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2256191, 2256192
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2256191, 2256192
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2256191, 2256192
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2256193, 2256194
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2256189, 2256190
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2256189, 2256190
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2256189, 2256190
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2256189, 2256190
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2256189, 2256190
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2256191, 2256192

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/22/2021			06/22/2021 15:04	Sarah A Nixon	2256189, 2256190
EPA 300.0 Rev. 2.1	06/22/2021			07/01/2021 23:08	Lipi Saha	2256189
	06/22/2021			07/01/2021 23:20	Lipi Saha	2256190
EPA 350.1 Rev. 2.0	06/22/2021			06/30/2021 13:48	Roberta Tatti	2256191
	06/22/2021			06/30/2021 13:51	Roberta Tatti	2256192
EPA 351.2 Rev. 2.0	06/22/2021	06/30/2021 12:28	Benjamin T. Nordmann	07/02/2021 13:28	Letuzia M De Oliveira	2256191
	06/22/2021	06/30/2021 12:28	Benjamin T. Nordmann	07/02/2021 13:30	Letuzia M De Oliveira	2256192
EPA 353.2 Rev. 2.0	06/22/2021			06/23/2021 08:58	Beverly Y. Sanders	2256191
	06/22/2021			06/23/2021 08:59	Beverly Y. Sanders	2256192
EPA 365.1 Rev. 2.0	06/22/2021	06/28/2021 17:29	Madeline Gruver	06/29/2021 12:46	Shengyu Ding	2256191
	06/22/2021	06/28/2021 17:29	Madeline Gruver	06/29/2021 12:47	Shengyu Ding	2256192
EPA 365.1 Rev. 2.0 dissolved	06/22/2021			06/22/2021 15:38	Ping Hua	2256193
	06/22/2021			06/22/2021 15:42	Ping Hua	2256194
SM 2120 B-2011	06/22/2021			06/22/2021 15:48	Benjamin T. Nordmann	2256189
	06/22/2021			06/22/2021 15:49	Benjamin T. Nordmann	2256190
SM 2320 B-2011	06/22/2021			06/29/2021 23:40	Dale L. Simmons	2256189
	06/22/2021			06/29/2021 23:55	Dale L. Simmons	2256190
SM 2540 C-2011	06/22/2021			06/24/2021 14:15	Yijie Li	2256189, 2256190
SM 2540 D-2011	06/22/2021			06/24/2021 14:15	Yijie Li	2256189, 2256190
SM 4500 F-C-2011	06/22/2021			06/29/2021 23:40	Dale L. Simmons	2256189
	06/22/2021			06/29/2021 23:55	Dale L. Simmons	2256190
SM 5310 B-2011	06/22/2021			06/29/2021 09:20	Madeline Gruver	2256191
	06/22/2021			06/29/2021 09:35	Madeline Gruver	2256192