

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Builds "X" this box if there are qualified data on this page.

Meter ID: 154177

RD: 2021-12-22-35

Project: SHP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification: 12-30-20

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	N DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Paul Blair	2-25-21	7:44	20.1	769.9	9.07	9.10	100.1	/	1.13	P/F	L/F
ICV	Paul Blair	"	7:45	20.1	769.9	9.07	9.14	100.1	/	"	P/F	L/F
CCV	"	"	13:22	20.3	769.0	9.54	9.55	100.3	/	"	P/F	L/F
CCV	"	"	"	"	"	"	"	"	"	"	P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15 (Pre DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Paul Blair	7-25-21	7:45	201209D	12-21	1000	1000	P/F	L/F
ICV	"	"	7:47	2905F63	5-21	100	101.0	P/F	L/F
CCV	"	"	13:02	2905F63	5-21	100	100.7	P/F	L/F
CCV	"	"	13:03	201209D	12-21	1000	1002	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Paul Blair	7-25-21	7:48	200407C	10-21	7.01	24.4	7.02	-20.7	P/F	L/F
Calibr.	"	"	7:49	200407A	5-22	4.01	20.5	4.00	161.8	P/F	L/F
Calibr.	"	"	7:52	200407A	10-21	10.05	21.2	10.04	-191.6	P/F	L/F
ICV	"	"	7:51	"	"	10.05	20.9	10.04	-191.7	P/F	L/F
CCV	Paul Blair	7-25-21	13:06	190405E	"	"	"	"	"	P/F	L/F
CCV	"	"	"	200594	9-21	1.00	21.6	1.17	805.5	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 .
If mV are recorded: slope from 7 to 10 170.9; slope from 4 to 7 172.6 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / ☒ / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 51, 51; Date of last verification: 12-30-20

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 2/25/2021 **Collected By:** Ben Harden, Paul Blair
Customer: TLH-DIST **Sampling Agency:** FDEP
RQ: RQ-2021-02-22-35 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Middle of Cherry Lake	 G1TLHR0048
Alapaha River Rise @ Suwannee River	 G1TLHR0046
Norton Creek dstream Bandit St	 G1TLHR0147

RELINQUISHED BY(SIGNATURE):

DATE/TIME:

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



2.25.21



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Middle of Cherry Lake **Date:** 2/25/2021
WIN/Field ID: G1TLHR0048 **WBID:** 3322A **ENR:** - **Latitude:** 30.6169
County: MADISON **Org ID:** 21FLTLHR **Longitude:** -83.4153
Receiving Body of Water: Suwannee **RQ:** 2021-02-22-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Paul Blair	x		x		
☒	Ben Harden		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☒ Syringe	☐ Pole
Depth measured with: Sonar	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☒ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# 154177		
Photos: NO	Flow: NA		Tide:	Tide Times: High: Low:				
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:								
SBIO ID Numbers:								
Notes:								

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: **WIN Import ID:**
Enter Field Parameters, Verify Station ID In LIMS DMT: **QC Station ID, Field Parameters In LIMS DMT:**
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets **Migrate Data to WIN:** **Verify Data In WIN:**
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-02-22-35

Customer: WQAP

Project ID: SMP

Collected By: Paul Blair, Ben Harden

Field Report Prepared By: Paul Blair

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G1TLHR0048

Location: Middle of Cherry Lake

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
2/25/2021	1005 EST	11.76	119.8	16.3	7.25	0.3	4 ☐ VOB (S)	4.4	45.8	0.02
	1008	11.82	119.4	15.9	7.27	2.0	Central Lab please use top row for login purposes		45.8	0.02
	1011 EST	11.77	117.4	15.3	7.38	3.9			45.9	0.02

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # 0293040 Exp: 10/19/21	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD Lot # Exp:	☐ Ice ☐ HNO3*	☐ <2		
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # N/A Syringe Lot # 0079900	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	A
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☐ Ice			
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C PKD-QN-BV	☐ Ice			

Name:
Paul Blair
Ben Harden

Signature:

Date:

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Alapaha River Rise @ Suwannee River **Date:** 2/25/2021
WIN/Field ID: G1TLHR0046 **WBID:** 3341X **ENR:** - **Latitude:** 30.4376
County: HAMILTON **Org ID:** 21FLTLHR **Longitude:** -83.0896
Receiving Body of Water: Suwannee **RQ:** 2021-02-22-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Ben Harden		x		x	
☒	Paul Blair	x		x		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Secchi	
Secchi Equipment ID: #1	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☐ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☒ Flooded		Meter ID# 154177	
Photos: NO	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A (WIN ID_DUP)
WIN DMT Activity ID:		

Blank

Time Zone:	Time:	Field ID: (Blank Type_DDDMMYY)
DI Source:	Location:	
DI Type:	Equipment ID:	
DI Container (Name/ID):		DI Container Type:

LIMS EVENT ID: WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)
Scan/Save Field Sheets Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2021-02-22-35

Customer: WQAP

Project ID: SMP

Collected By: Ben Harden, Paul Blair

Field Report Prepared By: Ben Harden

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G1TLHR0046

Location: Alapaha River Rise @ Suwannee River

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
2/25/2021	1110 EST	9.46	91.1	13.5	4.9	0.15	0.3 ☒ VOB (S)	0.3	34.6	0.02
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H2SO4*	☐ <2		
	Lot # Exp:				
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD	☒ Ice ☒ HNO3*	☒ <2	1	B
	Lot # 0344060 Exp:	1-4-22			
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
	Filter Lot # Syringe Lot #				
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☐ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☐ Ice			
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C PEN-QL-C DTY-QN-C PKD-QN-C DTM-QN-C PRN-QN-C PKD-QN-BV	☐ Ice			
Name: Ben Harden Paul Blair	Signature:	Date:			

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Norton Creek dstream Bandit St **Date:** 2/25/2021
WIN/Field ID: G1TLHR0147 **WBID:** 3366 **ENR:** - **Latitude:** 30.3935
County: MADISON **Org ID:** 21FLTLHR **Longitude:** -83.34
Receiving Body of Water: Suwannee **RQ:** 2021-02-22-35

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Ben Harden		x		x	
☒	Paul Blair	x		x		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Secchi	
Secchi Equipment ID: #1	
Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# 154177					
Photos: NO	Flow: FLOW			Tide:		Tide Times: High: Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN:
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-02-22-35

Customer: WQAP

Project ID: SMP

Collected By: Ben Harden, Paul Blair

Field Report Prepared By: Ben Harden

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G1TLHR0147

Location: Norton Creek dstream Bandit St

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
2/25/2021	1150 EST	6.75	66.9	14.9	3.51	0.25	0.2 ☐ VOB (S)	0.5	60.1	0.03
							Central Lab please use top row for login purposes			
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H2SO4*	☐ <2		
	Lot # Exp:				
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD	☐ Ice ☐ HNO3*	☐ <2		
	Lot # Exp:				
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
	Filter Lot #				
	Syringe Lot #				
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☒ Ice		1	B
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R	☐ Ice			
	☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN				
Phytoplankton/Periphyton	☐ DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C	☐ Ice			
	☐ DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV ☐				

Name:
Ben Harden
Paul Blair

Signature:

Date:

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution