

RQ-2021-05-24-24

Login Checklist

Login Station ID: #2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 5/27/2021 9:45 14:40

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
1.1 0.6				17 13		✓	✓		— —

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 °C (above 10.0 °C for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☐ No ☒

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☐ No ☒Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☐ No ☒ Init: IM

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☐ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ☒ No ☐ NA ☐ Init: IM

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes ☐ No ☐ NA ☒ Init: IM

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: IM

Date: 5/27/2021

Event contains NCRs (Y/N)? ☒

Event ID: Run 7
 TLH-ROC-2021-05-27-01 (SMP)
 Date Received: 27-MAY-2021 14:40

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Infrared Thermometer ID:	IR TEMP GUN #2 exp:10/28/21
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	219019 exp: 7/1/22
Chlorine Test Strips	0129 exp: 03/2023

Additional Comments:

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 5/27/2021
Customer: TLH-DIST
RQ: RQ-2021-05-24-24

Collected By: Paul Blair, Jonathan Woods
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: HandDelivered

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Alligator Harbor South	 G2TLHR0054
Alligator Harbor near clam beds	 G2TLHR0025
Oyster Bay1	 G1TLHR0151
Spring Creek Spring	 G1TLHR0131
Oyster Bay2	 G1TLHR0152
Spring Creek Springs Middle	 G1TLHR0129

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 5/27/2021
1500EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:



Rec'd/Inspected Date:

5/27/21

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM**

Date: 5/27/2021 **Collected By:** Paul Blair, Jonathan Woods
Customer: TLH-DIST **Sampling Agency:** FDEP
RQ: RQ-2021-05-24-24 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: HandDelivered

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Alligator Harbor South	 G2TLHR0054
Alligator Harbor near clam beds	 G2TLHR0025
Oyster Bay1	 G1TLHR0151
Spring Creek Spring	 G1TLHR0131
Oyster Bay2	 G1TLHR0152
Spring Creek Springs Middle	 G1TLHR0129

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 5/27/2021
1500 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

**Rec'd/Inspected
Date:**



Customer: FDEP

Field Report Prepared by: Paul Blair, Jonathan Woods

Project ID: SMP

Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: G1TLHR0152

Comments:

Location: Oyster Bay2

Matrix: Marine

(Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. ©	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPTH)			
5/27/2021 1225	7.5	108.2	28	8.01	0.3	1	1.3	34788	21.82			
Central Lab please use top row for login purposes												
Parameter Suite	Parameter Methods				Preservation		pH Check		# Bottles sent to Lab	Bottle Group		
Preserved Nutrients (P-500ML)	X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP			X	Ice	X	H2SO4*	X	<2	1	A
	Lot #		120529		Exp:		10-20-21					
Metals (P-500ML)		W-ICPMS		W-ICP	W-HARD		Ice	HNO3*		<2		
	Lot #				Exp:							
Filtered Nutrients (P125ML)	X	W-PO4-F		X	Field Filtered 0.45 µm Filter		X	Ice			1	A
	Filter Lot #		N/A									
	Syringe Lot #		79900									
Chlorophyll (BP-1L)	X	CHLSUITE-W				X	Ice				1	A
Physical/Aggregate (Fresh) (P-1L)		Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS					Ice					
Physical/Aggregate (Marine) (P-1L)		Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					Ice					
Macroinvertebrates (PJ-2L)		MI-FW-QLDC					Ice					
Periphyton (PT-50ML)		ALGAE-ID					Ice					
Microbiology (Contract Lab Bottle)		E Coli	X	F Coli	X	Enterococci	X	Ice			1	A
	Contract Lab											
Molecular (QPCR-P-500ML)		PCR-HF1B3		PCR-BacR		PCR-DG3		Ice				
	PCR-GULL2		PCR-GFD									
Tracers (BG-500ML)		W-E8321-DI		W-E8321-MS			Ice					
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI	W-E8321-MS		W-PSNP-TQ			Ice				
					W-DIELDIN							
	W-PCL-SQ-R	W-CARB-AA (5 ml of MCAA** Buffer Solution)		W-CT-CL-R								
Phytoplankton		DTY-QN-C		PKD-QN-BV			Ice					

RQ-2021-05-24-24

Collected By: Paul Blair, Jonathan Woods



Customer: FDEP

Field Report Prepared by: Paul Blair, Jonathan Woods

Project ID: SMP

Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: G1TLHR0129

Comments:

Location: Spring Creek Springs Middle

Matrix: Marine

(Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. ©	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPTH)			
5/27/2021 1250	2.45	31.4	23.6	7.33	0.3	1	1.1	16511	9.75			
Central Lab please use top row for login purposes												
Parameter Suite	Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group			
Preserved Nutrients (P-500ML)	X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP			X	Ice	X	H2SO4*	X	<2	1	B
	Lot #		120529		Exp:		10-20-21					
Metals (P-500ML)		W-ICPMS		W-ICP	W-HARD		Ice		HNO3*		<2	
	Lot #				Exp:							
Filtered Nutrients (P125ML)	X	W-PO4-F		X	Field Filtered 0.45 um Filter		X	Ice			1	B
	Filter Lot #		N/A									
	Syringe Lot #											
Chlorophyll (BP-1L)	CHLSUITE-W				Ice							
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W- TDS				Ice							
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				Ice							
Macroinvertebrates (PJ-2L)	MI-FW-QLDC				Ice							
Periphyton (PT-50ML)	ALGAE-ID				Ice							
Microbiology (Contract Lab Bottle)	E Coli		F Coli		Enterococci		Ice					
	Contract Lab											
Molecular (QPCR-P-500ML)	PCR-HF183		PCR-BacR		PCR-DG3		Ice					
	PCR-GULL2		PCR-GFD									
Tracers (BG-500ML)	W-E8321-DI		W-E8321-MS		Ice							
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI		W-E8321-MS		W-PSNP-TQ		Ice					
	W-PCL-SQ-R		W-CARB-AA (5 ml of MCAA** Buffer Solution)		W-DIELDRIIN							
					W-CT-CL-R							
Phytoplankton	DTY-QN-C		PKD-QN-BV		Ice							



Customer: FDEP

Field Report Prepared by: Paul Blair, Jonathan Woods

Project ID: SMP

Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: G1TLHR0151

Comments:

Location: Oyster Bay1

Matrix: Marine

(Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)			
5/27/2021 1240	6.87	98.2	27.5	7.96	0.3	1.3	1.7	33994	21.3			
Central Lab please use top row for login purposes												
Parameter Suite	Parameter Methods				Preservation		pH Check		# Bottles sent to Lab	Bottle Group		
Preserved Nutrients (P-500ML)	X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP			X	Ice	X	H2SO4*	X	<2	1	A
		Lot #	120529		Exp:	10-20-21						
Metals (P-500ML)		W-ICPMS		W-ICP	W-HARD		Ice	HNO3*		<2		
		Lot #			Exp:							
Filtered Nutrients (P125ML)	X	W-PO4-F		X	Field Filtered 0.45 um Filter		X	Ice			1	A
		Filter Lot #	N/A									
		Syringe Lot #	79900									
Chlorophyll (BP-1L)	X	CHL SUITE-W				X	Ice			1	A	
Physical/Aggregate (Fresh) (P-1L)		Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS					Ice					
Physical/Aggregate (Marine) (P-1L)		Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					Ice					
Macroinvertebrates (PJ-2L)		MI-FW-QLDC					Ice					
Periphyton (PT-50ML)		ALGAE-ID					Ice					
Microbiology (Contract Lab Bottle)		E Coli		F Coli		Enterococci		Ice				
		Contract Lab										
Molecular (QPCR-P-500ML)		PCR-HF183		PCR-BacR	PCR-DG3		Ice					
		PCR-GULL2		PCR-GFD								
Tracers (BG-500ML)		W-E8321-DI		W-E8321-MS			Ice					
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI	W-E8321-MS		W-PSNP-TQ		Ice					
					W-DIELDRIN							
		W-PCL-SQ-R	W-CARB-AA (5 ml of MCAA** Buffer Solution)		W-CT-CL-R							
Phytoplankton		DTY-QN-C		PKD-QN-BV			Ice					

RQ-2021-05-24-24

Collected By: Paul Blair, Jonathan Woods



RQ-2021-05-24-24	Collected By: Paul Blair, Jonathan Woods
Customer: FDEP	Field Report Prepared by: Paul Blair, Jonathan Woods
Project ID: SMP	Sampling Agency: FDEP
WIN ID/Field ID:  WIN ID/Field ID: G2TLHR0054	Comments:

Location: Alligator Harbor South

Matrix: Marine

(Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)	
5/27/2021 1110	6.95	103	26.7	7.85	0.3	0.8	1.2	46850	30.4	
Central Lab please use top row for login purposes										
Parameter Suite	Parameter Methods					Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					Ice	H2SO4*	<2		
	Lot #		Exp:							
Metals (P-500ML)	W-ICPMS		W-ICP		W-HARD	Ice	HNO3*	<2		
	Lot #		Exp:							
Filtered Nutrients (P125ML)	W-PO4-F		Field Filtered 0.45 µm Filter			Ice				
	Filter Lot #									
	Syringe Lot #									
Chlorophyll (BP-1L)	CHLSUITE-W					Ice			1	C
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W- TDS					Ice				
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					Ice				
Macroinvertebrates (PJ-2L)	MI-FW-QLDC					Ice				
Periphyton (PT-50ML)	ALGAE-ID					Ice				
Microbiology (Contract Lab Bottle)	E Coli	X	F Coll	X	Enterococci	X	Ice			
	Contract Lab									
Molecular (QPCR-P-500ML)	PCR-HF183		PCR-BacR		PCR-DG3	Ice				
	PCR-GULL2		PCR-GFD							
Tracers (BG-500ML)	W-E8321-DI		W-E8321-MS			Ice				
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI	W-E8321-MS			W-PSNP-TQ	Ice				
					W-DIELDRIN					
	W-PCL-SQ-R	W-CAR8-AA (5 ml of MCAA** Buffer Solution)			W-CT-CL-R					
Phytoplankton	DTY-QN-C		PKD-QN-BV			Ice				



RQ-2021-05-24-24

Customer: FDEP

Project ID: SMP

Collected By: Paul Blair, Jonathan Woods

Field Report Prepared by: Paul Blair, Jonathan Woods

Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: G2TLHR0025

Comments:

Location: Alligator Harbor near clam beds

Matrix: Marine

(Check One)

☒ Grab☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. ©	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)				
5/27/2021 1120	6.88	103.2	27.8	7.86	0.3	1.2	2.3	45692	29.53				
Central Lab please use top row for login purposes													
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group			
Preserved Nutrients (P-500ML)		X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				X	Ice	X	H2SO4*	<2	1	A
		Lot #		120529		Exp:		10-20-21					
Metals (P-500ML)			W-ICPMS		W-ICP	W-HARD		Ice		HNO3*	<2		
		Lot #				Exp:							
Filtered Nutrients (P125ML)		X	W-PO4-F		X	Field Filtered 0.45 um Filter		X	Ice			1	A
		Filter Lot #		N/A									
		Syringe Lot #		79900									
Chlorophyll (BP-1L)		x	CHLSUITE-W				x	Ice				1	A
Physical/Aggregate (Fresh) (P-1L)			Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W- TDS					Ice					
Physical/Aggregate (Marine) (P-1L)		X	Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				X	Ice					
Macroinvertebrates (PJ-2L)			MI-FW-QLDC					Ice					
Periphyton (PT-50ML)			ALGAE-ID					Ice					
Microbiology (Contract Lab Bottle)			E Coli		X	F Coli		X	Enterococci	X			
			Contract Lab										
Molecular (QPCR-P-500ML)			PCR-HF183			PCR-BacR			PCR-DG3		Ice		
			PCR-GULL2			PCR-GFD							
Tracers (BG-500ML)			W-E8321-DI			W-E8321-MS			Ice				
Pesticides (BG-500ML) (BG-1L)			W-E8321-DI			W-E8321-MS			W-PSNP-TQ		Ice		
									W-DIELDRIN				
			W-PCL-SQ-R			W-CARB-AA (5 ml of MCAA** Buffer Solution)			W-CT-CL-R				
Phytoplankton			DTY-QN-C			PKD-QN-BV			Ice				

RQ-2021-05-24-24

Collected By: Paul Blair, Jonathan Woods



Customer: FDEP

Field Report Prepared by: Paul Blair, Jonathan Woods

Project ID: SMP

Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: G1TLHR0131

Comments:

Location: Spring Creek Spring

Matrix: Marine

(Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)
5/27/2021 1300	0.66	7.8	21.6	7.22	0.3	3.8	9.6	11479	6.55

Central Lab please use top row for login purposes

Parameter Suite	Parameter Methods						Preservation			pH Check		# Bottles sent to Lab	Bottle Group			
Preserved Nutrients (P-500ML)	X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					X	Ice	X	H2SO4*		X	<2	1	B	
		Lot #		120529		Exp:		10-20-21								
Metals (P-500ML)		W-ICPMS		W-ICP		W-HARD			Ice		HNO3*			<2		
		Lot #				Exp:										
Filtered Nutrients (P125ML)	X	W-PO4-F		X	Field Filtered 0.45 um Filter			X	Ice					1	B	
		Filter Lot #		N/A												
		Syringe Lot #		7990												
Chlorophyll (BP-1L)	X	CHLSUITE-W					X	Ice			1			B		
Physical/Aggregate (Fresh) (P-1L)	X	Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W- TDS					X	Ice			1			B		
Physical/Aggregate (Marine) (P-1L)	X	Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY					X	Ice			1			B		
Macroinvertebrates (PJ-2L)		MI-FW-QLDC						Ice								
Periphyton (PT-50ML)		ALGAE-ID						Ice								
Microbiology (Contract Lab Bottle)		E Coli			F Coli		X	Enterococci		X	Ice			1	B	
		Contract Lab														
Molecular (QPCR-P-500ML)		PCR-HF183			PCR-BacR			PCR-DG3			Ice					
		PCR-GULL2			PCR-GFD											
Tracers (BG-500ML)		W-E8321-DI			W-E8321-MS				Ice							
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI			W-E8321-MS			W-P5NP-TQ			Ice					
							W-DIELDRIN									
		W-PCL-SQ-R			W-CARB-AA (5 ml of MCAA** Buffer Solution)		W-CT-CL-R									
Phytoplankton		DTY-QN-C			PKD-QN-BV			Ice								

RQ-2021-05-24-24

Collected By: Paul Blair, Jonathan Woods

Date of Request: 29-APR-2021
 Created By: ECKLES_M On: 29-APR-2021
 Modified By: JASROTIA_P On: 03-MAY-2021
 Customer: TLH-ROC
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Not Applicable
 Event Description: Run 7

Send Coolers To:

Phone: 850-245-7586
 2600 Blair Stone Rd
 Tallahassee, FL 32399
 Attn: Michael Eckles
 Cooler Ship EMail: michael.eckles@dep.state.fl.us

Priority: 3
 Event Status: S
 Criminal Investigation: NO
 Chemistry Request Reviewed By: WATTS_J On: 30-APR-2021
 Biology Request Reviewed By: JASROTIA_P On: 03-MAY-2021
 Sampling Kit Required: YES Ship on: 20-MAY-2021
 Sampling Kit Shipped: YES On: 13-MAY-2021
 Sampling Kit Packed By: MCCOY_I
 Preservatives Needed: NO
 Date to Receive Samples: 24-MAY-2021
 Logged In By:

Send Final Report To:

2600 Blair Stone Rd
 Tallahassee, FL 32399
 Attn: Michael Eckles
 Report Notification EMail:
 michael.eckles@floridadep.gov; terry.green@floridadep.gov

Comments: Group A WBIDs 1256 (0025) & 1176C (0052 & 0053)
 Group B WBID 1146Z (0129 & 0131)
 Group C WBID 1256 (0054)
 Group D WBID 1242 (0006)
 Group E Blank
 No Fedex Labels

Bottle Group A (Water) With 3 Samples:

Bottle Type: P-1L	Number of Bottles: 3	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-BR-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Bromide in aqueous matrices
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 3	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WO-THI	Number of Bottles: 3	Preserved With: ICE
ENT-24-QT	Template: DEFAULT	(Enterolert/QT) Enterococci by Enterolert Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-250ML-WO-THI	Number of Bottles: 3	Preserved With: ICE
FCOLI-MF	Template: DEFAULT	(SM 9222 D) Fecal coliforms by membrane filter method - non-chlorinated water systems
Bottle Type: P-500ML	Number of Bottles: 3	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO ₂ NO ₃	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group A (Water) With 3 Samples:

Bottle Type: P-125ML	Number of Bottles: 3	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group B (Water) With 2 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 2	Preserved With: ICE
FCOLI-MF	Template: DEFAULT	(SM 9222 D) Fecal coliforms by membrane filter method - non-chlorinated water systems

Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML	Number of Bottles: 2	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Type: P-1L	Number of Bottles: 2	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L	Number of Bottles: 2	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-250ML-WO-THI	Number of Bottles: 2	Preserved With: ICE
ENT-24-QT	Template: DEFAULT	(Enterolert/QT) Enterococci by Enterolert Quanti-Tray method - non-chlorinated water systems

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ENT-24-QT	Template: DEFAULT	(Enterolert/QT) Enterococci by Enterolert Quanti-Tray method - non-chlorinated water systems

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
FCOLI-MF	Template: DEFAULT	(SM 9222 D) Fecal coliforms by membrane filter method - non-chlorinated water systems

Bottle Group D (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Barium
Calcium
Iron
Magnesium
Manganese
Potassium
Sodium

Bottle Group D (Water) With 1 Samples:

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Beryllium		
Cadmium		
Chromium		
Copper		
Lead		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group E (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Barium		
Calcium		
Iron		
Magnesium		
Manganese		
Potassium		
Sodium		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Beryllium		
Cadmium		
Chromium		

Bottle Group E (Water) With 1 Samples:

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Copper
Lead
Nickel
Selenium
Silver
Thallium

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML	Number of Bottles: 1	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.