

Chemical Analysis Report

TLH-ROC-2021-06-10-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Run 11**
Request ID: **RQ-2021-05-31-19**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 02-JUL-2021 10:16

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the "T" and a long, sweeping underline.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Boggy Creek @ SR 22

Collection Date/Time: 06/10/2021 10:20

Field ID: G3TLHR0015

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2254423	EPA 180.1 Rev. 2.0	Turbidity	8.9		NTU	P399530	
	EPA 300.0 Rev. 2.1	Chloride	6.7		mg Cl/L	P399825	
	SM 2120 B-2011	Color (true)	180		PCU	P399533	
	SM 2320 B-2011	Total Alkalinity	11	J	mg CaCO3/L	P400153	LCS
	SM 2540 C-2011	TDS	54		mg/L	P400003	
	SM 2540 D-2011	TSS	11	I	mg/L	P399856	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P399739	
2254425	EPA 350.1 Rev. 2.0	Ammonia-N	0.022		mg N/L	P399990	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.66		mg N/L	P399715	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P399692	
	EPA 365.1 Rev. 2.0	Total-P	0.024		mg P/L	P400124	
	SM 5310 B-2011	Organic Carbon	13		mg C/L	P400167	
2254427	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P399513	

Ref. Method and Comment:

SM 2320 B-2011: Refer to the Lab Analysis Report for an explanation of QC Codes.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample. The PQL was elevated because of sample matrix interference.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Mule Creek @ Alvie Holmes Rd

Collection Date/Time: 06/10/2021 10:40

Field ID: G3TLHR0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2254424	EPA 180.1 Rev. 2.0	Turbidity	3.9		NTU	P399530	
	EPA 300.0 Rev. 2.1	Chloride	91		mg Cl/L	P399825	
	SM 2120 B-2011	Color (true)	23		PCU	P399533	
	SM 2320 B-2011	Total Alkalinity	7.4	J	mg CaCO3/L	P400153	LCS
	SM 2540 C-2011	TDS	177		mg/L	P400003	
	SM 2540 D-2011	TSS	3	I	mg/L	P399856	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P399739	
2254426	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P399991	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.16	I	mg N/L	P399715	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.010		mg N/L	P399692	
	EPA 365.1 Rev. 2.0	Total-P	0.009		mg P/L	P400124	
	SM 5310 B-2011	Organic Carbon	2.1		mg C/L	P400167	
2254428	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P399513	

Ref. Method and Comment:

SM 2320 B-2011: Refer to the Lab Analysis Report for an explanation of QC Codes.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P399530

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P399825

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P399990

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P399991

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P399715

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P399692

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P400124

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P399513

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P399533

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P400153

Component	Result	Code	Units
Total Alkalinity	0.72	I	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P400003

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P399856

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P399739

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P400167

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P399825

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P399990

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P399991

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P399715

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	92.7		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P399692

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P400124

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	99.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P399513

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	103		P	90 - 110

Reference Method: SM 2120 B-2011
 Batch ID: P399533

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-2011
 Batch ID: P400153

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	107		F	94 - 106

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P399739

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	102		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P400167

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.3		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P399825

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254247	Chloride	100		P	90 - 110
2254404	Chloride	92.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P399990

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254344	Ammonia-N	98.0	98.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P399991

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254436	Ammonia-N	97.8	99.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P399692

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254350	NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P400124

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254350	Total-P	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P399513

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254406	O-Phosphate-P	100	102	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P399739

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254134	Fluoride	102	102	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P400167

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2254351	Organic Carbon	96.2	104	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P399530

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254272	Turbidity	5.11	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P399825

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254247	Chloride	1.15	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P399990

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254344	Ammonia-N	0.0	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P399991

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254436	Ammonia-N	1.11	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P399715

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254378	Kjeldahl Nitrogen	5.32	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P399692

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254350	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P400124

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254350	Total-P	1.25	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P399513

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254406	O-Phosphate-P	1.29	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P399533

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254404	Color (true)	0.400	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P400153

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2255387	Total Alkalinity	0.495	Sample	P	0 - 5.2

Reference Method: SM 2540 C-2011
Batch ID: P400003

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254349	TDS	4.72	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P399739

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254134	Fluoride	0.479	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P400167

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2254351	Organic Carbon	3.88	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A105989

Included Lab Sample IDs: 2254427, 2254428

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.1	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A105999

Included Lab Sample IDs: 2254423, 2254424

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.0	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A106000

Included Lab Sample IDs: 2254423, 2254424

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	102	P/P	90 - 115
Color (true)	102	102	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A106065

Included Lab Sample IDs: 2254425, 2254426

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A106087

Included Lab Sample IDs: 2254423, 2254424

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	105	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A106105

Included Lab Sample IDs: 2254423, 2254424

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A106117

Included Lab Sample IDs: 2254425, 2254426

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	90.6	94.9	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A106195

Included Lab Sample IDs: 2254425, 2254426

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.2	98.8	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A106195

Included Lab Sample IDs: 2254425, 2254426

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.6	99.2	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A106248

Included Lab Sample IDs: 2254423, 2254424

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	106	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A106257

Included Lab Sample IDs: 2254425, 2254426

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A106258

Included Lab Sample IDs: 2254425, 2254426

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	102	102	P/P	90 - 110
Organic Carbon	102	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					5.11	
EPA 300.0 Rev. 2.1	Chloride	100	100	92.7		1.15	
EPA 350.1 Rev. 2.0	Ammonia-N	101	98.0	98.0			0.0
	Ammonia-N	99.4	97.8	99.0			1.11
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	92.7					5.32
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.5	100	100			0.0
EPA 365.1 Rev. 2.0	Total-P	99.0	102	103			1.25
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	103	100	102			1.29
SM 2120 B-2011	Color (true)	101				0.400	
SM 2320 B-2011	Total Alkalinity	107				0.495	
SM 2540 C-2011	TDS					4.72	
SM 4500 F-C-2011	Fluoride	102	102	102			0.479
SM 5310 B-2011	Organic Carbon	97.3	96.2	104			3.88

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2254423, 2254424
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2254423, 2254424
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2254425, 2254426
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2254425, 2254426
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2254425, 2254426
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2254425, 2254426
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2254427, 2254428
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2254423, 2254424
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2254423, 2254424
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2254423, 2254424
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2254423, 2254424
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2254423, 2254424
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2254425, 2254426

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/10/2021			06/10/2021 14:54	Sarah A Nixon	2254423, 2254424
EPA 300.0 Rev. 2.1	06/10/2021			06/17/2021 01:34	Lipi Saha	2254423
	06/10/2021			06/17/2021 01:46	Lipi Saha	2254424
EPA 350.1 Rev. 2.0	06/10/2021			06/21/2021 12:26	Ping Hua	2254425
	06/10/2021			06/21/2021 12:43	Ping Hua	2254426
EPA 351.2 Rev. 2.0	06/10/2021	06/15/2021 14:18	Benjamin T. Nordmann	06/16/2021 11:51	Virginia P. Brown	2254425
	06/10/2021	06/15/2021 14:18	Benjamin T. Nordmann	06/16/2021 11:53	Virginia P. Brown	2254426
EPA 353.2 Rev. 2.0	06/10/2021			06/15/2021 10:12	Beverly Y. Sanders	2254425
	06/10/2021			06/15/2021 10:13	Beverly Y. Sanders	2254426
EPA 365.1 Rev. 2.0	06/10/2021	06/23/2021 13:50	Shengyu Ding	06/24/2021 13:52	Shengyu Ding	2254425
	06/10/2021	06/23/2021 13:50	Shengyu Ding	06/24/2021 13:53	Shengyu Ding	2254426
EPA 365.1 Rev. 2.0 dissolved	06/10/2021			06/10/2021 14:01	Ping Hua	2254427, 2254428
SM 2120 B-2011	06/10/2021			06/10/2021 16:33	Madeline Gruver	2254424
	06/10/2021			06/10/2021 16:36	Madeline Gruver	2254423
SM 2320 B-2011	06/10/2021			06/23/2021 19:22	Patrick M. Roche	2254423
	06/10/2021			06/23/2021 19:29	Patrick M. Roche	2254424
SM 2540 C-2011	06/10/2021			06/15/2021 14:12	Yijie Li	2254423, 2254424
SM 2540 D-2011	06/10/2021			06/15/2021 14:12	Yijie Li	2254423, 2254424
SM 4500 F-C-2011	06/10/2021			06/15/2021 19:11	Dale L. Simmons	2254423
	06/10/2021			06/15/2021 19:16	Dale L. Simmons	2254424
SM 5310 B-2011	06/10/2021			06/23/2021 00:23	Madeline Gruver	2254425
	06/10/2021			06/23/2021 01:03	Madeline Gruver	2254426