

Chemical Analysis Report

TLH-ROC-2021-08-05-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Run 3**
Request ID: **RQ-2021-08-02-72**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Environmental Consultant

Date Certified: 30-AUG-2021 10:29



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Weaver Warrior @ Old RR Track

Collection Date/Time: 08/05/2021 10:15

Field ID: G1TLHR0118

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2267831	EPA 180.1 Rev. 2.0	Turbidity	1.5		NTU	P402070	
	EPA 300.0 Rev. 2.1	Chloride	3.6		mg Cl/L	P402413	
		Sulfate	0.22	I	mg SO4/L	P402416	
	SM 2120 B-2011	Color (true)	160		PCU	P402073	
	SM 2320 B-2011	Total Alkalinity	25		mg CaCO3/L	P402156	
	SM 2540 C-2011	TDS	59		mg/L	P402568	
	SM 2540 D-2011	TSS	3	I	mg/L	P402573	
	SM 4500 F-C-2011	Fluoride	0.041	I	mg F/L	P402158	
2267832	EPA 350.1 Rev. 2.0	Ammonia-N	0.012		mg N/L	P402386	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.91		mg N/L	P402785	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.007	I	mg N/L	P402232	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P402323	
	SM 5310 B-2011	Organic Carbon	18		mg C/L	P402839	
2267833	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.076		mg P/L	P402067	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P402070

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402413

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402416

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402386

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402785

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402232

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402323

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402067

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P402073

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P402156

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P402568

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P402573

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P402158

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P402839

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P402413

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P402416

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P402386

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P402785

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P402232

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P402323

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	108		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P402067

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2120 B-2011
 Batch ID: P402073

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.4		P	85 - 115

Reference Method: SM 2320 B-2011
 Batch ID: P402156

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	94 - 106

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P402158

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.1		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P402839

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402413

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267746	Chloride	98.4		P	90 - 110
2267940	Chloride	93.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402416

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267746	Sulfate	98.1		P	90 - 110
2267940	Sulfate	99.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402386

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267792	Ammonia-N	98.8	98.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402232

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267840	NO2NO3-N	97.5	97.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402323

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267941	Total-P	108	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402067

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267817	O-Phosphate-P	98.1	97.1	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P402158

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267837	Fluoride	99.8	99.2	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P402839

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267840	Organic Carbon	96.5	96.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P402070

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267704	Turbidity	6.56	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402413

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267746	Chloride	1.30	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402416

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267746	Sulfate	1.96	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402386

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267792	Ammonia-N	0.577	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402785

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267688	Kjeldahl Nitrogen	2.32	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402232

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267840	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402323

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267941	Total-P	0.724	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267817	O-Phosphate-P	0.619	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P402073

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267791	Color (true)	2.37	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P402156

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267837	Total Alkalinity	0.942	Sample	P	0 - 5.2

Reference Method: SM 2540 C-2011
Batch ID: P402568

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267837	TDS	6.37	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P402158

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267837	Fluoride	0.467	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P402839

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267840	Organic Carbon	0.0	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A107065

Included Lab Sample IDs: 2267833

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.4	99.2	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A107067

Included Lab Sample IDs: 2267831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.5	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A107068

Included Lab Sample IDs: 2267831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	99.4	P/P	90 - 115

Reference Method: SM 2320 B-2011

Run ID: A107112

Included Lab Sample IDs: 2267831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.9	105	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A107112

Included Lab Sample IDs: 2267831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A107136

Included Lab Sample IDs: 2267832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A107166

Included Lab Sample IDs: 2267831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.4	100	P/P	90 - 110
Sulfate	98.7	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A107211

Included Lab Sample IDs: 2267832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.6	98.7	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A107234

Included Lab Sample IDs: 2267832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	105	108	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A107399

Included Lab Sample IDs: 2267832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.4	95.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A107450

Included Lab Sample IDs: 2267832

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	96.7	99.5	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				6.56	
EPA 300.0 Rev. 2.1	Chloride	97.9	98.4	93.4	1.30	
	Sulfate	99.8	98.1	99.7	1.96	
EPA 350.1 Rev. 2.0	Ammonia-N	99.4	98.8	98.0		0.577
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104				2.32
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	101	97.5	97.5		0.0
EPA 365.1 Rev. 2.0	Total-P	108	108	107		0.724
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	98.1	97.1		0.619
SM 2120 B-2011	Color (true)	99.4			2.37	
SM 2320 B-2011	Total Alkalinity	102			0.942	
SM 2540 C-2011	TDS				6.37	
SM 4500 F-C-2011	Fluoride	99.1	99.8	99.2		0.467
SM 5310 B-2011	Organic Carbon	96.6	96.5	96.5		0.0

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2267831
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2267831
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2267831
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2267832
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2267832
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2267832
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2267832
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2267833
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2267831
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2267831
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2267831
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2267831
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2267831
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2267832

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/05/2021			08/05/2021 15:14	Sarah A Nixon	2267831
EPA 300.0 Rev. 2.1	08/05/2021			08/12/2021 02:21	Lipi Saha	2267831
EPA 350.1 Rev. 2.0	08/05/2021			08/12/2021 12:22	Roberta Tatti	2267832
EPA 351.2 Rev. 2.0	08/05/2021	08/20/2021 12:58	Benjamin T. Nordmann	08/24/2021 14:42	Virginia P. Brown	2267832
EPA 353.2 Rev. 2.0	08/05/2021			08/10/2021 09:04	Beverly Y. Sanders	2267832
EPA 365.1 Rev. 2.0	08/05/2021	08/11/2021 13:55	Simonne.V. Calhoun	08/13/2021 10:02	Shengyu Ding	2267832
EPA 365.1 Rev. 2.0 dissolved	08/05/2021			08/05/2021 16:01	Patrick M. Roche	2267833
SM 2120 B-2011	08/05/2021			08/05/2021 16:12	Sarah A Nixon	2267831
SM 2320 B-2011	08/05/2021			08/06/2021 20:10	Dale L. Simmons	2267831
SM 2540 C-2011	08/05/2021			08/11/2021 14:58	Yijie Li	2267831
SM 2540 D-2011	08/05/2021			08/11/2021 14:58	Yijie Li	2267831
SM 4500 F-C-2011	08/05/2021			08/06/2021 20:10	Dale L. Simmons	2267831
SM 5310 B-2011	08/05/2021			08/20/2021 23:57	Madeline Gruver	2267832