

RQ-2021-10-04-31

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 10)

SEST2110

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Arya

RQ-

2021-10-11-1060

2021-10-11-29

Project:

HABISMPITrend

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

07/22/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Charlie Hoey	10/12/21	813	23.1	754.0	8.5	8.49	99.2	—	1.06	P / F	L / F
ICV	"	"	815	23.8	754.0	8.4	8.39	99.2	—	1.06	P / F	L / F
CCV	Charlie Hoey	10/12/21	1558	23.3	752.5	8.4	8.51	99.8	—	1.06	P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. ET 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Charlie Hoey	10/12/21	818	1108F51	8/23	1000	1000	P / F	L / F
ICV	"	"	820	1102792	1/23	100	98.2	P / F	L / F
CCV	Charlie Hoey	10/12/21	1601	1105C52	5/23	5000	4851	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP ET 1100	Name	Date	Time GT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Charlie Hoey	10/12/21	825	210039	10/23	7.0	23.5	7.0	-34.4	P / F	L / F
Calibr.	"	"	827	2107E46	7/23	4.0	23.4	4.0	144.4	P / F	L / F
Calibr.	"	"	829	2012C64	6/22	10.0	23.6	10.0	-206.6	P / F	L / F
ICV	"	"	832	2012C64	6/22	10.0	23.6	9.98	-206.5	P / F	L / F
CCV	Charlie Hoey	10/12/21	1605	2107E46	7/23	4.0	23.2	3.98	144.2	P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;If mV are recorded: slope from 7 to 10 $+172.2$, slope from 4 to 7 $+178.8$ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 07/22/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Charlie Hoey	10/12/21	834	0.00	0.00	P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

RQ-2021-10-04-31

SEST2110

Form effective August 1, 2018