

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Acva RQ: 2021-10-25-67 Project: SCI-Training

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/19/2021

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Andrew Luering	10/26/21	720	22.9	752	8.5	8.51	99.0	✓	1.07	(P) F	(U) F
ICV	"	"	721	22.9	752	8.5	8.51	99.1	✓	1.07	(P) F	(U) F
CCV	S. Fischer	10/26/21	1520	26.8	750	7.9	7.91	98.8	✓	1.07	(P) F	(U) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Andrew Luering	10/26/21	722	1108FS1	8/2023	1000	1000	(P) F	(U) F
ICV	"	"	725	1102792	1/2023	100	98.0	(P) F	(U) F
CCV	S. Fischer	10/26/21	1522	1108FS1	08/23	1000	1001	(P) F	(U) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Andrew Luering	10/26/21	727	210039	10/2023	7.0	23.2	7.01	-32.9	(P) F	(U) F
Calibr.	"	"	729	2107E46	7/2023	4.0	23.1	4.00	146.5	(P) F	(U) F
Calibr.	"	"	731	2012064	6/2022	10.0	23.2	10.02	-205.4	(P) F	(U) F
ICV	"	"	731	2012064	6/2022	10.0	23.3	10.00	-205.6	(P) F	(U) F
CCV	S. Fischer	10/26/21	1524	2107E46	07/23	4.	23.2	4.02	143.7	(P) F	(U) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.5, slope from 4 to 7 179.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/19/2021

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Andrew Luering	10/26/21	735	0.00	0.00	(P) F	(U) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: