

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Arva

RQ: 2021-11-15-52

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/19/2021

DO BEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	DO Chart mg/L	Meter DO mg/L	%DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Andrew Luerling	11/30/21	755	18.8	758	9.3	9.26	99.8	—	1.06	P/F	L/F
ICV	"	"	758	14.1	758	9.2	9.23	99.8	—	1.06	P/F	L/F
CCV	Charlie Hoyer	11/30/21	1542	21.8	757.8	8.8	8.73	99.5	—	1.06	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Andrew Luerling	11/30/21	800	1108FS1	8/2023	1000	1000	P/F	L/F
ICV	"	"	802	1105C57	5/2023	100	99.7	P/F	L/F
CCV	Charlie Hoyer	11/30/21	1545	1105C52	5/23	5006	4802	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH BEP SOP FT 1000	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Andrew Luerling	11/30/21	804	213013	4/2023	7.0	21.8	7.02	-37.4	P/F	L/F
Calibr.	"	"	806	2107E46	7/2023	4.0	21.7	4.06	141.6	P/F	L/F
Calibr.	"	"	808	1106829	11/2022	10.0	21.8	10.04	-209.7	P/F	L/F
ICV	"	"	810	1106829	11/2022	10.0	21.8	10.02	-210.2	P/F	L/F
CCV	Charlie Hoyer	11/30/21	1547	2107E46	7/23	4.0	21.9	3.95	144.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.3, slope from 4 to 7 179 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/19/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Andrew Luerling	11/30/21	812	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: