

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hodor RQ: 2021-06-14-40 Project: SMPC CM

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 04/26/2021

DO DEP/SOP FD 1500	Name	Date	Time GT-ET	Temp °C	Baro meter mmHg	DO Chart mg/L	Meter DO mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	06/29/21	0730	23.1	758	8.5	8.5	99.7	—	1.06	P/F	L/F
ICV	//	//	0734	23.1	758	8.5	8.5	99.7	—	1.06	P/F	L/F
CCV	//	//	1650	26.5	758	8.0	8.11	99.4	—	1.06	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FD 1200	Name	Date	Time GT-ET	Lot #	Expir. Date	Standard umhos/cm	Meter Reading umhos/cm	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	06/29/21	0736	1105E62	05/23	1000	1000	P/F	L/F
ICV	//	//	0738	1102792	01/23	100	101.1	P/F	L/F
CCV	//	//	1652	2009A62	08/22	50K	49876	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP/SOP FD 1100	Name	Date	Time GT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	06/29/21	0740	2101B07	12/22	7.0	23.7	7.01	-20.1	P/F	L/F
Calibr.	I	I	0742	4012K91	12/22	4.0	23.6	4.01	158.7	P/F	L/F
Calibr.	I	I	0744	2012C64	06/22	10.0	23.6	10.01	188.6	P/F	L/F
ICV	I	I	0746	2012C64	06/22	10.0	23.6	10.12	187.3	P/F	L/F
CCV	//	//	1654	4012K91	12/22	4.0	23.6	4.13	151.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 -168.5, slope from 4 to 7 178.8 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 04/26/2021

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time GT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Drew Liddick	06/29/21	0748	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: