

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hado

RQ: 2021-04-19-37

Project: SMP & CM

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 04/26/2021

DO FDEP SOP FD 4500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	DO Chart mg/L	Meter DO mg/L	% DO	Probe Charge	Probe Gain	Pass Fail	Lab Field
Calibr.	Drew Liddick	07/12/21	0800	23.3	757	8.6	8.61	99.7	—	1.05	P/F	L/F
ICV	//	//	0804	23.3	757	8.6	8.61	99.7	—	1.05	P/F	L/F
CCV	//	//	1655	25.7	757	8.1	8.13	99.4	—	1.05	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond. EV 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass Fail	Lab Field
Calibr.	Drew Liddick	07/12/21	0806	1105 E62	05/23	1500	1500	P/F	L/F
ICV	//	//	0808	1102 A92	01/23	100	100	P/F	L/F
CCV	//	//	1652	2009 A62	08/23	50K	50111	P/F	L/F
CCV							30111	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP EV 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass Fail	Lab Field
Calibr.	Drew Liddick	07/12/21	0810	2101 B07	12/22	7.0	23.7	7.01	-22.1	P/F	L/F
Calibr.	I	I	0812	4012 K91	12/22	4.0	23.7	4.01	156.7	P/F	L/F
Calibr.	I	I	0814	2012 C64	06/22	10.0	23.7	10.01	-188.7	P/F	L/F
ICV	I	I	0816	2012 C64	06/22	10.0	23.7	10.12	-190.1	P/F	L/F
CCV	//	//	1654	4012 K91	12/22	4.0	23.4	4.11	154.7	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 -166.6, slope from 4 to 7 178.8 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 04/26/2021

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Drew Liddick	07/12/21	0818	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: